

```
#####
#
# NBP calibration data file for sensors
#
# NOTE:
# 1. In order for these calibrations to take affect, uwint must
# be restarted.
#
# 2. Please enter serial numbers for all sensors
#
# 3. Remember, when you check this file back into RCS, use the
# -u option. It MUST remain in /usr/local/packages/rvdas/config
#
# Revised August 10, 2000 K. Gavahan
# - Initial revision.
#
# Revised...
#
#####
#
# Ship - LMG or NBP
#
SHIP NBP
#
#####
#
# Science specific information
#
VESSEL: NBP
TITLE: NBP0206
NUMBER: NBP0206
START_DATE: 10/29/02
END_DATE: 11/09/02
CHIEF_SCIENTIST: Jamie Austin
PARTICIPATING SCIENTISTS: Suzanne O'Hara, Dave Caress, John Diebold
#
#-----
# Data specific information
#
# Base file name for data files
BASE_FILE: NBP0206
#
# NAVIGATION LOGGERS - loggers and data directory
NAV_LOGGERS: l_3df,l_furuno,l_gyr,l_ngl,l_pcode
DATA_LOC_NAV: /data/not-public/logger/nav
#
# UNDERWAY LOGGERS - loggers and data directory
UW_LOGGERS: l_met,l_sim,l_tsg,l_grv,l_bathy
DATA_LOC_UWAY: /data/not-public/logger/uw
#
NETWORK: science
#
#
LOGGER_LOC: /usr/local/packages/rvdas/bin
#
#
#####
```

```

#
# Geohysical information
#
# The value for YEAR can be obtained by running /usr/local/bin/decimal_year.
# It should be updated everytime the gravity offset is updated.
#
# Gravity offset information
#
CRUISE_ID: NBP0206
#GRAV_OFFSET      UPDATED 07-30-02
GRAV_OFFSET: 972345.03
DEPTH: BATHY
YEAR: 2002.81
#
#####
#
# Gravity QC
#
# LOCATION      : where the final data file is located
# GAP           : minimum allowable time gap      (in seconds)
# FIELD         : field where value can be found (starting at 1 after timestamp)
# NAME          : name of the field
# MIN           : minimum allowable value          (XXX means ignore)
# MAX           : maximum allowable value          (XXX means ignore)
# ROC           : maximum allowable rate of change (XXX means ignore)
# DELIMITER     : the delimiter for the body of data items (SPACE, COMMA, . . .)
#
#TAG LOCATION GAP FIELD NAME MIN MAX ROC DELIMITER
GRQC /data/current_cruise/geopdata/GRV/ 30 1 GRAVITY 5000 12000 100 SPACE
#####
#
# NBP PSP and PIR coefficients
# PSP serial number: 33090F3 Cal: 12-06-01
# PIR serial number: 33023F3 Cal: 12-07-01
#
# PSP and PIR
# PIR = mV / ( coeff V/wm^-2 * 10^3mV/V)
# PIR coefficient = 1/ (3.88x10^-6 * 10^3)
# PIR coefficient = 1/(3.88x10^-3)
# PIR coefficient = 257.7
#
# PIR = mV / ( coeff V/wm^-2 * 10^3mV/V)
# PIR coefficient = 1/ (4.13x10^-6 * 10^3)
# PIR coefficient = 1/(4.13x10^-3)
# PIR coefficient = 247.1
#
# PSP = mV / ( coeff v/wm^-2 * 10^3mV/V)
# PSP coefficient = 1/(8.19x10^-6 * 10^3)
# PSP coefficient = 1/(8.19x10^-3)
# PSP coefficient = 122.1
#NOTE: Pay attention to significant digits
PSP1 122.1
PIR1 247.1
#
#####
#
# NBP met

```

```

#
# PAR serial number: 6356 2/15/01
# PAR Calibration Factor = 6.08 V/uE/cm^2sec
# PAR Probe Dark offset = 0.3 mV
#
# PAR      1/6.08 (Dry V/uE/cm^2sec) 0.0003 (Probe Dark in Volts)
PAR 0.1645 0.0003
#
#####
#
# Transmissiometer
# Serial number CST-422PR
# Date Calibrated 12/20/01
#
# % transmission = (vsig - vd) / (vref - vd)
#
#   vd = 0.059
#   vref = 4.818
#
#   = (vsig - 0.059) / (4.818 - 0.059)
#
#   Vdark   Vref
TRAN 0.058   4.733
#
#####
#
# Engineering
#
# RPM pitch rudder
SENG 25.  10.  3400.  2500.  20.
PENG 25.  10.  3400.  2500.  20.
# Roll and Pitch Pot
POPI 4.0  4.0
# Seawater flow meter
SWFL 5.72
#
#####
#
# NBP winches
#
# Scale conversion information for the science
# winches on the NBP.  Sheave measurements made
# on 01/01/00.  Wire Pull tests done on dates
# indicated
#
# stbd winch      sheave diam= 28.125"  .714m
# 9/16" wire      wire diam  = 0.5625"  .014m
#      total circumference= 90.124"  2.289m
#      magnets      = 24
#      Payout Scale factor= 3.755      .095
#      Tension Scale Factor= 200
#      operation limit= 20,718 lb
#
# port winch      sheave diam= 28.125  .714m
# .680" wire      wire diam  = 0.680    .017m
#      total circumference= 90.493"  2.297m
#      magnets      = 24

```

```

#      Payout Scale factor= 3.77      0.096m
#      Tension Scale Factor= 180
#      operation limit= 20,150 lb
#
# baltic winch  sheave diam= 12.125    .308m
# .322" wire    wire diam  = 0.322     .008m
#      total circumference= 39.103" 0.993m
#      magnets    = 10
#      Payout Scale factor= 3.910     0.099m
#      Tension Scale Factor= 200
#      operation limit= 5,980 lb
#
# Load pin in waterfall winches is sending out an A/D
# value of 2 even under 0 tension
# Also, payout pos/neg is opposite other winches
# uwf winch      sheave diam= 12.125    .308m
#                wire diam  = 0.322     .00818m
#                total circumference= 39.103" 0.993m
#                magnets    = 10
#                Payout Scale factor= 3.910     0.0993m
#                Tension Scale Factor= 60
#                operation limit= 5,980 lb
#
# lwf winch      sheave diam= 12.125    0.308m
#                wire diam  = 0.3125 0.00794m
#                total circumference= 39.074" 0.993m
#                magnets    = 10
#                Payout Scale factor= 3.907     0.0993m
#                Tension Scale Factor= 60
#                operation limit= 6,565 lb
#
# wnc1 and WNC2 are old.
#
#winch  payout tension speed
#name fields are in format A:B where y=Ax+B
#
# new winch strings
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
#      a      b      c      d      e
SWNC -0.1    200    1.67  20718  -800
PWNC  0.1    180    1.67  20150   0
BWNC  0.1    62.5   1.67  5980   437.5
WWNC  -0.1    60    -1.67  5980   0
# old winch strings
LWF  -0.1    60    -1.67  6565   0
UWF  -0.1    60    -1.67  5980   0
WNC1  0.1    200   1.67  NAN    0
WNC2 -0.1    60    -1.67  NAN    0
#
#
#####
#####
##
***** Calibration factors for SBE 21 S/N 857 *****
***** Calibration Date of 12 Jul 02 *****

```

```
# currently in use
# Temperture calibration factors
%TEMPERTURE%
g 0.00424093119
h 0.000599112442
i 0.00000442824868
j -0.00000181049207
fo 1000.000
*
```

```
# conductivity calibration factors
%CONDUCTIVITY%
g -4.05383582
h 0.483602896
i 0.00125997786
j -0.0000345251231
p -0.0000000957
t 0.00000325
*
```

```
***** Remote Temperature Probe SN #4071 *****
***** Calibration Date of Apr. 16 2002 *****
# external temperature calibration factors
%EXTERNAL TEMPERATURE%
g 0.00435876516
h 0.00063986307
i 0.0000213393613
j 0.00000157577668
fo 1000.000
*
```

```
#
#####
#
```