

```
#####
# LMG calibration data file for sensors
#
# NOTE:
# 1. In order for these calibrations to take affect, uwint and rv_tsg must
# be restarted. (Remember, rv_tsg has parameters.)
#
# 2. Please enter serial numbers for all sensors
#
# 3. Remember, when you check this file back into RCS, use the
#    -u option. It MUST remain in /usr/local/packages/rvdas/config
#
# 4. The TSG calibration coefficients must be placed last in this file.
#
#
#####
# Ship - LMG or NBP
SHIP LMG
#
#####
#
# Cruise ID (i.e. LMG0505)
cruiseID LMG0713
#
#####
# LM Gould radiometer calibrations
# PSP ser#:31701F3 cal date: 25 Apr, 2007
# PIR ser#:32031F3 cal date: 03 Apr. 2007
# Instrument uVolts/W/m^2
PSP 8.51
PIR 3.91
#
#####
# Instrument Vdark Calib_Factor (ser#:6393, cal date: 1/16/2006
#instrument, Probe Dark(V), Calib Factor (Dry) (V/uE/cm^2sec)
# Dark value measured in situ 05 Jan, 2007
PAR -0.0050988 5.63
#
#####
# Transmisometer (ser#: CST-553DR, cal date: 30Jan06)
# Vdark Vref Path
TRAN 0.060 4.803 0.25
#
#
#####
```

```

#
# LMG winches
#
# Scale conversion information for the science winches on the LMG.
# Sheave measurements made on 01/01/00.
# Wire Pull tests done on dates indicated
#
# Dush 4 winch    sheave diam=
# 9/16" wire    wire diam =
#      total circumference=
#      magnets =
#      Payout Scale factor=
#      Tension Scale Factor=
#          operation limit= lb
#
# Dush xx winch    sheave diam= 28.125   .714m
# .680" wire    wire diam = 0.680   .017m
#      total circumference= 90.493" 2.297m
#      magnets = 24
#      Payout Scale factor= 3.77   0.096m
#      Tension Scale Factor= 180
#          operation limit= 20,150 lb
#
#
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
# operation limit = d
#     a      b      c      d      e
LDU4  1      0.465  1  20718  0
LDU5  1      1      1  20150  0
LD11  1      1      1  5980   0
LWN1  1      1      1  5980   0
#SWNC -0.1   200   1.67  20718 -800
#PWNC 0.1    180   1.67  20150  0
#BWNC  0.1   62.5  1.67  5980   437.5
#WWNC -0.1   60    -1.67  5980   0
#
#

```

```

***** Remote Temperature Probe SN #4015 *****
***** Calibration Date of 11-May-07 *****
# external temperature calibration factors
%EXTERNAL TEMPERATURE%
g 0.00436651867

```

```

h 0.000627057431
i 0.0000215854061
j 0.00000173345987
f0 1000.0
*

#####
##### Note, TSG calibrations must be last in this file #####
##### Do not change the formating, only the values. Thanks #####
#####

## **** Calibration factors for SBE 21 S/N 3208 ****
***** Calibration Date of 30 June 2005 ****
# currently in use
# Temperture calibration factors
%TEMPERATURE%

# g 0.00420097598
# h 0.000591574677
# i 0.00000218319415
# j -0.00000209040478
# f0 1000.000
g 0.00413343557
h 0.000615389618
i 0.0000197232448
j 0.00000135946639
f0 1000.000
*

# conductivity calibration factors
%CONDUCTIVITY%

# g -3.93320202
# h 0.467907879
# i 0.00171904247
# j -0.0000536817820
# p -0.0000000957
# t 0.00000325
g -3.99168762
h 0.471887572
i -0.000550516050
j 0.0000513922485
p -0.0000000957
t 0.00000325

```

*

#

#