

```
#####
# LMG calibration data file for sensors
#
# NOTE:
# 1. In order for these calibrations to take affect, uwint and rv_tsg must
# be restarted. (Remember, rv_tsg has parameters.)
#
# 2. Please enter serial numbers for all sensors
#
# 3. Remember, when you check this file back into RCS, use the
# -u option. It MUST remain in /usr/local/packages/rvdas/config
#
# 4. The TSG calibration coefficients must be placed last in this file.
#
#
#####
# Ship - LMG or NBP
SHIP LMG
#
#####
# Cruise ID (i.e. LMG0505)
cruiseID LMG0803
#
#####
# LM Gould radiometer calibrations
# PSP ser#:28933F3 cal date: 31 Aug, 2007
# PIR ser#:28903F3 cal date: 31 Aug. 2007
# Instrument uVolts/W/m^2
PSP 8.05
PIR 3.73
#
#####
# Instrument Vdark Calib_Factor (ser#:6394, cal date: 8/29/2006
#instrument, Probe Dark(V), Calib Factor (Dry) (V/uE/cm^2sec)
# Dark value measured in situ 28 Dec, 2007
PAR -0.0063735 5.26
#
#####
# Transmisometer (ser#: CST-891DR, cal date: 20Nov06)
# Vdark Vref Path
TRAN 0.057 4.829 0.25
#
#
#####
#
```

```

#
# LMG winches
#
# Scale conversion information for the science winches on the LMG.
# Sheave measurements made on 01/01/00.
# Wire Pull tests done on dates indicated
#
# Dush 4 winch    sheave diam=
# 9/16" wire    wire diam =
#      total circumference=
#      magnets   =
#      Payout Scale factor=
#      Tension Scale Factor=
#      operation limit=  lb
#
# Dush xx winch    sheave diam= 28.125   .714m
# .680" wire    wire diam = 0.680   .017m
#      total circumference= 90.493" 2.297m
#      magnets   = 24
#      Payout Scale factor= 3.77   0.096m
#      Tension Scale Factor= 180
#      operation limit= 20,150 lb
#
#
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
# operation limit = d
#      a      b      c      d      e
LDU4  1      0.465  1  20718  0
LDU5  1      1      1  20150  0
LD11  1      1      1  5980   0
LWN1  1      1      1  5980   0
#SWNC -0.1   200    1.67 20718 -800
#PWNC 0.1    180    1.67 20150  0
#BWNC 0.1    62.5   1.67 5980   437.5
#WWNC -0.1   60     -1.67 5980   0
#
#
#####
#####
#### Note, TSG calibrations must be last in this file #####
#### Do not change the formatting, only the values. Thanks #####
#####
##
#***** Calibration factors for SBE 21 S/N 3208 *****

```

***** Calibration Date of 30-Jun-05 2007 *****

currently in use

Temperture calibration factors

%TEMPERTURE%

g 0.00413343557

h 0.000615389618

i 0.0000197232448

j 0.00000135946639

fo 1000.000

*

conductivity calibration factors

%CONDUCTIVITY%

g -3.99168762

h 0.471887572

i -0.00055051605

j 0.0000513922485

p -0.0000000957

t 0.00000325

*

***** Remote Temperature Probe SN # 4015 *****

***** Calibration Date of 11-May-07 *****

external temperature calibration factors

%EXTERNAL TEMPERATURE%

g 0.0043665187

h 0.000627057431

i 0.0000215854061

j 0.00000173345987

fo 1000.000

*

#

#

#