

```
#####  
#  
# NBP calibration data file for sensors  
#  
# NOTE:  
# 1. In order for these calibrations to take affect, uwint must  
# be restarted.  
#  
# 2. Please enter serial numbers for all sensors  
#  
# 3. Remember, when you check this file back into RCS, use the  
# -u option. It MUST remain in /usr/local/packages/rvdas/config  
#  
# Revised August 10, 2000 K. Gavahan  
# - Initial revision.  
#  
# Revised...  
#  
#####  
#  
# Ship - LMG or NBP  
#  
SHIP NBP  
#  
#####  
#  
# Science specific information  
#  
VESSEL: NBP  
TITLE: NBP0801  
NUMBER: NBP0801  
START_DATE: 01/08/2008  
END_DATE: 01/30/2008  
CHIEF_SCIENTIST: Bruce Huber, Dave Caron  
PARTICIPATING SCIENTISTS:  
#  
#-----  
# Data specific information  
#  
# Base file name for data files  
BASE_FILE: NBP0801  
#  
# NAVIGATION LOGGERS - loggers and data directory  
NAV_LOGGERS: l_gyr,l_pcode,l_seap  
DATA_LOC_NAV: /data/logger/nav  
#
```

```
# UNDERWAY LOGGERS - loggers and data directory
UW_LOGGERS: l_met,l_sim,l_tsg,l_bathy
DATA_LOC_UWAY: /data/logger/uw
#
NETWORK: science
#
#
LOGGER_LOC: /usr/local/packages/rvdas/bin
#
#
#####
#
# Geophysical information
#
# The value for YEAR can be obtained by running /usr/local/bin/
decimal_year.
# It should be updated everytime the gravity offset is updated.
#
# Gravity offset information
#
CRUISE_ID: NBP0801
#GRAV_OFFSET: 972326.93 on 0401.
#GRAV_OFFSET: 972327.68 on 0402.
#GRAV_OFFSET: 972304.02 on 0501
#GRAV_OFFSET: 972305.39 on 0501b
#GRAV_OFFSET: 972308.37 on 0506
#GRAV_OFFSET: 972310.34 on 0601
#GRAV_OFFSET: 972311.94
#GRAV_OFFSET: 972309.88 on 0602
#GRAV_OFFSET: 972311.96
#GRAV_OFFSET: 972312.25 on 0603
#GRAV_OFFSET: 972311.95
#GRAV_OFFSET: 972318.07
#GRAV_OFFSET: 972319.71
#GRAV_OFFSET: 972320.26
#GRAV_OFFSET: 972320.07
#GRAV_OFFSET: 972319.42
#GRAV_OFFSET: 972316.86
#GRAV_OFFSET: 972318.65
#GRAV_OFFSET: 972316.67
#GRAV_OFFSET: 972302.95 on 0710
#GRAV_OFFSET: 972302.90
GRAV_OFFSET: 972309.39
YEAR: 2008.02
#
#####
```

```

#
# Gravity QC
#
# LOCATION      : where the final data file is located
# GAP           : minimum allowable time gap      (in seconds)
# FIELD         : field where value can be found (starting at 1 after
timestamp)
# NAME          : name of the field
# MIN           : minimum allowable value          (XXX means ignore)
# MAX           : maximum allowable value          (XXX means ignore)
# ROC           : maximum allowable rate of change (XXX means ignore)
# DELIMITER     : the delimiter for the body of data items (SPACE,
COMMA, . . .)
#
#TAG LOCATION GAP FIELD NAME MIN MAX ROC DELIMITER
GRQC /data/current_cruise/geopdata/GRV/ 30 1 GRAVITY 4000 12000 100 SPACE
#####
#
# NBP PSP and PIR coefficients
# PSP serial number: 33090F3    Cal: 27 Feb 2007
# PIR serial number: 33023F3    Cal: 8 Mar 2007
#
# PSP and PIR assumed to be cal number * 10^-6
#
# Calculations make in the programs....
# PIR = mV / ( coeff V/wm^-2 * 10^3mV/V)
# PSP = mV / ( coeff v/wm^-2 * 10^3mV/V)
#
# For example, program will calculate
# PIR coefficient = 1/ (3.92x10^-6 * 10^3) = 255.1
# PSP coefficient = 1/(8.05x10^-6 * 10^3) = 124.22
#
PSP1 8.16
PIR1 3.87
#####
#
# NBP met
#
# PAR serial number: 6356 8 Aug 07
# PAR Calibration Factor = 5.8644 V/uE/cm^2sec
# PAR Probe Dark offset = -0.1 mV
#
# old PAR      1/6.30 (Dry V/uE/cm^2sec) 0.0021 (Probe Dark in Volts)
#
# PAR          1/6.06 (Dry V/uE/cm^2sec) 0.0004 (Probe Dark in Volts)
# PAR          1/6.10 (Dry V/uE/cm^2sec) 0.0004 (Probe Dark in Volts)

```

```

# PAR      1/6.10 (Dry V/uE/cm^2sec) 0.0004 (Probe Dark in Volts)
# PAR      1/6.27 (Dry V/uE/cm^2sec) 0.0001 (Probe Dark in Volts)
# PAR      1/5.44 (Dry V/uE/cm^2sec) 0.00201 (Probe Dark in Volts)
# PAR      1/5.8644 (Dry V/uE/cm^2sec) -0.0001 (Probe Dark in Volts)
#
PAR 0.17 0.0001
#
#####
#
# Transmissometer HydroDAS
# Serial number CST-397DR
# Date Calibrated 02/13/2006
#
# % transmission = (vsig - vd) / (vref - vd)
#
#   vd = 0.061
#   vref = 4.772
#
#   = (vsig - 0.061) / (4.772 - 0.061)
#
#   Vdark   Vref
TRAN 0.061 4.772
#
#
#####
#
# Fluorometer
# Serial number AFL-044
# Date Calibrated 5/31/2006
#
# Chlorophyll concentration in mg/l(mg/m**3) =
#   (Vmeasured - CWO) * SF (scale factor)
#
#   CWO = Vblank = dark counts
#
#   Dark Ccounts = 0.177
#   SF = 9.8116
#
#   = (Vmeasured - 0.177) * 9.8116
#
#   Dark_Counts SF
FLUR 0.177 9.8116
#
#####
#
# Flow meters

```

```

#
FL01  0.026
FL02  0.024
FL03  0.06
FL04  0.055
#
#####
#
# Engineering
#
# RPM pitch rudder
SENG  25.  10.  3400.  2500.  20.
PENG  25.  10.  3400.  2500.  20.
# Roll and Pitch Pot
POPI  4.0  4.0
# Seawater flow meter
# swfl *c1  +c2
SWFL  48.0  0
#
#####
#
# NBP winches
#
# Scale conversion information for the science
# winches on the NBP.  Sheave measurements made
# on 01/01/00.  Wire Pull tests done on dates
# indicated
#
# stbd winch      sheave diam= 28.125"  .714m
# 9/16" wire      wire diam  = 0.5625"  .014m
#      total circumference= 90.124"  2.289m
#      magnets      = 24
#      Payout Scale factor= 3.755      .095
#      Tension Scale Factor= 200
#      operation limit= 20,718 lb
#
# port winch      sheave diam= 28.125  .714m
# .680" wire      wire diam  = 0.680    .017m
#      total circumference= 90.493"  2.297m
#      magnets      = 24
#      Payout Scale factor= 3.77      0.096m
#      Tension Scale Factor= 180
#      operation limit= 20,150 lb
#
# baltic winch    sheave diam= 12.125  .308m
# .322" wire      wire diam  = 0.322    .008m

```

```

#      total circumference= 39.103" 0.993m
#      magnets      = 10
#      Payout Scale factor= 3.910  0.099m
#      Tension Scale Factor= 200
#      operation limit= 5,980 lb
#
# Load pin in waterfall winches is sending out an A/D
# value of 2 even under 0 tension
# Also, payout pos/neg is opposite other winches
# uwf winch      sheave diam= 12.125  .308m
#      wire diam  = 0.322  .00818m
#      total circumference= 39.103" 0.993m
#      magnets      = 10
#      Payout Scale factor= 3.910  0.0993m
#      Tension Scale Factor= 60
#      operation limit= 5,980 lb
#
# lwf winch      sheave diam= 12.125  0.308m
#      wire diam  = 0.3125 0.00794m
#      total circumference= 39.074" 0.993m
#      magnets      = 10
#      Payout Scale factor= 3.907  0.0993m
#      Tension Scale Factor= 60
#      operation limit= 6,565 lb
#
# wnc1 and WNC2 are old.
#
# winch payout tension speed
# name fields are in format A:B where y=Ax+B
#
# new winch strings
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
#      a      b      c      d      e
SWNC -0.1  200  1.67  20718  -800
PWNC 0.1   180  1.67  20150  0
BWNC 0.1   62.5 1.67  5980   437.5
WWNC -0.1  60   -1.67 5980   0
# old winch strings
LWF -0.1  60   -1.67 6565   0
UWF -0.1  60   -1.67 5980   0
WNC1 0.1  200 1.67  NAN  0
WNC2 -0.1 60  -1.67  NAN  0
#
#

```

```
#####  
#####  
##  
#***** Calibration factors for SBE 21 S/N 3198 *****  
#***** Calibration Date of 20 May 06 *****  
# c:urrently in use  
# Temperature calibration factors  
%TEMPERTURE%  
g 0.00422473160  
h 0.000629770835  
i 0.0000205272425  
j 0.00000154706370  
fo 1000.000  
*  
  
# conductivity calibration factors  
%CONDUCTIVITY%  
g -4.27061383  
h 0.504384737  
i -0.000453257393  
j 0.0000472934991  
p -0.000000095700  
t 0.0000032500  
*  
  
#***** Remote Temperature Probe SN #1267 *****  
#***** Calibration Date of 12-Apr-06 *****  
# external temperature calibration factors  
%EXTERNAL TEMPERATURE%  
g 0.00476625066  
h 0.000664522185  
i 0.0000284261863  
j 0.00000262601374  
fo 1000.000  
*  
  
#  
#####
```