

# **Data Report NBP0812:**

**Drake Passage**

**Dr. Teresa Chereskin**

**November 19 – December 14, 2008**



**United States Antarctic Program**

**RVIB Nathaniel B. Palmer**

**Raytheon Polar Services**

**Prepared by Paul Huckins & Mike Coons**

## Table of Contents

<b>INTRODUCTION .....</b>	<b>1</b>
<b>DISTRIBUTION CONTENTS AT A GLANCE .....</b>	<b>2</b>
EXTRACTING DATA .....	3
<b>DISTRIBUTION CONTENTS .....</b>	<b>4</b>
CRUISE INFORMATION .....	4
<i>Cruise Track</i> .....	4
<i>Satellite Images</i> .....	4
NBP DATA PRODUCTS .....	4
<i>JGOFS</i> .....	5
<i>MGD77</i> .....	6
SCIENCE OF OPPORTUNITY .....	7
<i>ADCP</i> .....	7
<i>pCO<sub>2</sub></i> .....	7
CRUISE SCIENCE .....	8
<i>CTD</i> .....	8
RVDAS .....	8
<i>Sensors and Instruments</i> .....	8
Underway Sensors .....	9
Meteorology and Radiometry .....	9
Geophysics .....	9
Oceanography .....	9
Navigational Instruments .....	10
<i>Data</i> .....	10
Underway Data .....	11
Sound Velocity Probe (svp1) .....	11
Meteorology (mw1) .....	11
Knudsen (knud) .....	12
Fluorometer (flr1) .....	13
pCO <sub>2</sub> (pco2) .....	13
Simrad EM120 (mbdp) .....	13
Micro-TSG (mtsg) .....	14
Gravimeter (grv1) .....	14
Engineering (eng1) .....	15
Hydro-DAS (hdas) .....	16
GUV Data (pguv) .....	16
Remote Temperature (rtmp) .....	16
Navigational Data .....	17
Seapath GPS (seap) .....	17
Trimble (P-Code) GPS (PCOD) .....	19
Gyro Compass (gyr1) .....	20
ADCP Course (adcp) .....	20
Processed Data .....	21
pCO <sub>2</sub> -merged .....	21
<b>CALCULATIONS .....</b>	<b>22</b>
PAR .....	22
PSP .....	25
PIR .....	23
<b>ACQUISITION PROBLEMS AND EVENTS .....</b>	<b>24</b>

<b>APPENDIX: SENSORS AND CALIBRATIONS .....</b>	<b>25</b>
NBP0812 Shipboard Sensors: .....	25
METEOROLOGY & RADIOMETERS .....	25
UNDERWAY .....	25
NBP0812 CTD Sensors:.....	26
CALIBRATIONS .....	26
Gravity Tie Start (Punta Arenas) .....	27
METEOROLOGY SYSTEM.....	28
Anemometer (Bridge).....	28
Barometer .....	29
Humidity Sensor .....	30
Temperature Sensor .....	31
PIR.....	32
PSP .....	33
PAR.....	34
GUV.....	35
Underway Temperature Sensor (Primary Remote).....	36
Underway Temperature Sensor (Secondary Remote).....	37
MICRO-TSG CALIBRATION FILES .....	38
Temperature Sensor (Micro-TSG).....	38
Conductivity Sensor (Micro-TSG).....	39
Underway fluorometer (Primary) .....	40
Underway Transmissometer (Primary) .....	41
CTD Pressure Sensor (fish).....	42
CTD Primary Temperature Sensor.....	43
CTD Secondary Temperature Sensor.....	44
CTD Primary Conductivity Sensor.....	45
CTD Secondary Conductivity Senesor.....	48

## Introduction

The NBP data acquisition systems continuously log data from the instruments used during the cruise. This document describes:

- The structure and organization of the data on the distribution media
- The format and contents of the data strings
- Formulas for calculating values
- Information about the specific instruments in use during the cruise
- A log of acquisition problems and events during the cruise that may affect the data
- Scanned calibration sheets for the instruments in use during the cruise.

The data is distributed on a DVD-R or CD-ROM written in ISO9660 level-1 format. It is readable by virtually every computing platform.

All the data has been compressed using Unix “gzip,” identified by the “.gz” extension. It has been copied to the distribution media in the Unix tar archive format, “.tar” extension. Tools are available on all platforms for uncompressing and de-archiving these formats: On Macintosh, use Stuffit Expander with DropStuff. On Windows operating systems use WinZip.

MultiBeam and BathyW data, if collected, are distributed separately.

*IMPORTANT: Read the last section, “Acquisition Problems and Events,” for important information that may affect the processing of this data.*

## Distribution Contents at a Glance

Volume 1 of 1: NBP0812		Description
File		
/		Root level directory
	NBP0812.trk	Text file of cruise track (lat,lon)
	NBP0812.mgd	Full Cruise MGD77 data file
	NBP0812.gmt	GMT binary file of MGD77 data
	INSTCOEF.TXT	Instrument Coefficient File
	0812DATA.doc	Data Report NBP0812
	SIP_RSP.zip	Copies of the SIP and RSP
/plots		Cruise track plots
	NBP0812_Track.ps	Cruise track plot (PostScript format)
	NBP0812_Track.jpg	Cruise track plot (JPEG format)
/process		Processed data
	0812jgof.tar	JGOFS format data files
	0812qc.tar	Daily RVDAS QC postscript plots
	0812pco2.tar	Merged pCO <sub>2</sub> data files
	0812mgd.tar	MGD Data
	0812proc.tar	Other processed data
/rvdas/nav		Navigation data
	0812gyr1.tar	Gyro raw data
	0812PCOD.tar	Trimble P-code raw data
	0812seap.tar	Seapath data
	0812adcp.tar	ADCP Data Sets
/rvdas/uw		Underway data
	0812ctdd.tar	CTD Depth
	0812eng1.tar	Engineering Data
	0812grv1.tar	Gravimeter raw data
	0812hdas.tar	HydroDAS raw data
	0812knud.tar	Knudsen raw data
	0812mbdp.tar	Multi-beam Depth
	0812mtsg.tar	Micro TSG data
	0812mwx1.tar	Meteorology raw data
	0812pco2.tar	pCO <sub>2</sub> raw data
	0812pguv.tar	GUV raw data
	0812rtmp.tar	Remote temperature data
	0812svp1.tar	Sound velocity probe (in ADCP well)
/Imagery		Other Data, Files and Pictures
	0812lmag.tar	Satellite imagery
/ocean		
	0812ctd.tar	Ctd data

Volume 2 of the data set is composed of cruise specific data collected by the science party. As such it is not described in detail in this document. A general table of contents of that disk is as follows:

Volume 2 of 2: NBP0812_Supplemental	
File	Description
/	Root level directory
ADCP	<div style="border: 1px solid black; padding: 5px;">           This data is not described in detail in this report. It is included to as a convenience to the science party.         </div>
AutoSal	
Ladcp	
Modem	
surveys	
08 Telemetry	
Ptd	
photos	

## Extracting Data

The Unix tar command has many options. It is often useful to know exactly how an archive was produced when expanding its contents. All archives were created using the command,

```
tar cvf archive_filename files_to_archive
```

To create a list of the files in the archive, use the Unix command,

```
tar tvf archive_filename > contents.list
```

where `contents.list` is the name of the file to create

To extract the files from the archive:

```
tar xvf archive_filename file(s)_to_extract
```

G-zipped files will have a “.gz” extension on the filename. These files can be decompressed after de-archiving, using the Unix command,

```
gunzip filename.gz
```

## Distribution Contents

### Cruise Information

NBP0812 departed Punta Arenas, Chile on November 19, 2008. At NSF's request we departed a day early in order to visit Palmer Station for medical reasons. There were no major problems during the cruise. The weather and the seas were very cooperative for the majority of the cruise. The basic objective was to visit moorings deployed on NBP0710. The NBP returned to Punta Arenas on December 13, 2008. We were accompanied by an Argentine observer and collected data in Argentine waters by permit. A separate data set has been created for Argentina.

The cruise data distribution is in 2 DVDs. The second DVD is composed of cruise specific data collected by the science party. It is included for their convenience and as such it is not described in detail in this document.

Copies of the SIP and RSP have been provided in the root directory of Vol 1 of the data set.

### Cruise Track

The distribution DVD includes a GMT cruise track file (NBP0812.trk). It contains the longitude and latitude at one-minute intervals extracted from the NBP0812.gmt file.

JPEG and PostScript cruise track files have been produced and placed in the /plots directory.

### Satellite Images

Satellite Images received for this cruise can be found in the file called /Imagery/0812Imag.tar collected and processed on the ship is in two further subdirectories, Ice and WX (weather). Files are named using the convention, ssss\_fff\_mmddyy\_tttt\_ww.gif where:

ssss_fff	= satellite and flight number
mm	= month
dd	= day
yy	= year
tttt	= time in hours and minutes (UTC)
ww	= optional field for identifying wavelength, such as vis (visible) or IR (infrared)

### NBP Data Products

Two datasets are created on each cruise: JGOFS and MGD77.

The data processing scripts used to produce JGOFS and MGD77 data sets create a lot of intermediate files. These files are included on the data distribution media in a file called /process/0812proc.tar. They are included to make re-processing easier in the event of an error, but no extensive detail of the formats is included in this document. If you have any questions, please contact [itvessel@usap.gov](mailto:itvessel@usap.gov).

## JGOFS

The JGOFS data set can be found on the distribution media in the file /process/0812jgof.tar. The archive contains one file produced for each day named jgDDD.dat.gz, where DDD is the year-day the data was acquired. The “.gz” extension indicates that the individual files are compressed before archiving. Each daily file consists of 22 columnar fields in text format as described in the table below. The JGOFS data set is created from calibrated data decimated at one-minute intervals. Several fields are derived measurements from more than a single raw input. For example, Course Made Good (CMG) and Speed Over Ground (SOG) are calculated from gyro and GPS inputs. Daily plots during the cruise are produced from the JGOFS data set. Note: Null, unused, or unknown fields are indicated as “NAN” as 9999 in the JGOFS data.

Field	Data	Units
01	GMT date	dd/mm/yy
02	GMT time	hh:mm:ss
03	NGL latitude (negative is South)	tt.tttt
04	NGL longitude (negative is West)	ggg.gggg
05	Speed over ground	Knots
06	GPS HDOP	-
07	Gyro Heading	Degrees (azimuth)
08	Course made good	Degrees (azimuth)
09	Mast PAR	$\mu\text{Einsteins}/\text{meter}^2 \text{ sec}$
10	Sea surface temperature	$^{\circ}\text{C}$
11	Sea surface conductivity	siemens/meter
12	Sea surface salinity	PSU
13	Sea depth (uncorrected, calc. sw sound vel. 1500 m/s)	meters
14	True wind speed (max speed windbird)	meters/sec
15	True wind direction (max speed windbird)	degrees (azimuth)
16	Ambient air temperature	$^{\circ}\text{C}$
17	Relative humidity	%
18	Barometric pressure	mBars
19	Sea surface fluorometry	volts (0-5 FSO)
20	Not used	-
21	PSP	$\text{W}/\text{m}^2$
22	PIR	$\text{W}/\text{m}^2$



## MGD77

The MGD77 data set is contained in a single file for the entire cruise. It can be found in the top level of the distribution data structure as NBP0812.mgd. The file NBP0812.gmt is created from the MGD77 dataset using the “mgd77togmt” utility. NBP0812.gmt can be used with the GMT plotting package.

The data used to produce the NBP0812.mgd file can be found on the distribution media in the file /process/0812proc.tar. The data files in the archive contain a day’s data and follow the naming convention Dddd.fnl.gz, where ddd is the year-day. These files follow a space-delimited columnar format that may be more accessible for some purposes. They contain data at one-second intervals rather than one minute and are individually “gzipped” to save space. Below is a detailed description of the MGD77 data set format. The other files in the archive contain interim processing files and are included to simplify possible reprocessing of the data using the RVDAS NBP processing scripts.

All decimal points are implied. Leading zeros and blanks are equivalent. Unknown or unused fields are filled with 9’s. All “corrections”, such as time zone, diurnal magnetics, and EOTVOS, are understood to be added.

Col	Len	Type	Contents	Description, Possible Values, Notes
1	1	Int	Data record type	Set to “5” for data record
2-9	8	Char	Survey identifier	
10-12	3	int	Time zone correction	Corrects time (in characters 13-27) to GMT when added; 0 = GMT
13-16	4	int	Year	4 digit year
17-18	2	int	Month	2 digit month
19-20	2	int	Day	2 digit day
21-22	2	int	Hour	2 digit hour
23-27	5	real	Minutes x 1000	
28-35	8	real	Latitude x 100000	+ = North - = South. (–9000000 to 9000000)
36-44	9	real	Longitude x 100000	+ = East - = West. (–18000000 to 18000000)
45	1	int	Position type code	1=Observed fix 3=Interpolated 9=Unspecified
46-51	6	real	Bathymetry, 2- way travel time	In 10,000th of seconds. Corrected for transducer depth and other such corrections
52-57	6	real	Bathymetry, corrected depth	In tenths of meters.
58-59	2	int	Bathymetric correction code	This code details the procedure used for determining the sound velocity correction to depth
60	1	int	Bathymetric type code	1 = Observed 3 = Interpolated (Header Seq. 12) 9 = Unspecified
61-66	6	real	Magnetics total field, 1 <sup>ST</sup> sensor	In tenths of nanoteslas (gammas)
67-72	6	real	Magnetics total field, 2 <sup>ND</sup> sensor	In tenths of nanoteslas (gammas), for trailing sensor
73-78	6	real	Magnetics residual field	In tenths of nanoteslas (gammas). The reference field used is in Header Seq. 13
79	1	int	Sensor for residual	1 = 1 <sup>st</sup> or leading sensor

Col	Len	Type	Contents	Description, Possible Values, Notes
			field	2 = 2 <sup>nd</sup> or trailing sensor 9 = Unspecified
80-84	5	real	Magnetics diurnal correction	In tenths of nanoteslas (gammas). (In nanoteslas) if 9-filled (i.e., set to "+9999"), total and residual fields are assumed to be uncorrected; if used, total and residuals are assumed to have been already corrected.
85-90	6	F6.0	Depth or altitude of magnetics sensor	(In meters) + = Below sea level 3 = Above sea level
91-97	7	real	Observed gravity	In 10 <sup>th</sup> of mgals. Corrected for Eotvos, drift, tares
98-103	6	real	EOTVOS correction	In 10 <sup>th</sup> of mgals. $E = 7.5 V \cos \phi \sin \alpha + 0.0042 V^*V$
104-108	5	real	Free-air anomaly	In 10 <sup>th</sup> of mgals G = observed G = theoretical
109-113	5	char	Seismic line number	Cross-reference for seismic data
114-119	6	char	Seismic shot-point number	
120	1	int	Quality code for navigation	5= Suspected, by the originating institution 6= Suspected, by the data center 9= No identifiable problem found

## Science of Opportunity

### ADCP

The shipboard ADCP system measures currents in a depth range from about 30 to 300 m -- in good weather. In bad weather or in ice, the range is reduced, and sometimes no valid measurements are made. ADCP data collection is the OPP-funded project of Eric Firing (University of Hawaii) and Teri Chereskin (Scripps Institution of Oceanography). Data is collected on both the LMG and the NBP for the benefit of scientists on individual cruises, and for the long-term goal of building a profile of current structure in the Southern Ocean.

A data feed is sent from the ADCP system to RVDAS whenever a reference layer is acquired. This feed contains east and north vectors for ship's speed, relative to the reference layer, and ship's heading. Collected files (one per day) are archived in 0812adcp.tar in the directory /rvdas/nav.

### pCO<sub>2</sub>

The NBP carries a pCO<sub>2</sub> measurement system from Lamont-Doherty Earth Observatory (LDEO). pCO<sub>2</sub> data is recorded by RVDAS and transmitted to LDEO at the end of each cruise. You will find pCO<sub>2</sub> data in a file named 0812pco2.tar in the /process directory, which contains the pCO<sub>2</sub> instrument's data merged with GPS, meteorological and other oceanographic measurements. For more information contact Colm Sweeney (csweeney@ldeo.columbia.edu).

## Cruise Science

### ***CTD***

The ctd data has been placed in the tar file /ocean/0812ctd.tar. The archive contains tar files 0812proc.tar.

### **RVDAS**

The Research Vessel Data Acquisition System (RVDAS) was developed at Lamont-Doherty Earth Observatory of Columbia University and has been in use on its research ship for many years. It has been extensively adapted for use on the USAP research vessels.

Daily data processing of the RVDAS data is performed to calibrate and convert values into useable units and as a quality-control on operation of the DAS. Raw and processed data sets from RVDAS are included in the data distribution. The tables below provide detailed information on the sensors and data. Be sure to read the “Significant Acquisition Events” section for important information about data acquisition during this cruise.

### ***Sensors and Instruments***

RVDAS data is divided into two general categories, *underway and navigation*. They can be found on the distribution media as subdirectories under the top level rvdas directory: /rvdas/uw, and /rvdas/nav. Processed oceanographic data is in the top level directory, /process. Each instrument or sensor produces a data file named with its channel ID. Each data file is g-zipped to save space on the distribution media. Not all data types are collected every day or on every cruise.

The naming convention for data files produced by the sensors and instruments is

NBP[CruiseID][ChannelID].dDDD

Example:     NBP0812.mwx1.d330

- The CruiseID is the numeric name of the cruise, in this case, NBP0812.
- The ChannelID is a 4-character code representing the system being logged. An example is “mwx1,” the designation for meteorology.
- DDD is the day of year the data was collected

## Underway Sensors

### Meteorology and Radiometry

Measurement	Channel ID	Collect. Status	Rate	Instrument
Air Temperature	mwx1	continuous	1 sec	R.M. Young 41372LC
Relative Humidity	mwx1	continuous	1 sec	R.M. Young 41372LC
Wind Speed/Direction	mwx1	continuous	1 sec	Gill 1390-PK-007
Barometer	mwx1	continuous	1 sec	R.M. Young 61201
PIR (LW radiation)	mwx1	continuous	1 sec	Eppley PIR
PSP (SW radiation)	mwx1	continuous	1 sec	Eppley PSP
PAR	mwx1	continuous	1 sec	BSI QSR-240
GUV	pguv	continuous	2 sec	BSI PUV-2511
PUV	pguv	not collected		BSI PUG-2500

### Geophysics

Measurement	Channel ID	Collect. Status	Rate	Instrument
Gravimeter	grv1	continuous	10 sec*	LaCoste & Romberg
Magnetometer	mag1	not collected	15 sec	EG&G G-866
Bathymetry	bat1	not collected	Varies	ODEC Bathy 2000
Bathymetry	knud	continuous	Varies	Knudsen 320B/R
Bathymetry	sim1	not collected	Varies	Simrad EK500 Sonar

\*Data is output every second but it only changes every 10 seconds.

### Oceanography

Measurement	Channel ID	Collect. Status	Rate	Instrument
Conductivity	mtsg	Continuous	6 sec	SeaBird SBE-45
Salinity	mtsg	Continuous	6 sec	Calc. from pri. temp
Sea Surface Temp	mtsg	Continuous	6 sec	SeaBird 3-01/S
Fluorometry	hdas	Continuous	2 sec	WET Lab AFL
Transmissometry	hdas	Continuous	2 sec	WET Lab C-Star
pCO <sub>2</sub>	Pco2	Continuous	70 sec	(LDEO)
ADCP	adcp	Continuous	varies	RD Instruments

## **Navigational Instruments**

Measurement	Channel ID	Collect. Status	Rate	Instrument
Trimble GPS	PCOD	Continuous	1 sec	Trimble 20636-00SM
Gyro	gyr1	Continuous	0.2 sec	Yokogawa Gyro
SeaPath	seap	Continuous	1 sec	SeaPath 200

## ***Data***

Data is received from the RVDAS system via RS-232 serial connections. A time tag is added at the beginning of each line of data in the form,

```
yy+dd:hh:mm:ss.sss [data stream from instrument]
```

where

yy       = two-digit year  
ddd      = day of year  
hh       = 2 digit hour of the day  
mm       = 2 digit minute  
ss.sss   = seconds

All times are reported in UTC.

The delimiters that separate fields in the raw data files are often spaces and commas but can be other characters such as : = @. Occasionally no delimiter is present. Care should be taken when reprocessing the data that the field's separations are clearly understood.

In the sections below a sample data string is shown, followed by a table that lists the data contained in the string.

## Underway Data

Each section below describes a type of data file (file name extension in parentheses) followed by a typical line of data in the file. In the table(s) for each section is a description of the fields within each line of data. Note: most data files listed below will be included with each cruise's data distribution, however some types of files may be omitted if the instrument was not operating during the cruise. The available data files can be found in the /rvdas/uw directory on the distribution disc.

### Sound Velocity Probe (svp1)

08+330:00:00:49.011 1519.35

Field	Data	Units
1	RVDAS Time tag	
2	Sound velocity in ADCP sonar well	m/s

### Meteorology (mwx1)

There are actually 3 different data strings in the mwx1 data file:

MET

08+330:23:59:57.725 MET,12.1,-54,6.64,88.7,111.3374,0.02414567,-  
0.4827508,282.9581,281.8823,1003.119

PUS

08+330:23:59:58.546 PUS,A,020,008.53,M,+337.12,+009.00,00,0F

SUS

08+330:23:59:58.779 SUS,A,017,008.76,M,+335.53,+006.35,00,02

### MET string

Field	Data	Units
1	RVDAS time tag	
2	MET (string flag)	
3	Power Supply Voltage	V
4	Enclosure Relative Humidity	%
5	Air temperature	°C
6	Air Relative Humidity	%
7	PAR (photosynthetically available radiation)*	mV
8	PSP (short wave radiation)*	mV
9	PIR Thermopile (long wave radiation)*	mV
10	PIR Case Temperature	°Kelvin
11	PIR Dome Temperature	°Kelvin
12	Barometer	mBar

\*See page 21 for calculations.

## PUS string

Field	Data	Units
1	RVDAS time tag	
2	PUS (string flag)	
3	A (unit identification)	
4	Port Wind direction relative	deg
5	Port Wind speed relative	m/s
6	Units	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Unit Status (00 or 60 are good, any other value indicates fault)	
10	Check Sum	

## SUS string

Field	Data	Units
1	RVDAS time tag	
2	SUS (string flag)	
3	A (unit identification)	
4	Starboard Wind direction relative	deg
5	Starboard Wind speed relative	m/s
6	Units	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Unit Status (00 or 60 are good, any other value indicates fault)	
10	Check Sum	

## Knudsen (knud)

99+099:00:18:19.775 HF,305.2,LF,304.3

Field	Data	Units
1	RVDAS time tag	
2	HF = High frequency flag (12 kHz)	
3	High frequency depth	meters
4	LF = Low frequency flag (3.5 kHz)	
5	Low frequency depth	meters

**Fluorometer (flr1)**

This Fluorometer is not in use. Current Fluorometer goes to the hdas string.

00+019:23:59:58.061 0 0818 :: 1/19/00 17:23:17 = 0.983 (RAW) 1.2 (C)

Field	Data	Units
1	RVDAS time tag	
2	Marker 0 to 8	
3	4-digit index	
4	Date	mm/dd/yy
5	Time	hh:mm:ss
6	Signal	
7	Signal units of measurement	
8	Cell temperature (if temperature compensation package is installed)	
9	Temperature units (if temperature compensation package is installed)	

**pCO<sub>2</sub> (pco2)**

00+021:23:59:43.190 2000021.99920 2382.4 984.2 30.73 50.8 345.9 334.1 -1.70  
-68.046 -144.446 Equil

Field	Data	Units
1	RVDAS time tag	
2	pCO <sub>2</sub> time tag (decimal is fractional time of day)	yyyyddd.ttt
3	Raw voltage (IR)	mV
4	Cell temperature	°C
5	Barometer	MBar
6	Concentration	ppm
7	Equilibrated temperature	°C
8	pCO <sub>2</sub> pressure	microAtm
9	Flow rate	ml / min
10	Source ID #	1 or 2 digits
11	Valve position	1 or 2 digits
12	Flow source (Equil = pCO <sub>2</sub> measurement)	text

**Simrad EM120 (mbdp)**

Field	Data	Units
1	LDTDS	
2	\$EMDPT	
3	Depth (corrected)	Meters



**Micro-TSG (mtsg)**

08+330:23:59:40.894 5.9322, 3.34685, 34.0550, 1473.281

Field	Data	Units
1	RVDAS time tag	
2	Temperature	°C
3	Conductivity	s/m
4	Salinity	PSU
5	Sound velocity	m/s

**Gravimeter (grv1)**

There are now two sets of fields output by the gravity meter. The data record is output once per second, and identified by "\$DAT" in the id field. A summary of sensor environmental data is output every ten seconds, identified by "%ENV" in the id field.

**Data record (\$DAT):**

05+194:00:00:27.995 \$DAT,2005/ 7/13, 0: 7: 7.36,194, 9050.37, 9050.06, 5410.86, -0.00, -0.01, -0.02, 0.00, 0.00, 0.70, 0.19, -0.12, -0.25, 0.00, -69.45711315, -54.32181487, 0.000, 285.200,

Field	Data	Conversion	Units
1	RVDAS time tag		
2	Text string (id field)	\$DAT for data record	
3	Date	YYYY/MM/DD	
4	Time	HH:MM:SS.SS	
5	Day of Year	DDD	
6	Gravity count	mgal = count x 1.0046 + offset	count
7	Spring Tension		CU
8	Beam Position	Volts x 750,000	
9	VCC		
10	AL		
11	AX		
12	VE		
13	AX2		
14	XACC2		
15	LACC2		
16	CROSS ACCEL		GAL
17	LONG ACCEL		GAL
18	EOTVOS CORR		MGAL
19	LONGITUDE		Degrees
20	LATITUDE		Degrees
21	HEADING		Degrees
22	VELOCITY		Knots

**Environmental record (\$ENV)**

05+183:19:13:10.945 %ENV,2005/ 7/ 2,19:19:52.16,183,S-036/V1.5, 3.34,  
 47.19, 20.34,1.111840E-1,-0.57700,-0.10591, 0.40180, 2.55260, 0.43000, 1,  
 300

Field	Data	Conversion	Units
1	RVDAS time tag		
2	Text string (id field)	\$ENV for environmental record	
3	Date	YYYY/MM/DD	
4	Time	HH:MM:SS.SS	
5	Day of Year	DDD	
6	Meter ID		
7	Meter Pressure		inch-Hg
8	Meter temp		°C
9	Ambient temp		°C
10	K-Factor		
11	VCC Coeff		
12	AL Coeff		
13	AX Coeff		
14	VE Coeff		
15	AX2 Coeff		
16	Serial Filter Length		Seconds
17	QC Filter Length		Seconds

**Engineering (eng1)**

08+330:23:59:50.899 12.25684 23.89813 0.4029922 0.2541656 233.4218 -  
 751.9 -8145.28 -1.386184 23.37653 23.37653 NAN

Field	Data	Units
1	RVDAS time tag	
2	Power Supply Voltage	V
3	Internal Case Temperature	°C
4	Pump #1 flow rate	L/min
5	Pump #2 flow rate	L/min
6	Pump #3 flow rate	L/min
7	Seismic air pressure	Lbs/sq-in
8	PIR case resistance (not currently hooked up, data is irrelevant)	Kohm
9	PIR case ratiometric output (not currently hooked up, data is irrelevant)	mV
10	Freezer #1 temperature	°C
11	Freezer #2 temperature	°C
12	Freezer #3 temperature	°C

\*See page 25 for PIR calculations.

**Hydro-DAS (hdas)**

08+330:23:59:41.877 12.15836 14.22853 368.9655 4060.69 -1 65.5 65.5 80  
57

Field	Data	Units
1	RVDAS time tag	
2	Supply voltage	V
3	Panel temperature	°C
4	Fluorometer	mV
5	Transmissometer	mV
6	Sea Water Valve (-1 = stern thruster valve, 0 = moon pool valve)	
7	Flow meter 1 frequency	Hz
8	Flow meter 2 frequency	Hz
9	Flow meter 3 frequency	Hz
10	Flow meter 4 frequency	Hz

**GUV Data (pguv)**

08+330:23:59:40.328 112508 235940 .000197 1.856E-1 1.116E0 4.987E-2 -  
1.959E-4 1.637E0 4.153E-3 1.76E0 42.296 17.844

Field	Data	Units
1	RVDAS time tag	
2	Date	mmddyy
3	Time (UTC)	hhmmss
4	Ed0Gnd	V
5	Ed0320	uW (cm <sup>2</sup> nm)
6	Ed0340	uW (cm <sup>2</sup> nm)
7	Ed0313	uW (cm <sup>2</sup> nm)
8	Ed0305	uW (cm <sup>2</sup> nm)
9	Ed0380	uW (cm <sup>2</sup> nm)
10	Ed0PAR	uE (cm <sup>2</sup> nm)
11	Ed0395	uW (cm <sup>2</sup> nm)
12	Ed0Temp	°C
13	Ed0Vin	V

**Remote Temperature (rtmp)**

07+272:00:00:15.960 -1.7870

Field	Data	Units
1	RVDAS time tag	
2	Temperature at seawater intake	°C

## Navigational Data

### Seapath GPS (seap)

The Seapath GPS outputs the following data strings, four in NMEA format and two in proprietary PSXN format:

- GPZDA
- GPGGA
- GPVTG
- GPHDT
- PSXN, 20
- PSXN, 22
- PSXN, 23

#### GPZDA

02+253:00:00:00.772 \$GPZDA,235947.70,09,09,2002,,\*7F

Field	Data	Units
1	RVDAS time tag	
2	\$GPZDA	
3	time	hhmmss.ss
4	Day	dd
5	Month	mm
6	Year	yyyy
7	(empty field)	
8	Checksum	

#### GPGGA

02+253:00:00:00.938

GPGGA,235947.70,6629.239059,S,06827.668899,W,1,07,1.0,11.81,M,,M,,\*6F

Field	Data	Units
1	RVDAS time tag	
2	\$GPGGA	
3	time	hhmmss.ss
4	Latitude	ddmm.mmmmmm
5	N or S for north or south latitude	
6	Longitude	ddmm.mmmmmm
7	E or W for east or west longitude	
8	GPS quality indicator, 0=invalid, 1=GPS SPS, 2=DGPS, 3=PPS, 4=RTK, 5=float RTK, 6=dead reckoning	
9	number of satellites in use (00-99)	
10	HDOP	x.x
9	height above ellipsoid in meters	m.mm
11	M	
12	(empty field)	
13	M	
14	age of DGPS corrections in seconds	s.s
15	DGPS reference station ID (0000-1023)	
16	Checksum	



## GPVTG

02+253:00:00:00.940 \$INVTG,19.96,T,,M,4.9,N,,K,A\*39

Field	Data	Units
1	RVDAS time tag	
2	\$GPVTG	
3	course over ground, degrees true	d.dd
4	T	
5	,	
6	M	
7	speed over ground in knots	k.k
8	N	
9	,	
10	K	
11	Mode	
12	Checksum	

## GPHDT

02+253:00:00:00.941 \$GPHDT,20.62,T\*23

Field	Data	Units
1	RVDAS time tag	
2	\$GPHDT	
3	Heading, degrees true	d.dd
4	T	
5	Checksum	

## PSXN,20

02+253:00:00:00.942 \$PSXN,20,0.43,0.43\*39

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	20	
4	Horizontal position & velocity quality: 0=normal, 1=reduced performance, 2=invalid data	
5	Height & vertical velocity quality: 0=normal, 1=reduced performance, 2=invalid data	
6	Heading quality: 0=normal, 1=reduced performance, 2=invalid data	
7	Roll & pitch quality: 0=normal, 1=reduced performance, 2=invalid data	
8	Checksum	

## PSXN,22

02+253:00:00:00.942 \$PSXN,22,0.43,0.43\*39

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	22	
4	gyro calibration value since system start-up in degrees	d.dd
5	short term gyro offset in degrees	d.dd
6	Checksum	



**PSXN,23**

02+253:00:00:02.933 \$PSXN,23,0.47,0.57,20.62,0.03\*0C

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	23	
4	roll in degrees, positive with port side up	d.dd
5	pitch in degrees, positive with bow up	d.dd
6	Heading, degrees true	d.dd
7	heave in meters, positive down	m.mm
8	Checksum	

**Trimble (P-Code) GPS (PCOD)**

The Trimble GPS, which formerly output Precise Position (*P-Code*) strings, but now only outputs Standard Position (*Civilian*) strings, outputs three NMEA standard data strings:

- Position fix (GGA)
- Latitude / longitude (GLL),
- Track and ground speed (VTG)

**GGA: GPS Position Fix – Geoid/Ellipsoid**

01+319:00:04:11.193 \$GPGGA,000410.312,6227.8068,S,06043.6738,W,1,06,1.0,031.9,M,-017.4,M,,\*49

Field	Data	Units
1	RVDAS Time tag	
2	\$GPGGA	
3	UTC time at position	hhmmss.sss
4	Latitude	ddmm.mmm
5	North (N) or South (S)	
6	Longitude	ddmm.mmm
7	East (E) or West (W)	
8	GPS quality: 0 = Fix not available or invalid 1 = GPS, SPS mode, fix valid 2 = DGPS (differential GPS), SPS mode, fix valid 3 = P-CODE PPS mode, fix valid	
9	Number of GPS satellites used	
10	HDOP (horizontal dilution of precision)	
11	Antenna height	meters
12	M for meters	
13	Geoidal height	meters
14	M for meters	
15	Age of differential GPS data (no data in the sample string)	
16	Differential reference station ID (no data in the sample string)	
17	Checksum (no delimiter before this field)	





**GLL: GPS Latitude/Longitude**

01+319:00:04:11.272 \$GPGLL,6227.8068,S,06043.6738,W,000410.312,A\*32

Field	Data	Units
1	RVDAS Time tag	
2	\$GPGLL	
3	Latitude	degrees
4	North or South	
5	Longitude	degrees
6	East or West	
7	UTC of position	hhmmss.sss
8	Status of data (A = valid)	
9	Checksum	

**VTG: GPS Track and Ground Speed**

01+319:00:04:11.273 \$GPVTG,138.8,T,126.0,M,000.0,N,000.0,K\*49

Field	Data	Units
1	RVDAS time tag	
2	\$GPVTG	
3	Heading	degrees
4	Degrees true (T)	
5	Heading	degrees
6	Degrees magnetic (M)	
7	Ship speed	knots
8	N = knots	
9	Speed	km/hr
10	K = km per hour	
11	Checksum	

**Gyro Compass (gyr1)**

00+019:23:59:59.952 \$HEHDT 25034,-020\*73

Field	Data	Units
1	RVDAS time tag	
2	\$HEHDT	
3	Heading, Degrees True	degrees
5	Checksum	

**ADCP Course (adcp)**

00+019:23:59:59.099 \$PUHAW,UVH,-1.48,-0.51,250.6

Field	Data	Units
1	RVDAS time tag	
2	\$PUHAW	
3	UVH (E-W, N-S, Heading)	
4	Ship Speed relative to reference layer, east vector	knots
5	Ship Speed relative to reference layer, north vector	knots
6	Ship heading	degrees

## Processed Data

The processed data sets can be found in the /process directory and subdirectories. Note: many of the subdirectories contain intermediate datasets to facilitate further processing and are not intended to be end-products. Only the final product files and datasets are described below.

### pCO<sub>2</sub>-merged

00+346:23:58:20.672 2000346.9991 2398.4 1008.4 0.01 45.4 350.3 342.6 15.77

Equil -43.6826 173.1997 15.51 33.90 0.33 5.28 9.05 1007.57 40.0 14.87 182.44

Field	Data	Units
1	RVDAS time tag	
2	pCO <sub>2</sub> time tag (decimal is fractional time of day)	yyyyddd.ttt
3	Raw voltage (IR)	mV
4	Cell temperature	°C
5	Barometer	MBar
6	Concentration	ppm
7	Equilibrated temperature	°C
8	pCO <sub>2</sub> pressure	microAtm
9	Flow rate	ml / min
10	Source ID #	1 or 2 digits
11	Valve position	1 or 2 digits
12	Flow source (Equil = pCO <sub>2</sub> measurement)	text
13	RVDAS latitude	degrees
14	RVDAS longitude	degrees
15	TSG external temperature	°C
16	TSG salinity	PSU
17	TSG fluorometry	V
18	RVDAS true wind speed	m/s
19	RVDAS true wind direction	degrees
20	Barometric Pressure	mBars
21	Uncontaminated seawater pump flow rate	l/min
22	Speed over ground	knots
23	Course made good	degrees

## Calculations

The file `instrument.coeff` located in the `/` directory contains the calibration factors for shipboard instruments. This was the file used by the RVDAS processing software.

### PAR

Coefficients `parc1` and `parcv` for this cruise can be found in the `instrument.coeff` file as the variable labeled PAR, respectively. Variable `par` is the raw data in mV, as described in the “mw1” file description. The calibration scale and probe offset dark are values taken from the PAR Cal Sheet.

```
par = raw data mV
calibration scale = 5.8644 V/(μEinstiens/cm2sec)
parc1 = 1 / scale = .17
probe offset dark = -.1 mV
parcv = dark x 1000 mV/V = -0.0001 V
((par / 1000 mV/V) - parcv) x parc1 x 10000 cm2/m2 = μEinstiens/m2sec
```

Calculations (extracted from the C code):

```
/* Convert from mV to V */
par /= 1000;
/* (par V - vdark V) / Calibration Scale Factor V/uE/cm2sec */
parCalc = (par - parcv) * parc1 * 10000;
```

### PSP

Coefficient `pspCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled PSP1. Variable `psp` is the raw data in mV, as described in the “mw1” file description.

```
psp = raw data mV
calibration scale = pspCoeff x 10^-6 V/(W/m2)
psp / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* Convert from mV to W/m^2 */
pspCalc = (psp * 1000 / pspCoeff);
```

## PIR

Coefficient `pirCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled `PIR1`. Variable `pir_thermo` is the raw data in mV, `pir_case` is the PIR case temperature in Kelvins and `pir_dome` is the PIR dome temperature in Kelvins, as described in the “`mw1`” file description. Hard-coded “C” coefficients are shown below:

```
Dome constant = 3.5
```

```
Sigma = 5.6704e-8
```

```
pir_thermo = raw data mV
```

```
calibration scale = pirCoeff x 10-6 V/(W/m2)
```

```
pir_thermo / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* convert mV to W/m^2 */  
pirCalc = (pir_thermo * 1000 / pirCoeff)  
/* correct for case temperature */  
pirCalc += sigma * pow(pir_case,4)  
/* correct for dome temperature */  
pirCalc -= 3.5 * sigma * (pow(pir_dome, 4) - pow(pir_case, 4))
```

## Acquisition Problems and Events

This section lists problems with acquisition noted during this cruise including instrument failures, data acquisition system failures and any other factor affecting this data set. The format is ddd:hh:mm (ddd is year-day, hh is hour, and mm is minute). Times are reported in GMT.

Start	End	Description
325:07:50		68° West – started loggers
326:20:05		Primary CTD Conductivity Sensor, serial # 042513 failed on CTD cast 003. Replaced with new Primary CTD Conductivity Sensor, serial # 041143.
326:23:24		Discovery (secondary DAS server) rebooted itself for no apparent reason. We restarted loggers on Challenger (primary DAS server) just to make sure all was well. Loss of data was insignificant to none.
328:15:10		Secondary CTD Conductivity Sensor, serial # 041798, was lost on CTD cast 009. Replaced with new Secondary CTD Conductivity Sensor, serial # 040924.
334:23:15		Exited Argentinean EEZ and entered international waters.
340:05:49	340:06:45	Challenger (primary DAS server) crashed. About an hour's worth of data was not recorded.
340:18:09	340:18:14	Challenger (primary DAS server) crashed again. There was 5 minutes of data not recorded.
341:08:29		Entered Argentinean EEZ
342:16:03	342:16:57	The antenna cable for the Trimble GPS (PCOD) had been jarred loose. This resulted in intermittent loss of signal.
347:19:57		68° West – stopped loggers

## Appendix: Sensors and Calibrations

### NBP0812 Shipboard Sensors:

<i>Sensor</i>	<i>Serial Number</i>	<i>Last Calibration Date</i>	<i>Comments</i>
<b>Meteorology &amp; Radiometers</b>			
Port Anemometer (Gill Ultrasonic)	836076	5/15/2007	Installed 11/17/2008
Stbd Anemometer (Gill Ultrasonic)	836077	5/15/2007	Installed 11/17/2008
Bridge Anemometer	WM 45835	2/28/2007	Bridge (center)
Barometer	01705	12/1/2006	Installed 5/26/2008
Humidity/Wet Temp	06134	9/29/2006	Installed 5/26/2008
PIR	33023F3	6/20/2008	Installed 7/9/2008
PSP	33090F3	6/11/2008	Installed 7/9/2008
Mast PAR	6356	8/8/2007	Installed 8/28/2007
GUV (Mast)	25110203113	3/18/2008	Installed 4/18/2008
<b>Underway</b>			
Micro TSG	4549120-0226	1/12/2008	Installed 6/1/2008
Fluorometer	AFL-044	5/31/2006	Installed 6/23/2007
Digital Remote Temp	3849120-0178	1/12/2008	
Remote Temp	031267	4/12/2006	Installed 1/29/2007
Transmissometer	CST-557DR	12/20/2006	Installed 3/14/08

**NBP0812 CTD Sensors:**

Sensor	Serial Number	Last Calibration Date	Comments
CTD Fish	09P7536-0328	8/1/2008	
CTD Fish Pressure	53980	8/1/2008	
CTD Deck Unit	11P47914-0768	n/a	
Primary Temperature Sensor	031238	2/21/2008	
Secondary Temperature Sensor	031649	2/21/2008	
Primary Conductivity Sensor (A)	042513	1/10/2008	Failed during Cast 003, 11/21/2008
Secondary Conductivity Sensor (A)	041798	8/01/2007	Was lost on Cast 009, 11/23/2008
Primary Conductivity Sensor (B)	041143	6/18/2008	Installed as new Primary Conductivity Sensor prior to Cast 004
Secondary Conductivity Sensor (B)	040924	7/09/2008	Installed as new Secondary Conductivity Sensor prior to cast 010
CTD Pump (Primary)	051626	4/17/2008	In House Cal
CTD Pump (Secondary)	051646	4/17/2008	In House Cal
Altimeter	42434	n/a	
Slip Ring Assembly	n/a	n/a	
Carousel Water Sampler	3211265-0066	n/a	
Pinger 12khz	5518	n/a	In House Cal

**Calibrations**

The following pages are replicas of current calibration sheets for the sensors used during this cruise.



### ***Gravity Tie Start (Punta Arenas)***

[illegible]

**Anemometer (Bridge)****RM Young Anemometer Calibration, Model 05106**

S/N: 45835

Date: 28-Feb-07

Cal'd By: George Aukon

Clockwise Cal Motor RPM	Calculated Windspeed m/s	Measured Windspeed m/s	Delta m/s	Knots
0	0.00	0.00	0.00	0.0
200	0.98	0.90	0.08	1.9
500	2.45	2.40	0.05	4.8
1000	4.90	4.80	0.10	9.5
1500	7.35	7.30	0.05	14.3
2000	9.80	9.80	0.00	19.0
3000	14.70	14.60	0.10	28.6
4000	19.60	19.50	0.10	38.1
5000	24.50	24.30	0.20	47.6
6000	29.40	29.20	0.20	57.1
7000	34.30	34.10	0.20	66.6
8000	39.20	39.00	0.20	76.2
9000	44.10	43.90	0.20	85.7
10000	49.00	48.80	0.20	95.2
12000	58.80	58.60	0.20	114.2

Direction	Measured Direction	Delta Direction
0	359	0
30	29	1
60	59	1
90	89	1
120	119	1
150	148	2
180	179	1
210	210	0
240	240	0
270	270	0
300	301	-1
330	331	-1
0	1	-1

Note: Delta direction should not  
exceed + or - 3 degrees.

Counter Clockwise Cal Motor RPM	Calculated Windspeed m/s	Measured Windspeed m/s	Delta m/s
0	0.00	0.00	0.00
200	0.98	1.00	-0.02
500	2.45	2.50	-0.05
1000	4.90	4.90	0.00
1500	7.35	7.40	-0.05
2000	9.80	9.80	0.00
3000	14.70	14.80	-0.10
4000	19.60	19.80	-0.20
5000	24.50	24.60	-0.10
6000	29.40	29.50	-0.10
7000	34.30	34.50	-0.20
8000	39.20	39.40	-0.20
9000	44.10	44.40	-0.30
10000	49.00	49.30	-0.30
12000	58.80	59.60	-0.80

Caution: Do Not exceed 12000 rpm during Wind  
Speed test.

Wind Speed Threshold < 2.9 gm? ☒ yes  
Wind Direction Threshold < 30 gm? ☒ yes

**Additional Comments**

Potentiometer and potentiometer coupling  
were replaced, vertical shaft bearings were  
cleaned and lubricated.

Note: Delta Windspeed should not exceed  
+ or - 0.3 m/s for 0 - 5000 rpm

**Barometer**

**R.M. Young Company**  
2801 Aero Park Drive  
Traverse City, Michigan 49686 USA

**Barometric Pressure Sensor Calibration Report**

Customer: *Raytheon Technical Services Co*

Test Number: 60621

Customer PO: RM31245-50

Test Date: 1 December 2006

Sales Order: 8800

**Test Sensor:**

Model: 61201

Serial Number: *BP01705*

Description: Barometric Pressure Sensor

Report of calibration comparison of test barometric pressure sensor with National Institute of Standards and Technology traceable standard pressure calibrator at five pressures in the R.M. Young Company controlled pressure facility. Calibration accuracy  $\pm 1.0$  hPa.

Reference Pressure (hPa)	Voltage Output (millivolts)	Indicated (1) Pressure (hPa)
800.0	0	800.0
875.0	1251	875.1
950.0	2501	950.1
1025.0	3750	1025.0
1100.0	4998	1099.9

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

**Reference Instrument**

Druck Pressure Controller Model DPI515  
Fluke Multimeter Model 8060A

**Serial # NIST Test Reference**

51500497 UKAS Lab 0221  
4865407 234027

Tested By: *E. Chumney*

METEOROLOGICAL INSTRUMENTS

Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: www.youngusa.com

**Humidity Sensor**

**R.M. Young Company**  
2801 Aero Park Drive  
Traverse City, Michigan 49686 USA

3-04  
**COPY**

**Relative Humidity Sensor Calibration Report**

Customer: *Raytheon Technical Services Co LLC*

Test Number: 69292R  
Test Date: 29 September 2006

Customer PO: RM29548.50  
Sales Order: 8708

**Test Sensor:**

Model: 41372LC Serial Number: 6134  
Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test relative humidity sensor with National Institute of Standards and Technology traceable standard relative humidity sensor at five humidity levels in the R.M. Young Company controlled humidity chamber facility. Calibration accuracy  $\pm 2.0$  %.

Reference Humidity (%)	Current Output (milliamps)	Indicated (1) Humidity (%)
10.4	5.9	11.9
30.4	8.8	30.2
49.7	12.0	50.0
69.9	15.2	69.8
89.6	17.9	86.6

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

**Reference Instrument**

Vaisala Humidity Sensor Model 35AC  
Fluke Multimeter Model 8060A

**Serial # NIST Test Reference**

N475040 TN 266152  
4865407 234027

Tested By: *E. Channing*

*page 2*

**METEOROLOGICAL INSTRUMENTS**

Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: www.youngusa.com

**Temperature Sensor**

**R.M. Young Company**  
2801 Aero Park Drive  
Traverse City, Michigan 49686 USA

3-04  
met

**Temperature Sensor Calibration Report**

Customer: **Raytheon Technical Services Co LLC**

Test Number: 69292

Customer PO: RM29548-50

Test Date: 29 September 2006

Sales Order: 8708

**Test Sensor:**

Model: 41372LC

Serial Number: 6134

Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test temperature sensor with National Institute of Standards and Technology traceable standard thermometers at three temperatures in the R.M. Young Company controlled temperature calibration bath facilities. Calibration accuracy  $\pm 0.1^\circ$  Celsius.

Bath Temperature (degrees C)	Current Output (milliamps)	Indicated (1) Temperature (degrees C)
-50.03	3.992	-50.05
0.03	12.003	0.02
50.03	20.004	50.03

(1) Calculated from current output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

**Reference Instrument**

Brooklyn Thermometer Model 43-FC  
Brooklyn Thermometer Model 22332-D5-FC  
Brooklyn Thermometer Model 2X400-D7-FC  
Keithley Multimeter Model 191

**Serial # NIST Test Reference**

8006-118 204365  
25071 249763  
77532 228060  
15232 234027

Tested By:

*EChenning*

page 1

**METEOROLOGICAL INSTRUMENTS**

Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: www.youngusa.com

**PIR****THE EPPLEY LABORATORY, INC.**

12 Sheffield Ave., P.O. Box 419, Newport, RI 02840 USA

Telephone: 401-847-1020

Fax: 401-847-1031

Email: info@eppleylab.com

Internet: www.eppleylab.com



Scientific Instruments  
for Precision Measurements  
Since 1917

## STANDARDIZATION OF EPPLEY PRECISION INFRARED RADIOMETER Model PIR

Serial Number: 33023F3

Resistance: 739  $\Omega$  at 23  $^{\circ}\text{C}$   
 Temperature Compensation Range:  $-20^{\circ}$  to  $+40^{\circ}$   $^{\circ}\text{C}$

This pyrgometer has been compared against Eppley's Blackbody Calibration System under radiation intensities of approximately 200 watts meter<sup>-2</sup> and an average ambient temperature of 25  $^{\circ}\text{C}$  as measured by Standard Omega Temperature Probe, RTD#1.

As a result of a series of comparisons, it has been found to have a sensitivity of:

3.90  $\times 10^{-6}$  volts/watts meter<sup>-2</sup>

The calculation of this constant is based on the fact that the relationship between radiation intensity and emf is rectilinear to intensities of 700 watts meter<sup>-2</sup>. This radiometer is linear to within  $\pm 1.0\%$  up to this intensity.

The calibration of this instrument is traceable to the International Practical Temperature Scale (IPTS) through a precision low-temperature blackbody.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".

Shipped to: Raytheon Technical Services Date of Test: June 13, 2008  
 National Science Foundation  
 Port Hueneme, CA

S.O. Number: 61667  
 Date: June 20, 2008

In Charge of Test: *R.T. Gorman*  
 Reviewed by: *Thomas J. Kuhn*

Remarks:



**PSP****THE EPPLEY LABORATORY, INC.**

12 Sheffield Ave., P.O. Box 419, Newport, RI 02840 USA

Telephone: 401-847-1020

Fax: 401-847-1031

Email: info@eppleylab.com

Internet: www.eppleylab.com

Scientific Instruments  
for Precision Measurements  
Since 1917**STANDARDIZATION OF  
EPPLEY PRECISION SPECTRAL PYRANOMETER  
Model PSP**

Serial Number: 33090F3

Resistance: 700  $\Omega$  at 23  $^{\circ}\text{C}$ Temperature Compensation Range:  $-20^{\circ}$  to  $+40^{\circ}$   $^{\circ}\text{C}$ 

This radiometer has been compared with Standard Precision Spectral Pyranometer, Serial Number 21231F3 in Eppley's Integrating Hemisphere under radiation intensities of approximately 700 watts meter<sup>-2</sup> (roughly one half a solar constant).

As a result of a series of comparisons, it has been found to have a sensitivity of:

$$8.06 \times 10^{-6} \text{ volts/watts meter}^{-2}$$


The calculation of this constant is based on the fact that the relationship between radiation intensity and emf is rectilinear to intensities of 1400 watts meter<sup>-2</sup>. This radiometer is linear to within  $\pm 0.5\%$  up to this intensity.

The calibration of this instrument is traceable to standard self-calibrating cavity pyrhemometers in terms of the Systems Internationale des Unites (SI units), which participated in the Tenth International Pyrheliometric Comparisons (IPC X) at Davos, Switzerland in September-October 2005.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".


Useful conversion facts: 1 cal cm<sup>-2</sup> min<sup>-1</sup> = 697.3 watts meter<sup>-2</sup>  
1 BTU/ft<sup>2</sup>-hr<sup>-1</sup> = 3.153 watts meter<sup>-2</sup>

Shipped to: Raytheon Technical Services Date of Test: June 11, 2008  
National Science Foundation  
Port Hueneme, CA

In Charge of Test: 

S.O. Number: 61666

Date: June 20, 2008

Reviewed by: 

Remarks:

**PAR****Biospherical Instruments Inc.****CALIBRATION CERTIFICATE**

Calibration Date 8/8/2007  
 Model Number QSR-240  
 Serial Number 6356  
 Operator TPC  
 Standard Lamp HEC-1630(10/25/2006)  
 Probe Excitation Voltage Range: 6 to 18 VDC(+)  
 Output Polarity: Positive

Probe Conditions at Calibration(In air):

Calibration Voltage: 6 VDC(+)  
 Probe Current: 1.2 mA

Probe Output Voltage:

Probe Illuminated 93.1 mV  
 Probe Dark -0.1 mV  
 Probe Net Response 93.1 mV  
 RG780 Filter 0.1 mV

Corrected Lamp Output:

Output In Air (same condition as calibration):

9.56E+15 quanta/cm<sup>2</sup>sec  
0.01588 uE/cm<sup>2</sup>sec

Calibration Scale Factor:

(To calculate irradiance, divide the net voltage reading in Volts by this value.)

Dry: 9.7383E-18 V/(quanta/cm<sup>2</sup>sec)  
5.8644E+00 V/(uE/cm<sup>2</sup>sec)

Notes:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).
3. The collector should be cleaned frequently with alcohol.
4. Calibration was performed with customer cable, when available.

QSR240R 05/24/95



**GUV**

Biospherical Instruments Inc.

GUV-2511 Calibration Certificate

System Serial Number		25110203113				Date of Calibration		3/18/08			
Calibration database		25110203113v4.mdb				Date of Certificate		3/18/2008			
DASSN		0068				Standard of Spectral Irradiance		91537(10/25/06)			
Microprocessor Tag Number		2				Operator		TC			
Monochromatic Channels		Wavelength [nm]	Responsivity [Amps per $\mu\text{W}/(\text{cm}^2\cdot\text{nm})$ ]	ScaleSmall [Volts per $\mu\text{W}/(\text{cm}^2\cdot\text{nm})$ ]	ScaleMedium [Volts per $\mu\text{W}/(\text{cm}^2\cdot\text{nm})$ ]	ScaleLarge [Volts per $\mu\text{W}/(\text{cm}^2\cdot\text{nm})$ ]	OffsetSmall [volts]	OffsetMedium [volts]	OffsetLarge [volts]	Measurement Units	
Ed0320	2	320	2.7010E-10	2.7551E-05	8.0491E-03	2.8497E+00	6.5000E-05	6.4000E-05	2.2300E-04	$\mu\text{W}/(\text{cm}^2\cdot\text{nm})$	
Ed0340	6	340	2.2027E-10	2.2467E-05	6.5640E-03	2.3144E+00	3.7400E-04	3.7600E-04	1.1590E-03	$\mu\text{W}/(\text{cm}^2\cdot\text{nm})$	
Ed0313	8	313	2.4100E-10	2.4539E-05	7.1692E-03	2.5594E+00	1.0000E-04	1.0000E-04	6.9100E-04	$\mu\text{W}/(\text{cm}^2\cdot\text{nm})$	
Ed0305	10	305	1.6600E-11	1.6943E-06	4.9499E-04	1.5313E-01	4.1300E-04	4.0500E-04	-2.2080E-03	$\mu\text{W}/(\text{cm}^2\cdot\text{nm})$	
Ed0380	12	380	7.0759E-11	7.2174E-06	2.1086E-03	7.0320E-01	1.1140E-03	1.1050E-03	-1.4780E-03	$\mu\text{W}/(\text{cm}^2\cdot\text{nm})$	
Ed0395	18	395	3.4926E-10	3.5624E-05	1.0408E-02	3.3997E+00	1.3100E-04	1.3000E-04	2.2300E-04	$\mu\text{W}/(\text{cm}^2\cdot\text{nm})$	
Broadband Channels		Wavelength [nm]	Responsivity [Amps per $\mu\text{E}/(\text{cm}^2\cdot\text{s})$ ]	ScaleSmall [Volts per $\mu\text{E}/(\text{cm}^2\cdot\text{s})$ ]	ScaleMedium [Volts per $\mu\text{E}/(\text{cm}^2\cdot\text{s})$ ]	ScaleLarge [Volts per $\mu\text{E}/(\text{cm}^2\cdot\text{s})$ ]	OffsetSmall [volts]	OffsetMedium [volts]	OffsetLarge [volts]	Measurement Units	
Ed0PAR	13	400-700	1.9205E-05	1.9589E+00	5.7230E+02	2.0376E+05	8.8600E-04	8.7700E-04	-1.3030E-03	$\mu\text{E}/(\text{cm}^2\cdot\text{sec})$	
Auxiliary Channels		Address	Wavelength	Responsivity	ScaleS	ScaleM	ScaleL	OffsetS	OffsetM	OffsetL	Measurement Units
Ed0Gnd	0	0	1	1	1	1	1	0	0	0	V
Ed0Temp	22	0	1	0.01	0.01	0.01	0.01	0	0	0	C
Ed0Vin	27	0	1	-0.25	-0.25	-0.25	-0.25	0	0	0	V

© Biospherical Instruments Inc., 5340 Riley Street, San Diego, California 92110 USA. Contact [support@biospherical.com](mailto:support@biospherical.com) for more information.

**Underway Temperature Sensor (Primary Remote)**

40

**SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0389  
CALIBRATION DATE: 06-Jun-08SBE 38 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE**ITS-90 COEFFICIENTS**

a0 = 5.665621e-005

a1 = 2.720183e-004


a2 = -2.302051e-006

a3 = 1.478988e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.50010	749675.6	-1.50007	0.00003
0.99990	667866.6	0.99986	-0.00004
4.49990	569784.5	4.49990	-0.00000
7.99990	487738.4	7.99992	0.00002
11.50000	418866.9	11.49996	-0.00004
14.99990	360860.1	14.99992	0.00002
18.49990	311841.2	18.49995	0.00005
21.99990	270286.7	21.99992	0.00002
25.49990	234949.8	25.49987	-0.00003
29.00000	204808.3	28.99992	-0.00008
32.49990	179023.6	32.49995	0.00005

Temperature ITS-90 =  $1/[a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]] - 273.15$  (°C)

Residual = instrument temperature - bath temperature



**Underway Temperature Sensor (Secondary Remote)****SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1267  
CALIBRATION DATE: 12-Apr-06SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

$g = 4.76625066e-003$   
 $h = 6.64522185e-004$   
 $i = 2.84261863e-005$   
 $j = 2.62601374e-006$   
 $f_0 = 1000.0$

## ITS-68 COEFFICIENTS


$a = 3.68121498e-003$   
 $b = 5.89543636e-004$   
 $c = 1.47312260e-005$   
 $d = 2.62748536e-006$   
 $f_0 = 5707.029$

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	5707.029	-1.5002	-0.00010
1.0000	6042.027	1.0001	0.00005
4.4999	6534.652	4.5000	0.00014
7.9999	7055.608	8.0000	0.00007
11.4999	7605.659	11.4998	-0.00008
14.9999	8185.578	14.9998	-0.00014
18.4999	8796.088	18.4998	-0.00011
22.0000	9437.895	22.0000	-0.00003
25.4999	10111.624	25.5001	0.00019
28.9999	10817.898	29.0001	0.00019
32.4999	11557.257	32.4997	-0.00019

$$\text{Temperature ITS-90} = 1 / \{ g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)] \} - 273.15 \text{ (}^\circ\text{C)}$$

$$\text{Temperature ITS-68} = 1 / \{ a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)] \} - 273.15 \text{ (}^\circ\text{C)}$$
Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35  $^\circ\text{C}$ )

Residual = instrument temperature - bath temperature

**POST CRUISE  
CALIBRATION**

## Micro-TSG Calibration Files

### Temperature Sensor (Micro-TSG)

63

#### SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0226  
 CALIBRATION DATE: 12-Jan-08

SBE 45 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE


#### ITS-90 COEFFICIENTS

a0 = -6.625307e-005  
 a1 = 2.870318e-004  
 a2 = -3.279891e-006  
 a3 = 1.750784e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	744351.4	1.0000	-0.0000
4.4999	636501.4	4.5000	0.0001
15.0000	405773.8	14.9998	-0.0002
18.5000	351383.7	18.5001	0.0001
24.0000	281927.0	24.0001	0.0001
29.0000	232171.7	28.9998	-0.0002
32.5000	203331.2	32.5001	0.0001

Temperature ITS-90 =  $1/[a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]] - 273.15$  (°C)

Residual = instrument temperature - bath temperature



**Conductivity Sensor (Micro-TSG)**

64

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0226  
 CALIBRATION DATE: 12-Jan-08

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -1.014919e+000  
 h = 1.569437e-001  
 i = -4.469846e-004  
 j = 6.064779e-005

CPcor = -9.5700e-008  
 CTcor = 3.2500e-006  
 WBOTC = 9.8072e-007


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2549.02	0.00000	0.00000
1.0000	34.9254	2.98440	5059.48	2.98442	0.00002
4.4999	34.9056	3.29231	5250.05	3.29229	-0.00001
15.0000	34.8618	4.27660	5816.78	4.27659	-0.00001
18.5000	34.8524	4.62264	6003.02	4.62263	-0.00001
24.0000	34.8419	5.18201	6292.11	5.18204	0.00003
29.0000	34.8360	5.70518	6550.56	5.70519	0.00001
32.5000	34.8334	6.07863	6728.76	6.07862	-0.00001

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = instrument conductivity - bath conductivity



**Underway fluorometer (Primary)**

PO Box 518  
620 Applegate St.  
Philomath OR 97370



(541) 929-5650  
Fax (541) 929-5277  
<http://www.wetlabs.com>

**Chlorophyll Fluorometer Characterization**

**Date:** 05/31/06  
**Serial #:** AFL-044  
**Job#:** 0209007  
**Tech:** K.C

**Dark Counts** 0.177 volts  
**CEV** 2.725 volts  
**SF** 9.8116

**FSV** 5.45 volts

**Linearity:** 0.999 R<sup>2</sup> (0–1.5 volts)  
0.995 R<sup>2</sup> (0– 5.45 volts)

**Notes:**

**Dark Counts:** Signal output of the meter in clean water with black tape over detector.

**CEV** is the chlorophyll equivalent voltage. This value is the signal output of the fluorometer when using a fluorescent proxy that has been determined to be approximately equivalent to **25 µg/l** of a *Thalassiosira weissflogii* phytoplankton culture.

**SF** is the scale factor used to derive chlorophyll concentration from the signal voltage output of the fluorometer. The scale factor is determined by using the following equation:  
 $SF = (25) / (CEV - \text{dark})$  e.g.  $(25) / (2.865 - 0.238) = 9.516$

**FSV** is the maximum signal voltage output that the fluorometer is capable of.

Chlorophyll concentration expressed in µg/l (mg/m<sup>3</sup>) can be derived by using the following equation: (µg/l) = (V<sub>measured</sub> – CWO) \* SF

The relationship between fluorescence and chlorophyll-*a* concentrations in-situ is highly variable. The scale factor listed on this document was determined by using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer you must perform secondary measurements on the populations of interest. This is typically done using extraction based measurement techniques on discrete samples. For additional information on determination of chlorophyll concentration see [ Standard Methods For The Examination Of Water And Wastewater] part 10200 H published jointly by: American Public Health Association, American Water Works Association and Water Environment Federation.

**Underway Transmissometer**

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

**C-Star Calibration and Repairs**

Date	12/20/2006	Customer	Raytheon Polar Services
Job #	021020	S/N#	CST-557DR
		Work order	007

**Repairs and Modifications:**

Replaced bulkhead connector, 38 mm lens and o-rings. Baked to remove moisture. Recalibrated.

**Comments:**

- Shake-tested unit
- Pressure-tested unit
- Noise test: 1 sample/sec for 60 sec
- Stability test: 1 sample/min for 12 hrs
- Performed water calibration
- Temperature test, 27–2 °C
- Updated unit's calibration sheet

cstarwkbkf1.xls

Revision F

6/12/03



**CTD Pressure Sensor (fish)****SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0328  
CALIBRATION DATE: 01-Aug-07SBE9plus PRESSURE CALIBRATION DATA  
10000 psia S/N 53980

## DIGIQUARTZ COEFFICIENTS:

C1 = -5.847002e+004  
 C2 = 6.910390e-001  
 C3 = 1.753360e-002  
 D1 = 4.241600e-002  
 D2 = 0.000000e+000  
 T1 = 3.026040e+001  
 T2 = -1.938830e-004  
 T3 = 4.330190e-006  
 T4 = 2.020250e-009  
 T5 = 0.000000e+000


## AD590M, AD590B, SLOPE AND OFFSET:

AD590M = 1.13300e-002  
 AD590B = -8.47592e+000  
 Slope = 0.99999  
 Offset = -0.8347 (dbars)

PRESSURE (PSIA)	INST OUTPUT(Hz)	INST TEMP(C)	INST OUTPUT (PSIA)	CORRECTED INST OUTPUT (PSIA)	RESIDUAL (PSIA)
14.647	33053.30	22.4	15.793	14.583	-0.064
2001.665	33609.50	22.7	2002.748	2001.525	-0.140
3988.943	34155.20	22.7	3990.098	3988.862	-0.081
5976.002	34690.70	22.8	5977.067	5975.820	-0.182
7962.994	35216.70	22.8	7964.504	7963.244	0.250
9950.680	35733.40	22.9	9951.508	9950.236	-0.444
7962.873	35216.70	23.0	7964.485	7963.225	0.352
5975.724	34690.70	23.0	5977.040	5975.792	0.068
3988.622	34155.20	23.1	3990.073	3988.837	0.215
2001.487	33609.50	23.2	2002.738	2001.515	0.028
14.639	33053.30	23.9	15.848	14.638	-0.001

Residual = corrected instrument pressure - reference pressure

Date, Avg Offset (ps:





**CTD Primary Temperature Sensor**

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643-9866 Fax (425) 643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1238  
 CALIBRATION DATE: 21-Feb-08

SBE3 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

**ITS-90 COEFFICIENTS**

g = 4.82476162e-003  
 h = 6.70818317e-004  
 i = 2.57308739e-005  
 j = 2.05952356e-006  
 f0 = 1000.0

**IPTS-68 COEFFICIENTS**

a = 3.68121354e-003  
 b = 5.97988390e-004  
 c = 1.45594709e-005  
 d = 2.06092428e-006  
 f0 = 6125.349


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	6125.349	-1.5001	0.00001
0.9999	6479.620	0.9999	-0.00003
4.4999	7000.092	4.4999	-0.00000
7.9999	7549.851	7.9999	0.00005
11.4999	8129.653	11.4999	-0.00000
14.9999	8740.270	14.9999	-0.00003
18.4999	9382.440	18.4999	0.00001
21.9999	10056.839	21.9999	-0.00003
25.4999	10764.175	25.4999	0.00001
28.9999	11505.081	29.0000	0.00006
32.4999	12280.148	32.4999	-0.00003

Temperature ITS-90 =  $1/[g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]] - 273.15$  (°C)

Temperature IPTS-68 =  $1/[a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]] - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature



**CTD Secondary Temperature Sensor**

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1649  
 CALIBRATION DATE: 21-Feb-08

SBE3 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

**ITS-90 COEFFICIENTS**

g = 4.80668457e-003  
 h = 6.64493310e-004  
 i = 2.15692409e-005  
 j = 1.43428317e-006  
 f0 = 1000.0

**IPTS-68 COEFFICIENTS**

a = 3.68120751e-003  
 b = 6.01350146e-004  
 c = 1.39161708e-005  
 d = 1.43556304e-006  
 f0 = 5958.871


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	5958.871	-1.4996	0.00045
0.9999	6301.380	0.9994	-0.00048
4.4999	6804.480	4.4994	-0.00045
7.9999	7335.658	8.0000	0.00012
11.4999	7895.510	11.5002	0.00030
14.9999	8484.823	15.0004	0.00053
18.4999	9104.147	18.4999	-0.00002
21.9999	9754.322	21.9995	-0.00038
25.4999	10436.063	25.4996	-0.00026
28.9999	11149.939	28.9999	-0.00004
32.4999	11896.546	32.5001	0.00023

Temperature ITS-90 =  $1/[g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]] - 273.15$  (°C)

Temperature IPTS-68 =  $1/[a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]] - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature



**CTD Primary Conductivity Sensor** (Casts 001 through 003)

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643-9866 Fax (425) 643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2513  
 CALIBRATION DATE: 10-Jan-08

SBE4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**GHJ COEFFICIENTS**

g = -1.05743446e+001  
 h = 1.62917155e+000  
 i = -5.67710821e-004  
 j = 1.56305556e-004  
 CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

**ABCDM COEFFICIENTS**

a = 5.36180356e-005  
 b = 1.62799472e+000  
 c = -1.05724932e+001  
 d = -8.43534708e-005  
 m = 4.3  
 CPcor = -9.5700e-008 (nominal)


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.54801	0.00000	0.00000
-1.0000	34.9569	2.81484	4.87385	2.81482	-0.00002
1.0000	34.9569	2.98683	4.98079	2.98686	0.00002
15.0000	34.9578	4.28713	5.72442	4.28712	-0.00000
18.5000	34.9569	4.63500	5.90741	4.63502	0.00001
29.0001	34.9510	5.72190	6.44539	5.72187	-0.00003
32.5000	34.9401	6.09513	6.61997	6.09515	0.00002

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients



**CTD Primary Conductivity Sensor** (From cast 004 on)

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1143  
 CALIBRATION DATE: 18-Jun-08

SBE4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**GHIJ COEFFICIENTS**

g = -3.93882237e+000  
 h = 5.02987900e-001  
 i = 1.11349944e-005  
 j = 2.56523028e-005  
 CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

**ABCDM COEFFICIENTS**

a = 3.40762494e-005  
 b = 5.02941077e-001  
 c = -3.93840704e+000  
 d = -7.94651103e-005  
 m = 3.9  
 CPcor = -9.5700e-008 (nominal)


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.79772	0.00000	0.00000
-1.0001	34.9126	2.81160	7.96942	2.81159	-0.00001
0.9999	34.9127	2.98341	8.17952	2.98342	0.00001
14.9999	34.9139	4.28231	9.61850	4.28227	-0.00003
18.4999	34.9136	4.62987	9.96789	4.62990	0.00003
28.9999	34.9108	5.71604	10.98685	5.71604	0.00000
32.4999	34.9039	6.08952	11.31558	6.08952	-0.00000

Conductivity =  $(g + hf^2 + if^3 + jf^4) / (10(1 + \delta t + \epsilon p))$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients





**CTD Secondary Conductivity Sensor** (Casts 001 through 009)**SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1798  
CALIBRATION DATE: 10-Jan-08SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHJ COEFFICIENTS

$g = -4.12050701e-000$   
 $h = 4.81308717e-001$   
 $i = -4.84959069e-004$   
 $j = 4.99382645e-005$   
 $CPcor = -9.5700e-008$  (nominal)  
 $CTcor = 3.2500e-006$  (nominal)

## ABCDM COEFFICIENTS


$a = 2.22551267e-006$   
 $b = 4.79814915e-001$   
 $c = -4.11470326e+000$   
 $d = -8.17173015e-005$   
 $m = 4.9$   
 $CPcor = -9.5700e-008$  (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.92894	0.00000	0.00000
-1.0000	34.9569	2.81484	8.19336	2.81487	0.00003
1.0000	34.9569	2.98683	8.40822	2.98684	0.00000
15.0000	34.9578	4.28713	9.88021	4.28702	-0.00011
18.5000	34.9569	4.63503	10.23762	4.63503	0.00003
29.0001	34.9510	5.72190	11.27937	5.72206	0.00016
32.5000	34.9401	6.09513	11.61461	6.09502	-0.00011

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (10(1 + \delta t + \epsilon p)) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure[decibars]}; \delta = CTcor; \epsilon = CPcor;$$

$$\text{Residual} = (\text{instrument conductivity} - \text{bath conductivity}) \text{ using } g, h, i, j \text{ coefficients}$$
**POST CRUISE  
CALIBRATION**

**CTD Secondary Conductivity Sensor** (From cast 010 on)

**SEA-BIRD ELECTRONICS, INC.**  
 1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0924  
 CALIBRATION DATE: 09-Jul-08

SBE4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**GHIJ COEFFICIENTS**

g = -3.99907272e+000  
 h = 5.35284626e-001  
 i = -6.25325544e-004  
 j = 6.12685156e-005  
 CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

**ABCDM COEFFICIENTS**

a = 1.18553196e-006  
 b = 5.33195976e-001  
 c = -3.99302487e+000  
 d = -9.02082056e-005  
 m = 5.2  
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.73650	0.00000	0.00000
-0.9253	34.9561	2.82114	7.76564	2.82114	0.00000
1.0564	34.9561	2.99168	7.96809	2.99169	0.00002
14.9999	34.9577	4.28711	9.36181	4.28706	-0.00005
18.4999	34.9578	4.63510	9.70156	4.63510	0.00000
29.0000	34.9557	5.72257	10.69214	5.72266	0.00009
32.4999	34.9487	6.09645	11.01138	6.09639	-0.00006

Conductivity =  $(g + hf^2 + if^3 + jf^4) / (10(1 + \delta t + \epsilon p))$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

