

```
#####
# LMG calibration data file for sensors
#
# NOTE:
# 1. In order for these calibrations to take affect, uwint and rv_tsg must
# be restarted. (Remember, rv_tsg has parameters.)
#
# 2. Please enter serial numbers for all sensors
#
# 3. Remember, when you check this file back into RCS, use the
# -u option. It MUST remain in /usr/local/packages/rvdas/config
#
# 4. The TSG calibration coefficients must be placed last in this file.
#
#
#####
# Ship - LMG or NBP
SHIP LMG
#
#####
# Cruise ID (i.e. LMG0810)
cruiseID LMG0811
#
#####
# LM Gould radiometer calibrations
# PSP ser#:31701F3 cal date: 22 May, 2008
# PIR ser#:32031F3 cal date: 22 May. 2008
# Instrument      uVolts/W/m^2
PSP    8.40
PIR    3.88
#
#####
# PAR Instrument Vdark Calib_Factor (ser#:6393, cal date: 2/7/2008
#instrument, Probe Dark(V), Calib Factor (Dry) (V/uE/cm^2sec)
PAR    .0008  5.9891
#
#####
# Transmisometer (ser#: CST-891DR, cal date: 20Nov06)
# Vdark Vref Path
TRAN  0.057  4.829  0.25
#
#
#####
# LMG winches
#
# Scale conversion information for the science winches on the LMG.
# Sheave measurements made on 01/01/00.
# Wire Pull tests done on dates indicated
#
# Dush 4 winch    sheave diam=
# 9/16" wire      wire diam =
#           total circumference=
```

```

#           magnets      =
#       Payout Scale factor=
#       Tension Scale Factor=
#           operation limit=   lb
#
# Dush xx winch    sheave diam= 28.125   .714m
# .680" wire      wire diam = 0.680   .017m
#           total circumference= 90.493" 2.297m
#           magnets      = 24
#       Payout Scale factor= 3.77     0.096m
#       Tension Scale Factor= 180
#           operation limit= 20,150 lb
#
#
# meters out = mout * a
# speed = speed * c
# tension = (tension * b) - e
# operation limit = d
#     a     b     c     d     e
LDU4  1     0.465  1   20718  0
LDU5  1     1       1   20150  0
LD11  1     1       1   5980   0
LWN1  1     1       1   5980   0
#SWNC -0.1   200   1.67  20718 -800
#PWNC  0.1    180   1.67  20150  0
#BWNC  0.1    62.5  1.67  5980   437.5
#WWNC -0.1    60    -1.67  5980   0
#
#
#####
##### Note, TSG calibrations must be last in this file #####
##### Do not change the formating, only the values. Thanks #####
#####
## **** Calibration factors for SBE 21 S/N 3208 ****
**** Calibration Date of 30-Jun-05 2007 ****
# currently in use
# Temperture calibration factors
%TEMPERTURE%
#g 0.00413343557
#h 0.000615389618
#i 0.0000197232448
#j 0.00000135946639
#fo 1000.000
#*
#
# conductivity calibration factors
%CONDUCTIVITY%
#g -3.99168762
#h 0.471887572
#i -0.00055051605
#j 0.0000513922485
#p -0.0000000957
#t 0.00000325

```

```
/*
***** Remote Temperature Probe SN # 4015 *****
***** Calibration Date of 11-May-07 *****
# external temperature calibration factors
#%EXTERNAL TEMPERATURE%
#g 0.0043665187
#h 0.000627057431
#i 0.0000215854061
#j 0.00000173345987
#fo 1000.000
#*
#
#
#
```