

Data Report NBP0902

Weddell Sea

Dr. Ken Smith

March 6 – April 15, 2009



RVIB Nathaniel B. Palmer
United States Antarctic Program
Raytheon Polar Services Corporation
Prepared by Kris Merrill

Table of Contents

INTRODUCTION.....	1
DISTRIBUTION CONTENTS AT A GLANCE	2
EXTRACTING DATA.....	3
DISTRIBUTION CONTENTS	4
CRUISE INFORMATION.....	4
<i>Cruise Track</i>	4
<i>Satellite Images</i>	4
NBP DATA PRODUCTS.....	4
<i>JGOFS</i>	5
<i>MGD77</i>	6
SCIENCE OF OPPORTUNITY.....	7
<i>ADCP</i>	7
<i>pCO₂</i>	7
CRUISE SCIENCE	8
<i>CTD</i>	8
<i>XBT</i>	8
<i>Mocness</i>	8
<i>Nutrients</i>	8
RVDAS	9
<i>Sensors and Instruments</i>	9
Underway Sensors	10
Meteorology and Radiometry.....	10
Geophysics	10
Oceanography.....	10
Navigational Instruments	11
<i>Data</i>	11
Underway Data	12
Sound Velocity Probe (svp1).....	12
Meteorology (mwx1).....	12
MET string	12
PUS string	13
SUS string	13
Bathy 2000 (bat1).....	13
Knudsen (knud).....	15
Fluorometer (flr1).....	15
pCO ₂ (peo2).....	15
Simrad EM120 (mbdp).....	16
Micro-TSG (mtsg).....	16
Gravimeter (grv1).....	16
Engineering (eng1).....	17
Hydro-DAS (hdas)	18
GUV Data (pguv).....	18
Remote Temperature (rtmp)	18
Navigational Data	19
Seapath GPS (seap)	19
Trimble (P-Code) GPS (PCOD)	21
Gyro Compass (gyr1)	22
ADCP Course (adcp)	22
Processed Data.....	23
pCO ₂ -merged	23
CALCULATIONS.....	24
PAR	24
PSP	24
PIR.....	25

ACQUISITION PROBLEMS AND EVENTS.....	26
APPENDIX: SENSORS AND CALIBRATIONS.....	27
<i>NBP0902 Shipboard Sensors</i>	27
Meteorology & Radiometers	27
Underway	27
<i>NBP0902 CTD Sensors</i>	28
<i>NBP0902 MOCNESS Sensors</i>	29
MOCNESS 1 Meter.....	29
MOCNESS 10 Meter.....	29
CALIBRATIONS	29
<i>Gravity Tie Start (Punta Arenas)</i>	30
<i>Anemometer (Bridge)</i>	31
<i>Barometer.....</i>	32
<i>Humidity Sensor</i>	33
<i>Temperature Sensor</i>	34
<i>PIR.....</i>	35
<i>PSP</i>	36
<i>PAR</i>	37
<i>GUV.....</i>	38
MICRO-TSG CALIBRATION FILES	39
<i>Temperature Sensor (Micro-TSG) #0226.....</i>	39
<i>Conductivity Sensor (Micro-TSG) #0226.....</i>	40
<i>Temperature Sensor (Micro-TSG) #0242.....</i>	41
<i>Conductivity Sensor (Micro-TSG) #0242</i>	42
<i>Underway Temperature Sensor (Primary Remote)</i>	43
<i>Underway Temperature Sensor (Secondary Remote)</i>	44
<i>Underway fluorometer (Primary).....</i>	45
<i>Underway Transmissometer.....</i>	46
<i>CTD Pressure Sensor (Fish)</i>	47
<i>CTD Primary Pump.....</i>	49
<i>CTD Secondary Pump</i>	49
<i>CTD Primary Conductivity Sensor.....</i>	50
<i>CTD Primary Conductivity Sensor B</i>	51
<i>CTD Secondary Conductivity Sensor</i>	52
<i>CTD Primary Temperature Sensor.....</i>	53
<i>CTD Secondary Temperature Sensor.....</i>	54
<i>CTD Primary Dissolved Oxygen Sensor</i>	55
<i>CTD Secondary Dissolved Oxygen Sensor.....</i>	56
<i>CTD Fluorometer.....</i>	57
<i>CTD Transmissometer.....</i>	58
<i>CTD PAR</i>	59
MOCNESS	60
<i>Mocness 1 Temperature Sensor.....</i>	60
<i>Mocness 10 Temperature Sensor.....</i>	61
<i>Mocness 1 Conductivity Sensor.....</i>	62
<i>Mocness 10 Conductivity Sensor.....</i>	63
<i>Mocness Fluorometer Sensor.....</i>	64
<i>Mocness Transmissometer Sensor.....</i>	65
<i>Mocness Depth Sensor</i>	66

Introduction

The NBP data acquisition systems continuously log data from the instruments used during the cruise. This document describes:

- The structure and organization of the data on the distribution media
- The format and contents of the data strings
- Formulas for calculating values
- Information about the specific instruments in use during the cruise
- A log of acquisition problems and events during the cruise that may affect the data
- Scanned calibration sheets for the instruments in use during the cruise.

The data is distributed on a DVD-R or CD-ROM written in ISO9660 level-1 format. It is readable by virtually every computing platform.

All the data has been compressed using Unix “gzip,” identified by the “.gz” extension. It has been copied to the distribution media in the Unix tar archive format, “.tar” extension. Tools are available on all platforms for uncompressing and de-archiving these formats: On Macintosh, use Stuffit Expander with DropStuff. On Windows operating systems use WinZip.

MultiBeam, BathyW data, and raw ADCP data are distributed separately.

IMPORTANT: Read the last section, “Acquisition Problems and Events,” for important information that may affect the processing of this data.

Distribution Contents at a Glance

Volume 1 of 1: NBP0902

File	Description
/	Root level directory NBP0902.trk Text file of cruise track (lat,lon) NBP0902.mgd Full Cruise MGD77 data file NBP0902.gmt GMT binary file of MGD77 data INSTCOEF.TXT Instrument Coefficient File 0902DATA.doc Data Report NBP0902 (MS word) 0902DATA.pdf Data Report NBP0902 (pdf) Schedule.doc Daily cruise schdule
/plots	Cruise track plots 0902_trk.ps Cruise track plot (PostScript format) 0902_trk.jpg Cruise track plot (JPEG format)
/process	Processed data 0902jgof.tar JGOFS format data files 0902qc.tar Daily RVDAS QC postscript plots 0902pco2.tar Merged pCO2 data files 0902mgd.tar MGD Data 0902proc.tar Other processed data
/rvdas/nav	Navigation data 0902gyr1.tar Gyro raw data 0902PCOD.tar Trimble P-code raw data 0902seap.tar Seapath data 0902adcp.tar ADCP Data Sets
/rvdas/uw	Underway data 0902ctdd.tar CTD Depth 0902eng1.tar Engineering Data 0902grv1.tar Gravimeter raw data 0902hdas.tar HydroDAS raw data 0902knud.tar Knudsen raw data 0902mtsg.tar Micro TSG data 0902mwx1.tar Meteorology raw data 0902pco2.tar pCO2 raw data 0902pguv.tar GUV raw data 0902rtmp.tar Remote temperature data 0902svp1.tar Sound velocity probe (in ADCP well)
/Imagery	Other data, files and pictures 0902lImag.tar Satellite imagery
/ocean	Ocean data 0902ctd.tar CTD data 0902xbt.tar XBT data
/MOCNESS	MOCNESS data mocness.tar MOCNESS data
/adcps	ADCP data adcp.tar ADCP data

Extracting Data

The Unix tar command has many options. It is often useful to know exactly how an archive was produced when expanding its contents. All archives were created using the command,

```
tar cvf archive_filename files_to_archive
```

To create a list of the files in the archive, use the Unix command,

```
tar tvf archive_filename > contents.list
```

where contents.list is the name of the file to create

To extract the files from the archive:

```
tar xvf archive_filename file(s)_to_extract
```

G-zipped files will have a “.gz” extension on the filename. These files can be decompressed after de-archiving, using the Unix command,

```
gunzip filename.gz
```

Distribution Contents

Cruise Information

NBP0902 departed Punta Arenas, Chile on March 6, 2009 and spent roughly six weeks studying large icebergs in the Weddell Sea. The aim of the project was to study the importance of free drifting icebergs on natural fertilization, pelagic community response and organic carbon export. We returned to Punta Arenas on April 15, 2009.

Cruise Track

The distribution DVD includes a GMT cruise track file (NBP0902.trk). It contains the longitude and latitude of the ship's position at one-minute intervals extracted from the NBP0902.gmt file.

JPEG and PostScript cruise track files have been produced and placed in the /plots directory.

Satellite Images

Satellite Images received for this cruise can be found in the file called /Imagery/0902Imag.tar collected and processed on the ship is in two further subdirectories, Ice and WX (weather). Files are named using the convention, ssss_fff_mmddyy_tttt_ww.gif where:

ssss_fff = satellite and flight number
mm = month
dd = day
yy = year
tttt = time in hours and minutes (UTC)
ww = optional field for identifying wavelength, such as vis (visible) or IR (infrared)

NBP Data Products

Two datasets are created on each cruise: JGOFS and MGD77.

The data processing scripts used to produce JGOFS and MGD77 data sets create a lot of intermediate files. These files are included on the data distribution media in a file called /process/0902proc.tar. They are included to make re-processing easier in the event of an error, but no extensive detail of the formats is included in this document. If you have any questions, please contact itvessel@usap.gov.

JGOFS

The JGOFS data set can be found on the distribution media in the file /process/0902jgof.tar. The archive contains one file produced for each day named jgDDD.dat.gz, where DDD is the year-day the data was acquired. The ".gz" extension indicates that the individual files are compressed before archiving. Each daily file consists of 22 columnar fields in text format as described in the table below. The JGOFS data set is created from calibrated data decimated at one-minute intervals. Several fields are derived measurements from more than a single raw input. For example, Course Made Good (CMG) and Speed Over Ground (SOG) are calculated from gyro and GPS inputs. Daily plots during the cruise are produced from the JGOFS data set. Note: Null, unused, or unknown fields are indicated as "NAN" as 9999 in the JGOFS data.

Field	Data	Units
01	GMT date	dd/mm/yy
02	GMT time	hh:mm:ss
03	NGL latitude (negative is South)	tt.tttt
04	NGL longitude (negative is West)	ggg.gggg
05	Speed over ground	Knots
06	GPS HDOP	-
07	Gyro Heading	Degrees (azimuth)
08	Course made good	Degrees (azimuth)
09	Mast PAR	μ Einstens/meter ² sec
10	Sea surface temperature	°C
11	Sea surface conductivity	siemens/meter
12	Sea surface salinity	PSU
13	Sea depth (uncorrected, calc. sw sound vel. 1500 m/s)	meters
14	True wind speed (max speed windbird)	meters/sec
15	True wind direction (max speed windbird)	degrees (azimuth)
16	Ambient air temperature	°C
17	Relative humidity	%
18	Barometric pressure	mBars
19	Sea surface fluorometry	μ g/l (mg/m ³)
20	Transmissometry	%
21	PSP	W/m ²
22	PIR	W/m ²

MGD77

The MGD77 data set is contained in a single file for the entire cruise. It can be found in the top level of the distribution data structure as NBP0902.mgd. The file NBP0902.gmt is created from the MGD77 dataset using the "mgd77togmt" utility. NBP0902.gmt can be used with the GMT plotting package.

The data used to produce the NBP0902.mgd file can be found on the distribution media in the file /process/0902proc.tar. The data files in the archive contain a day's data and follow the naming convention Dddd.fnl.gz, where ddd is the year-day. These files follow a space-delimited columnar format that may be more accessible for some purposes. They contain data at one-second intervals rather than one minute and are individually "gzipped" to save space. Below is a detailed description of the MGD77 data set format. The other files in the archive contain interim processing files and are included to simplify possible reprocessing of the data using the RVDAS NBP processing scripts.

All decimal points are implied. Leading zeros and blanks are equivalent. Unknown or unused fields are filled with 9's. All "corrections", such as time zone, diurnal magnetics, and EOTVOS, are understood to be added.

Col	Len	Type	Contents	Description, Possible Values, Notes
1	1	Int	Data record type	Set to "5" for data record
2-9	8	Char	Survey identifier	
10-12	3	int	Time zone correction	Corrects time (in characters 13-27) to GMT when added; 0 = GMT
13-16	4	int	Year	4 digit year
17-18	2	int	Month	2 digit month
19-20	2	int	Day	2 digit day
21-22	2	int	Hour	2 digit hour
23-27	5	real	Minutes x 1000	
28-35	8	real	Latitude x 100000	+ = North - = South. (-9000000 to 9000000)
36-44	9	real	Longitude x 100000	+ = East - = West. (-18000000 to 18000000)
45	1	int	Position type code	1=Observed fix 3=Interpolated 9=Unspecified
46-51	6	real	Bathymetry, 2- way travel time	In 10,000th of seconds. Corrected for transducer depth and other such corrections
52-57	6	real	Bathymetry, corrected depth	In tenths of meters.
58-59	2	int	Bathymetric correction code	This code details the procedure used for determining the sound velocity correction to depth
60	1	int	Bathymetric type code	1 = Observed 3 = Interpolated (Header Seq. 12) 9 = Unspecified
61-66	6	real	Magnetics total field, 1 st sensor	In tenths of nanoteslas (gammas)
67-72	6	real	Magnetics total field, 2 nd sensor	In tenths of nanoteslas (gammas), for trailing sensor
73-78	6	real	Magnetics residual field	In tenths of nanoteslas (gammas). The reference field used is in Header Seq. 13
79	1	int	Sensor for residual field	1 = 1 st or leading sensor 2 = 2 nd or trailing sensor 9 = Unspecified

Col	Len	Type	Contents	Description, Possible Values, Notes
80-84	5	real	Magnetics diurnal correction	In tenths of nanoteslas (gammas). (In nanoteslas) if 9-filled (i.e., set to “+9999”), total and residual fields are assumed to be uncorrected; if used, total and residuals are assumed to have been already corrected.
85-90	6	F6.0	Depth or altitude of magnetics sensor	(In meters) + = Below sea level 3 = Above sea level
91-97	7	real	Observed gravity	In 10^{-1} of mgals. Corrected for Eotvos, drift, tares
98-103	6	real	EOTVOS correction	In 10^{-1} of mgals. $E = 7.5 V \cos \phi \sin \alpha + 0.0042 V^*V$
104-108	5	real	Free-air anomaly	In 10^{-1} of mgals G = observed G = theoretical
109-113	5	char	Seismic line number	Cross-reference for seismic data
114-119	6	char	Seismic shot-point number	
120	1	int	Quality code for navigation	5=Suspected, by the originating institution 6=Suspected, by the data center 9=No identifiable problem found

Science of Opportunity

ADCP

The shipboard ADCP system measures currents in a depth range from about 30 to 300 m -- in good weather. In bad weather or in ice, the range is reduced, and sometimes no valid measurements are made. ADCP data collection is the OPP-funded project of Eric Firing (University of Hawaii) and Teri Chereskin (Scripps Institution of Oceanography). Data is collected on both the LMG and the NBP for the benefit of scientists on individual cruises, and for the long-term goal of building a profile of current structure in the Southern Ocean.

A data feed is sent from the ADCP system to RVDAS whenever a reference layer is acquired. This feed contains east and north vectors for ship's speed, relative to the reference layer, and ship's heading. Collected files (one per day) are archived in 0902adcp.tar in the directory /rvdas/nav.

pCO₂

The NBP carries a pCO₂ measurement system from Lamont-Doherty Earth Observatory (LDEO). pCO₂ data is recorded by RVDAS and transmitted to LDEO at the end of each cruise. You will find pCO₂ data in a file named 0902pco2.tar in the /process directory, which contains the pCO₂ instrument's data merged with GPS, meteorological and other oceanographic measurements. For more information contact Colm Sweeney (csweeney@ldeo.columbia.edu).

Cruise Science

CTD

Data acquired from conductivity, temperature and depth (CTD) casts has been placed in the tar file /ocean/0902ctd.tar. The archive contains ".gz" files of raw, processed and sv data with the file name beginning

nbp0902XXX.yyy

where XXX represents the cast number, and yyy is the data format.

XBT

Data acquired from expendable bathymetric temperature probes has been placed in the tar file /ocean/0902xbt.tar. The tar archive contains ".gz" files of the output files from the probe in the format

T7_000XX.EDF

where the first two digits indicate the probe type, the string after the underscore indicates the cast number, and the extension indicates the data format.

Mocness

Multiple Opening Closing Net Environmental Sensing System (MOCNESS). The mocness is a system of nets that can be opened and closed remotely from the deck. It is used to catch small marine organisms. All of the sensors on the Mocness are the same Seabird and Wetlabs sensors used on the CTD.

Nutrients

A summary report of the nutrient analysis from NBP0902 is provided in an excel spreadsheet in /Nutrient/0806Nuts.tar. This spreadsheet summarized the data found in the .OMN data file from the Lachat Quickchem 8000 Nutrient Analyzer, which are also provided in the archive /Nutrient/0806OMN.tar .

RVDAS

The Research Vessel Data Acquisition System (RVDAS) was developed at Lamont-Doherty Earth Observatory of Columbia University and has been in use on its research ship for many years. It has been extensively adapted for use on the USAP research vessels.

Daily data processing of the RVDAS data is performed to calibrate and convert values into useable units and as a quality-control on operation of the DAS. Raw and processed data sets from RVDAS are included in the data distribution. The tables below provide detailed information on the sensors and data. Be sure to read the “Significant Acquisition Events” section for important information about data acquisition during this cruise.

Sensors and Instruments

RVDAS data is divided into two general categories, *underway and navigation*. They can be found on the distribution media as subdirectories under the top level rvdas directory: /rvdas/uw, and /rvdas/nav. Processed oceanographic data is in the top level directory, /process. Each instrument or sensor produces a data file named with its channel ID. Each data file is g-zipped to save space on the distribution media. Not all data types are collected every day or on every cruise.

The naming convention for data files produced by the sensors and instruments is

NBP[CruiseID][ChannelID].dDDD

Example: NBP0902.mwx1.d330

- The CruiseID is the numeric name of the cruise, in this case, NBP0902.
- The ChannelID is a 4-character code representing the system being logged. An example is “mwx1,” the designation for meteorology.
- DDD is the day of year the data was collected

Underway Sensors

Meteorology and Radiometry

Measurement	Channel ID	Collect. Status	Rate	Instrument
Air Temperature	mwx1	continuous	1 sec	R.M. Young 41372LC
Relative Humidity	mwx1	continuous	1 sec	R.M. Young 41372LC
Wind Speed/Direction	mwx1	continuous	1 sec	Gill 1390-PK-007
Barometer	mwx1	continuous	1 sec	R.M. Young 61201
PIR (LW radiation)	mwx1	continuous	1 sec	Eppley PIR
PSP (SW radiation)	mwx1	continuous	1 sec	Eppley PSP
PAR	mwx1	continuous	1 sec	BSI QSR-240
GUV	pguv	continuous	2 sec	BSI PUV-2511
PUV	pguv	not collected		BSI PUG-2500

Geophysics

Measurement	Channel ID	Collect. Status	Rate	Instrument
Gravimeter	grv1	continuous	10 sec*	LaCoste & Romberg
Magnetometer	mag1	not collected	15 sec	EG&G G-866
Bathymetry	bat1	continuous**	Varies	ODEC Bathy 2000
Bathymetry	knud	continuous	Varies	Knudsen 320B/R
Bathymetry	sim1	not collected	Varies	Simrad EK500 Sonar

*Data is output every second but it only changes every 10 seconds.

**Primary underway bathymetry was switched from Knudsen to Bathy 2000 at 016:16:33.

Oceanography

Measurement	Channel ID	Collect. Status	Rate	Instrument
Conductivity	mtsg	Continuous	6 sec	SeaBird SBE-45
Salinity	mtsg	Continuous	6 sec	Calc. from pri. temp
Sea Surface Temp	mtsg	Continuous	6 sec	SeaBird 3-01/S
Fluorometry	hdas	Continuous	2 sec	WET Lab AFL
Transmissometry	hdas	Continuous	2 sec	WET Lab C-Star
pCO ₂	Pco2	Continuous	70 sec	(LDEO)
ADCP	adcp	Continuous	varies	RD Instruments

Navigational Instruments

Measurement	Channel ID	Collect. Status	Rate	Instrument
Trimble GPS	PCOD	Continuous	1 sec	Trimble 20636-00SM
Gyro	gyr1	Continuous	0.2 sec	Yokogawa Gyro
SeaPath	seap	Continuous	1 sec	SeaPath 200

Data

Data is received from the RVDAS system via RS-232 serial connections. A time tag is added at the beginning of each line of data in the form,

```
yy+dd:hh:mm:ss.sss [data stream from instrument]
```

where

yy	= two-digit year
dd	= day of year
hh	= 2 digit hour of the day
mm	= 2 digit minute
ss.sss	= seconds

All times are reported in UTC.

The delimiters that separate fields in the raw data files are often spaces and commas but can be other characters such as : = @. Occasionally no delimiter is present. Care should be taken when reprocessing the data that the field's separations are clearly understood.

In the sections below a sample data string is shown, followed by a table that lists the data contained in the string.

Underway Data

Each section below describes a type of data file (file name extension in parentheses) followed by a typical line of data in the file. In the table(s) for each section is a description of the fields within each line of data. Note: most data files listed below will be included with each cruise's data distribution, however some types of files may be omitted if the instrument was not operating during the cruise. The available data files can be found in the /rvdas/uw directory on the distribution disc.

Sound Velocity Probe (svp1)

08+330:00:00:49.011 1519.35

Field	Data	Units
1	RVDAS Time tag	
2	Sound velocity in ADCP sonar well	m/s

Meteorology (mwx1)

There are 3 different data strings in the mwx1 data file:

MET

08+330:23:59:57.725 MET,12.1,-54,6.64,88.7,111.3374,0.02414567,-
0.4827508,282.9581,281.8823,1003.119

PUS

08+330:23:59:58.546 PUS,A,020,008.53,M,+337.12,+009.00,00,0F

SUS

08+330:23:59:58.779 SUS,A,017,008.76,M,+335.53,+006.35,00,02

MET string

Field	Data	Units
1	RVDAS time tag	
2	MET (string flag)	
3	Power Supply Voltage	V
4	Enclosure Relative Humidity	%
5	Air temperature	°C
6	Air Relative Humidity	%
7	PAR (photosynthetically available radiation)*	mV
8	PSP (short wave radiation)*	mV
9	PIR Thermopile (long wave radiation)*	mV
10	PIR Case Temperature	°Kelvin
11	PIR Dome Temperature	°Kelvin
12	Barometer	mBar

*See page 21 for calculations.

PUS string

Field	Data	Units
1	RVDAS time tag	
2	PUS (string flag)	
3	A (unit identification)	
4	Port Wind direction relative	deg
5	Port Wind speed relative	m/s
6	Units	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Unit Status (00 or 60 are good, any other value indicates fault)	
10	Check Sum	

SUS string

Field	Data	Units
1	RVDAS time tag	
2	SUS (string flag)	
3	A (unit identification)	
4	Starboard Wind direction relative	deg
5	Starboard Wind speed relative	m/s
6	Units	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Unit Status (00 or 60 are good, any other value indicates fault)	
10	Check Sum	

Bathy 2000 (bat1)

00+019:23:59:53.901 ;I04485.3ME -23.0, I00000.0,-99.9,0000@01/11/00,
23:59:52.08 PW2 PF1 SF1 PL3 MO4 SB3 PO0 TX1 TR: GM5 1500 06.7 -72.1

Field	Data	Format / Possible Values	Units
1	RVDAS time tag		
2	Flagged low frequency chn. depth w/ units	;FDDDDDD.Dun where F = flag (V for valid, I for invalid), D=depth, un = units	meters
3	Low Frequency echo strength	EEE.EE	dB
4	Flagged high freq. chn. depth	not used	
5	High frequency echo strength	not used	
6	Signed heave data	SHHHH	cm
7	Date	mm/dd/yy	
8	Time	hh:mm:ss	
9	Transmit pulse window type	PW1=Rectangular PW2=Hamming PW3=Cosine PW4=Blackman	
10	Primary transmit frequency	PF1=3.5 kHz PF2=12.0 kHz	kHz
11	Parametric mode secondary	SF1=3.5 kHz	kHz

Field	Data	Format / Possible Values	Units
	frequency	SF2=12.0 kHz	
12	Pulse length	PL1=200usec PL2=500usec PL3=1msec PL4=2msec PL5=5msec PL6=10msec PL7=25msec If transmit mode is FM: PL1=25msec PL2=50msec PL3=100msec	
13	Operating mode	MO1=CW parametric MO2=CW MO3=FM parametric MO4=FM	
14	Frequency sweep bandwidth	SB1=1 kHz SB2=2 kHz SB3=5 kHz	kHz
15	Power level	PO1 = 0dB PO2 = -6dB PO3 = -12dB PO4 = -18dB PO5 = -24dB PO6 = -30dB PO6 = -30 dB PO7 = -36dB PO8 = -42dB	
16	Transmit mode	TX1=single ping active TX2=pinger listen TX3=multipinging TR TX4=multipinging TR TX5=multipinging TTRR TX6=multipinging TTTTRRRR TX7=multipinging TTTTTRRRRR	
17	Transmit Rate	TR3 = 4Hz TR4 = 2Hz TR5 = 1Hz TR6 = .5Hz TR7 = .33Hz TR8 = .25Hz TR9 = .20Hz TR: = .10Hz TR; = .05Hz	Hz
18	System gain mode	GM0=hydrographic AGC GM1 to GM9=hydrographic +3db to + 27db manual. GMA to GMD=hydrographic + 30db through + 60db manual GME to GMK=sub-bottom 1 through sub-bottom 7	
19	Speed of sound		m/sec
20	Depth of sonar window below sea-level		meters
21	Background noise level in fixed point reference		dB/V

Knudsen (knud)

99+099:00:18:19.775 HF,305.2,LF,304.3

Field	Data	Units
1	RVDAS time tag	
2	HF = High frequency flag (12 kHz)	
3	High frequency depth	meters
4	LF = Low frequency flag (3.5 kHz)	
5	Low frequency depth	meters

Fluorometer (flr1)

This Fluorometer is not in use. Current Fluorometer goes to the hdas string.

00+019:23:59:58.061 0 0818 :: 1/19/00 17:23:17 = 0.983 (RAW) 1.2 (C)

Field	Data	Units
1	RVDAS time tag	
2	Marker 0 to 8	
3	4-digit index	
4	Date	mm/dd/yy
5	Time	hh:mm:ss
6	Signal	
7	Signal units of measurement	
8	Cell temperature (if temperature compensation package is installed)	
9	Temperature units (if temperature compensation package is installed)	

pCO₂ (pco2)00+021:23:59:43.190 2000021.99920 2382.4 984.2 30.73 50.8 345.9 334.1 -1.70
-68.046 -144.446 Equil

Field	Data	Units
1	RVDAS time tag	
2	pCO ₂ time tag (decimal is fractional time of day)	yyyyddd.ttt
3	Raw voltage (IR)	mV
4	Cell temperature	°C
5	Barometer	MBar
6	Concentration	ppm
7	Equilibrated temperature	°C
8	pCO ₂ pressure	microAtm
9	Flow rate	ml / min
10	Source ID #	1 or 2 digits
11	Valve position	1 or 2 digits
12	Flow source (Equil = pCO ₂ measurement)	text

Simrad EM120 (mbdp)

Field	Data	Units
1	LDTDS	
2	\$EMDPT	
3	Center Beam Depth (corrected)	Meters

Micro-TSG (mtsg)

08+330:23:59:40.894 5.9322, 3.34685, 34.0550, 1473.281

Field	Data	Units
1	RVDAS time tag	
2	Internal Temperature	°C
3	Conductivity	s/m
4	Salinity	PSU
5	Sound velocity	m/s

Gravimeter (grv1)

There are now two sets of fields output by the gravity meter. The data record is output once per second, and identified by "\$DAT" in the id field. A summary of sensor environmental data is output every ten seconds, identified by "%ENV" in the id field.

Data record (\$DAT):

05+194:00:00:27.995 \$DAT,2005/7/13,0:7:7.36,194,9050.37,9050.06, 5410.86, -0.00, -0.01, -0.02, 0.00, 0.00, 0.70, 0.19, -0.12, -0.25, 0.00,-69.45711315,-54.32181487, 0.000, 285.200,

Field	Data	Conversion	Units
1	RVDAS time tag		
2	Text string (id field)	\$DAT for data record	
3	Date	YYYY/MM/DD	
4	Time	HH:MM:SS.SS	
5	Day of Year	DDD	
6	Gravity count	mgal = count x 1.0046 + offset	count
7	Spring Tension		CU
8	Beam Position	Volts x 750,000	
9	VCC		
10	AL		
11	AX		
12	VE		
13	AX2		
14	XACC2		
15	LACC2		
16	CROSS ACCEL		GAL
17	LONG ACCEL		GAL
18	EOTVOS CORR		MGAL

Field	Data	Conversion	Units
19	LONGITUDE		Degrees
20	LATITUDE		Degrees
21	HEADING		Degrees
22	VELOCITY		Knots

Environmental record (\$ENV)

05+183:19:13:10.945 %ENV, 2005/ 7/ 2, 19:19:52.16, 183, S-036/V1.5, 3.34,
 47.19, 20.34, 1.111840E-1, -0.57700, -0.10591, 0.40180, 2.55260, 0.43000, 1,
 300

Field	Data	Conversion	Units
1	RVDAS time tag		
2	Text string (id field)	\$ENV for environmental record	
3	Date	YYYY/MM/DD	
4	Time	HH:MM:SS.SS	
5	Day of Year	DDD	
6	Meter ID		
7	Meter Pressure		inch-Hg
8	Meter temp		°C
9	Ambient temp		°C
10	K-Factor		
11	VCC Coeff		
12	AL Coeff		
13	AX Coeff		
14	VE Coeff		
15	AX2 Coeff		
16	Serial Filter Length		Seconds
17	QC Filter Length		Seconds

Engineering (eng1)

08+330:23:59:50.899 12.25684 23.89813 0.4029922 0.2541656 233.4218 -
 751.9 -8145.28 -1.386184 23.37653 23.37653 NAN

Field	Data	Units
1	RVDAS time tag	
2	Power Supply Voltage	V
3	Internal Case Temperature	°C
4	Pump #1 flow rate	L/min
5	Pump #2 flow rate	L/min
6	Pump #3 flow rate	L/min
7	Seismic air pressure	Lbs/sq-in
8	PIR case resistance (not currently hooked up, data is irrelevant)	Kohm
9	PIR case ratiometric output (not currently hooked up, data is irrelevant)	mV
10	Freezer #1 temperature	°C
11	Freezer #2 temperature	°C
12	Freezer #3 temperature	°C

*See page 25 for PIR calculations.

Hydro-DAS (hdas)

08+330:23:59:41.877 12.15836 14.22853 368.9655 4060.69 -1 65.5 65.5 80
57

Field	Data	Units
1	RVDAS time tag	
2	Supply voltage	V
3	Panel temperature	°C
4	Fluorometer	mV
5	Transmissometer	mV
6	Sea Water Valve (-1 = stern thruster valve, 0 = moon pool valve)	
7	Flow meter 1 frequency	Hz
8	Flow meter 2 frequency	Hz
9	Flow meter 3 frequency	Hz
10	Flow meter 4 frequency	Hz

GUV Data (pguv)

08+330:23:59:40.328 112508 235940 .000197 1.856E-1 1.116E0 4.987E-2 -
1.959E-4 1.637E0 4.153E-3 1.76E0 42.296 17.844

Field	Data	Units
1	RVDAS time tag	
2	Date	mmddyy
3	Time (UTC)	hhmmss
4	Ed0Gnd	V
5	Ed0320	uW (cm^2 nm)
6	Ed0340	uW (cm^2 nm)
7	Ed0313	uW (cm^2 nm)
8	Ed0305	uW (cm^2 nm)
9	Ed0380	uW (cm^2 nm)
10	Ed0PAR	uE (cm^2 nm)
11	Ed0395	uW (cm^2 nm)
12	Ed0Temp	°C
13	Ed0Vin	V

Remote Temperature (rtmp)

07+272:00:00:15.960 -1.7870

Field	Data	Units
1	RVDAS time tag	
2	Temperature at seawater intake	°C

Navigational Data

Seapath GPS (seap)

The Seapath GPS outputs the following data strings, four in NMEA format and two in proprietary PSXN format:

- GPZDA
- GPGGA
- GPVTG
- GPHDT
- PSXN, 20
- PSXN, 22
- PSXN, 23

GPZDA

02+253:00:00:00.772 \$GPZDA,235947.70,09,09,2002,,*7F

Field	Data	Units
1	RVDAS time tag	
2	\$GPZDA	
3	time	hhmmss.ss
4	Day	dd
5	Month	mm
6	Year	yyyy
7	(empty field)	
8	Checksum	

GPGGA

02+253:00:00:00.938

GPGGA,235947.70,6629.239059,S,06827.668899,W,1,07,1.0,11.81,M,,M,,*6F

Field	Data	Units
1	RVDAS time tag	
2	\$GPGGA	
3	time	hhmmss.ss
4	Latitude	ddmm.mmeeeeeee
5	N or S for north or south latitude	
6	Longitude	ddmm.mmeeeeeee
7	E or W for east or west longitude	
8	GPS quality indicator, 0=invalid, 1=GPS SPS, 2=DGPS, 3=PPS, 4=RTK, 5=float RTK, 6=dead reckoning	
9	number of satellites in use (00-99)	
10	HDOP	x.x
9	height above ellipsoid in meters	m.mm
11	M	
12	(empty field)	
13	M	
14	age of DGPS corrections in seconds	s.s
15	DGPS reference station ID (0000-1023)	
16	Checksum	

GPVTG

02+253:00:00:00.940 \$INVTG,19.96,T,,M,4.9,N,,K,A*39

Field	Data	Units
1	RVDAS time tag	
2	\$GPVTG	
3	course over ground, degrees true	d.dd
4	T	
5	,	
6	M	
7	speed over ground in knots	k.k
8	N	
9	,	
10	K	
11	Mode	
12	Checksum	

GPHDT

02+253:00:00:00.941 \$GPHDT,20.62,T*23

Field	Data	Units
1	RVDAS time tag	
2	\$GPHDT	
3	Heading, degrees true	d.dd
4	T	
5	Checksum	

PSXN,20

02+253:00:00:00.942 \$PSXN,20,0.43,0.43*39

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	20	
4	Horizontal position & velocity quality: 0=normal, 1=reduced performance, 2=invalid data	
5	Height & vertical velocity quality: 0=normal, 1=reduced performance, 2=invalid data	
6	Heading quality: 0=normal, 1=reduced performance, 2=invalid data	
7	Roll & pitch quality: 0=normal, 1=reduced performance, 2=invalid data	
8	Checksum	

PSXN,22

02+253:00:00:00.942 \$PSXN,22,0.43,0.43*39

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	22	
4	gyro calibration value since system start-up in degrees	d.dd
5	short term gyro offset in degrees	d.dd
6	Checksum	

PSXN,23

02+253:00:00:02.933 \$PSXN,23,0.47,0.57,20.62,0.03*0C

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	23	
4	roll in degrees, positive with port side up	d.dd
5	pitch in degrees, positive with bow up	d.dd
6	Heading, degrees true	d.dd
7	heave in meters, positive down	m.mm
8	Checksum	

Trimble (P-Code) GPS (PCOD)

The Trimble GPS, which formerly output Precise Position (*P*-Code) strings, but now only outputs Standard Position (*Civilian*) strings, outputs three NMEA standard data strings:

- Position fix (GGA)
- Latitude / longitude (GLL),
- Track and ground speed (VTG)

GGA: GPS Position Fix – Geoid/Ellipsoid

01+319:00:04:11.193 \$GPGGA,000410.312,6227.8068,S,06043.6738,W,1,06,1.0,
031.9,M,-017.4,M,,*49

Field	Data	Units
1	RVDAS Time tag	
2	\$GPGGA	
3	UTC time at position	hhmmss.sss
4	Latitude	ddmm.mmm
5	North (N) or South (S)	
6	Longitude	ddmm.mmm
7	East (E) or West (W)	
8	GPS quality: 0 = Fix not available or invalid 1 = GPS, SPS mode, fix valid 2 = DGPS (differential GPS), SPS mode, fix valid 3 = P-CODE PPS mode, fix valid	
9	Number of GPS satellites used	
10	HDOP (horizontal dilution of precision)	
11	Antenna height	meters
12	M for meters	
13	Geoidal height	meters
14	M for meters	
15	Age of differential GPS data (no data in the sample string)	
16	Differential reference station ID (no data in the sample string)	
17	Checksum (no delimiter before this field)	

GLL: GPS Latitude/Longitude

01+319:00:04:11.272 \$GPGLL,6227.8068,S,06043.6738,W,000410.312,A *32

Field	Data	Units
1	RVDAS Time tag	
2	\$GPGLL	
3	Latitude	degrees
4	North or South	
5	Longitude	degrees
6	East or West	
7	UTC of position	hhmmss.sss
8	Status of data (A = valid)	
9	Checksum	

VTG: GPS Track and Ground Speed

01+319:00:04:11.273 \$GPVTG,138.8,T,126.0,M,000.0,N,000.0,K *49

Field	Data	Units
1	RVDAS time tag	
2	\$GPVTG	
3	Heading	degrees
4	Degrees true (T)	
5	Heading	degrees
6	Degrees magnetic (M)	
7	Ship speed	knots
8	N = knots	
9	Speed	km/hr
10	K = km per hour	
11	Checksum	

Gyro Compass (gyr1)

00+019:23:59:59.952 \$HEHDT 25034,-020 *73

Field	Data	Units
1	RVDAS time tag	
2	\$HEHDT	
3	Heading, Degrees True	degrees
5	Checksum	

ADCP Course (adcp)

00+019:23:59:59.099 \$PUHAW,UVH,-1.48,-0.51,250.6

Field	Data	Units
1	RVDAS time tag	
2	\$PUHAW	
3	UVH (E-W, N-S, Heading)	
4	Ship Speed relative to reference layer, east vector	knots
5	Ship Speed relative to reference layer, north vector	knots
6	Ship heading	degrees

Processed Data

The processed data sets can be found in the /process directory and subdirectories. Note: many of the subdirectories contain intermediate datasets to facilitate further processing and are not intended to be end-products. Only the final product files and datasets are described below.

pCO₂-merged

```
00+346:23:58:20.672 2000346.9991 2398.4 1008.4 0.01 45.4 350.3 342.6 15.77
Equil -43.6826 173.1997 15.51 33.90 0.33 5.28 9.05 1007.57 40.0 14.87 182.44 -1
```

Field	Data	Units
1	RVDAS time tag	
2	pCO ₂ time tag (decimal is fractional time of day)	yyyyddd.ttt
3	Raw voltage (IR)	mV
4	Cell temperature	°C
5	Barometer	MBar
6	Concentration	ppm
7	Equilibrated temperature	°C
8	pCO ₂ pressure	microAtm
9	Flow rate	ml / min
10	Source ID #	1 or 2 digits
11	Valve position	1 or 2 digits
12	Flow source (Equil = pCO ₂ measurement)	text
13	RVDAS latitude	degrees
14	RVDAS longitude	degrees
15	TSG external temperature	°C
16	TSG salinity	PSU
17	TSG fluorometry	V
18	RVDAS true wind speed	m/s
19	RVDAS true wind direction	degrees
20	Barometric Pressure	mBars
21	Uncontaminated seawater pump flow rate	l/min
22	Speed over ground	knots
23	Course made good	degrees
24	Input Source	-1 stem Thurston; 0 moonpool

Calculations

The file `instrument.coeff` located in the `/` directory contains the calibration factors for shipboard instruments. This was the file used by the RVDAS processing software.

PAR

Coefficients `parc1` and `parcv` for this cruise can be found in the `instrument.coeff` file as the variable labeled PAR, respectively. Variable `par` is the raw data in mV, as described in the "mwx1" file description. The calibration scale and probe offset dark are values taken from the PAR Cal Sheet.

```
par = raw data mV
calibration scale = 5.8644 V/(μEinstiens/cm2sec)
parc1 = 1 / scale = .17
probe offset dark = -.1 mV
parcv = dark x 1000 mV/V = -0.0001 V
((par / 1000 mV/V) - parcv) x parc1 x 10000 cm2/m2 = μEinstiens/m2sec
```

Calculations (extracted from the C code):

```
/* Convert from mV to V */
par /= 1000;
/* (par V - vdark V) / Calibration Scale Factor V/uE/cm2sec */
parCalc = (par - parcv) * parc1 * 10000;
```

PSP

Coefficient `pspCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled PSP1. Variable `psp` is the raw data in mV, as described in the "mwx1" file description.

```
psp = raw data mV
calibration scale = pspCoeff x 10-6 V/(W/m2)
psp / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* Convert from mV to W/m2 */
pspCalc = (psp * 1000 / pspCoeff);
```

PIR

Coefficient pirCoeff for this cruise can be found in the instrument.coeff file as the variable labeled PIR1. Variable pir_thermo is the raw data in mV, pir_case is the PIR case temperature in Kelvins and pir_dome is the PIR dome temperature in Kelvins, as described in the "mwx1" file description. Hard-coded "C" coefficients are shown below:

Dome constant = 3.5

Sigma = 5.6704e-8

```
pir_thermo = raw data mV
calibration scale = pirCoeff x 10^-6 V/(W/m^2)
pir_thermo / (scale x 1000 mV/V) = W/m^2
```

Calculations (extracted from the C code):

```
/* convert mV to W/m^2 */
pirCalc = (pir_thermo * 1000 / pirCoeff)
/* correct for case temperature */
pirCalc += sigma * pow(pir_case, 4)
/* correct for dome temperature */
pirCalc -= 3.5 * sigma * (pow(pir_dome, 4) - pow(pir_case, 4))
```

Acquisition Problems and Events

This section lists problems with acquisition noted during this cruise including instrument failures, data acquisition system failures and any other factor affecting this data set. The format is ddd:hh:mm (ddd is year-day, hh is hour, and mm is minute). Times are reported in GMT.

Start	End	Description
066:15:00		Departed Chilean EEZ -52.7626 -67.9974 and started loggers.
067:09:23		Departed Argentinean EEZ at -57.1018 -61.1639.
075:08:28 075:12:16 075:15:07	075:09:15 075:12:23 075:15:29	ECO turned off water well. Sea water flow affected. PCO2 data gap between 075:15:34 and 075:15:42.
077:12:40 077:15:58	077:13:12 077:16:56	Water well turned off for cleaning. Sea water flow affected.
079:12:57	079:14:07	Water well turned off for cleaning. Sea water flow affected.
078:23:12 078:23:23	078:12:40 078:23:26	Turned off MicroTSG to clean filters.
079:11:59	079:12:00	Restarted Knudsen.
091:17:10	091:17:19	Replaced Micro TSG.
094:13:10	094:13:43	ECO changed valves for saltwater system.
094:15:45	094:15:54	MET data interrupted due to mast power supply maintenance.
095:17:02	096:17:09	Unknown PCO2 data gap.
100:12:13	100:12:36	ADCP stopped for Simrad testing.
101:11:59		End of cruise data distribution – stopped loggers.

Appendix: Sensors and Calibrations

NBP0902 Shipboard Sensors

Sensor	Serial Number	Last Calibration Date	Comments
Meteorology & Radiometers			
Port Anemometer (Gill Ultrasonic)	836076	5/15/2007	Installed 11/17/2008
Stbd Anemometer (Gill Ultrasonic)	836077	5/15/2007	Installed 11/17/2008
Bridge Anemometer	WM 45835	2/28/2007	Bridge (center)
Barometer	0872	9/26/2006	Installed 12/13/2008
Humidity/Wet Temp	06733	9/26/2007	Installed 12/13/2008
PIR	33023F3	6/20/2008	Installed 7/9/2008
PSP	33090F3	6/11/2008	Installed 7/9/2008
Mast PAR	6356	8/8/2007	Installed 8/28/2007
GUV (Mast)	25110203113	3/18/2008	Installed 3/18/2008
PRR-800/810 System	8000803115	6/20/2007	Installed at beginning of cruise
Underway			
Micro TSG	4549120-0226	1/12/2008	Installed 6/1/2008
Micro TSG	4550449-0242	1/12/2008	Installed 3/31/2009
Remote Temp (primary)	3849120-0178	1/12/2008	Installed 7/1/2008
Fluorometer	AFLD-011	10/29/2008	Installed 1/1/2009
Transmissometer	CST-557DR	11/14/2007	Installed 12/13/008

NBP0902 CTD Sensors

Sensor	Serial Number	Last Calibration Date	Comments
CTD Fish	097536-0328	08/01/07	
CTD Fish Pressure	53980	08/01/07	
CTD Deck Unit	11P19858-0768	N/A	
Slip-Ring Assembly	1.406	N/A	
Carousel Water Sampler	3211265-0066	N/A	
Pump	051626	4/17/08	
Pump (secondary)	051646	4/17/08	
Conductivity (primary)	041314	9/18/08	Reading high on Casts 033, 034
Conductivity (primary "B")	041431	9/30/08	Installed for cast 035
Conductivity (secondary)	041852	9/18/08	
Temperature (primary)	03P2168	07/09/08	
Temperature (secondary)	03P2367	08/12/08	
Dissolved Oxygen	0082	09/12/08	
Dissolved Oxygen (sec)	0152	09/19/08	
Fluorometer	FLRTD-0397	12/18/08	
Transmissometer	CST-831DR	4/9/08	
PAR	4471	07/28/08	
Altimeter	42434	N/A	
Bottom Contact Switch	#1	N/A	

NBP0902 MOCNESS Sensors

Sensor	Serial Number	Last Calibration Date	Comments
MOCNESS 1 Meter			
Electronics Unit	139	01/28/2001	
Conductivity	042067	09/23/08	
Temperature	031497	03/01/07	
Fluorometer	FLRTD-855	12/18/08	
Transmissometer	CST-889DR	07/27/07	
MOCNESS 10 Meter			
Electronics Unit	149	02/24/2001	Failed during testing & cast 6.
Electronics Unit	139	01/28/2001	Used as primary for most casts.
Conductivity	042069	09/23/08	
Temperature	034080	06/11/05	
Fluorometer	FLRTD-855	12/18/08	
Transmissometer	CST-439DR	07/27/07	

Calibrations

The following pages are replicas of current calibration sheets for the sensors used during this cruise.

Gravity Tie Start (Punta Arenas)**Gravity Tie Spreadsheet**

The fields outlined in **BOLD** MUST BE FILLED IN for this spreadsheet to operate properly.
The automatically calculated values show up in the shaded fields.

Date: **1/3/2009**Location: **Punta Arenas, Chile**

Station: Harbour Admin. Bldg.

Latitude: 53 09 S

Longitude: 070 55 W

Elevation:

Gravity: 981320.82

Reference Code Numbers:

Station no. 9337-50

ISGN no. 51230N

Ship's meter before gravity tie (Gravity (cu))
 Ship's meter after gravity tie (Gravity (cu))
 Average
 Ship Gravimeter's Calibration Constant
Corrected ship's meter (QC Grav (mgal))

Value	Time (GMT)
8974.4	17:37
8974.6	18:38
8974.5	
1.0046	
9015.8	

Ship's meter before gravity tie (serial, RVDAS)
 Ship's meter after gravity tie (serial, RVDAS)
 Average (for comparison check only)

Value	Time (GMT)
8974.5	17:37
8974.6	18:38
8974.6	

Portable Gravimeter Interval Factor

1.01007 From Table 1 of Model G #807 Meter

Station**Value** **Time (GMT)** **Temp** **Date**

Pier measurement 1

4913.36	17:51	53	January 3, 2009	OBS mgal, averaged
4913.33	17:53	53	January 3, 2009	4962.82
4913.35	17:54	53	January 3, 2009	
4913.35				

Pier measurement 2

Pier measurement 3

Average

Station measurement 1

Station measurement 2

Station measurement 3

Average

Pier measurement 4

Pier measurement 5

Pier measurement 6

Average

4914.18	17:51	53	January 3, 2009	OBS mgal, averaged
4914.14	17:53	53	January 3, 2009	4963.64
4914.14	17:53	53	January 3, 2009	
4914.15				

4913.40	18:30	53	January 3, 2009	OBS mgal, averaged
4913.45	18:31	53	January 3, 2009	4962.91
4913.45	18:31	53	January 3, 2009	
4913.43				

Gravity offset from last tie**Drift since last tie**

972304.81

-0.50

OBS Differences

Station to Pier (1, 2, & 3 averaged)

Station to Pier (4, 5, & 6 averaged)

Averaged Differences

Gravity at pier

Elevation of pier above gravimeter, meters

Earth differential gravity, mgal/meter

Gravity at ship's gravimeter

Gravity Offset (for RVDAS)

Comments	
-0.81	
-0.73	
-0.77	
981320.05	
0.1	
0.3	
981320.09	
972304.31	

Gravity Tie performed by Greg Watson at pier in Punta Arenas and at port admin building.

Anemometer (Bridge)**RM Young Anemometer Calibration, Model 05106**

S/N: 45835

Date: 28-Feb-07

Cal'd By: George Aukon

Clockwise Cal Motor RPM	Calculated Windspeed m/s	Measured Windspeed m/s	Delta m/s	Knots
0	0.00	0.00	0.00	0.0
200	0.98	0.90	0.08	1.9
500	2.45	2.40	0.05	4.8
1000	4.90	4.80	0.10	9.5
1500	7.35	7.30	0.05	14.3
2000	9.80	9.80	0.00	19.0
3000	14.70	14.60	0.10	28.6
4000	19.60	19.50	0.10	38.1
5000	24.50	24.30	0.20	47.6
6000	29.40	29.20	0.20	57.1
7000	34.30	34.10	0.20	66.6
8000	39.20	39.00	0.20	76.2
9000	44.10	43.90	0.20	85.7
10000	49.00	48.80	0.20	95.2
12000	58.80	58.60	0.20	114.2

Direction	Measured Direction	Delta Direction
0	359	0
30	29	1
60	59	1
90	89	1
120	119	1
150	148	2
180	179	1
210	210	0
240	240	0
270	270	0
300	301	-1
330	331	-1
0	1	-1

Note: Delta direction should not exceed + or - 3 degrees.

Counter Clockwise Cal Motor RPM	Calculated Windspeed m/s	Measured Windspeed m/s	Delta m/s
0	0.00	0.00	0.00
200	0.98	1.00	-0.02
500	2.45	2.50	-0.05
1000	4.90	4.90	0.00
1500	7.35	7.40	-0.05
2000	9.80	9.80	0.00
3000	14.70	14.80	-0.10
4000	19.60	19.80	-0.20
5000	24.50	24.60	-0.10
6000	29.40	29.50	-0.10
7000	34.30	34.50	-0.20
8000	39.20	39.40	-0.20
9000	44.10	44.40	-0.30
10000	49.00	49.30	-0.30
12000	58.80	59.60	-0.80

Caution: Do Not exceed 12000 rpm during Wind Speed test.

Wind Speed Threshold < 2.9 gm? yes
 Wind Direction Threshold < 30 gm? yes

Additional Comments	
Potentiometer and potenborometer coupling were replaced , vertical shaft bearings were cleaned and lubricated.	

Note: Delta Windspeed should not exceed + or - 0.3 m/s for 0 - 5000 rpm

Barometer

R.M. Young Company
2801 Aero Park Drive
Traverse City, Michigan 49686 USA

CALIBRATION REPORT**Barometric Pressure Sensor**

Customer: *Raytheon Technical Services Co LLC*

Test Number: 79261 Customer PO: RM35605-50
Test Date: 26 September 2007 Sales Order: 9248

Test Sensor:

Model: 612C1 Serial Number: BP00872

Description: Barometric Pressure Sensor

Report of calibration comparison of test barometric pressure sensor with National Institute of Standards and Technology traceable standard pressure calibrator at five pressures in the R.M. Young Company controlled pressure facility. Calibration accuracy ± 1.0 hPa.

Reference Pressure (hPa)	Voltage Output (millivolts)	Indicated (1) (hPa)
800.0	0	800.0
875.0	1251	875.1
950.0	2501	950.1
1025.0	3750	1025.0
1100.0	4997	1099.8

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology

Reference Instrument	Serial #	NIST Test Reference
Druck Pressure Controller Model DPC1515	51500497	UKAS Lab 0221
Fluke Multimeter Model 8060A	4865407	234027

Tested By: *EChamney*

M E T E O R O L O G I C A L I N S T R U M E N T S
Tel: 231-946-3983 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: www.youngusa.com

Humidity Sensor

R.M. Young Company
2801 Aero Park Drive
Traverse City, Michigan 49686 USA

CALIBRATION REPORT
Relative Humidity Sensor

Customer: *Raytheon Technical Services Co LLC*

Test Number: 78223R
Test Date: 26 September 2007

Customer PC: RM35605-50
Sales Order: 9248

Test Sensor:	
Model: 41372LC	Serial Number: 6733
Description: Temperature/Relative Humidity Sensor	

Report of calibration comparison of test relative humidity sensor with National Institute of Standards and Technology traceable standard relative humidity sensor at five humidity levels in the R.M. Young Company controlled humidity chamber facility. Calibration accuracy $\pm 2.0\%$.

Reference Humidity (%)	Current Output (milliamps)	Indicated (1) Humidity (%)
10.3	6.0	12.2
30.3	6.8	29.9
50.3	12.0	50.0
70.2	15.1	69.2
89.6	17.9	88.8

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument	Serial # NIST Test Reference
Vaisala Humidity Sensor Model 35AC	N475040 TN 266152
Fluke Multimeter Model 8060A	4885407 234027

Tested By: *ECherryway*

M E T E O R O L O G I C A L I N S T R U M E N T S

Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: www.youngusa.com

Temperature Sensor

R.M. Young Company
2801 Aero Park Drive
Traverse City, Michigan 49686 USA

CALIBRATION REPORT
Temperature Sensor

Customer: Raytheon Technical Services Co LLC

Test Number: 78223
Test Date: 26 September 2007

Customer PO: RM35605-50
Sales Order: 9248

Test Sensor:	
Model: 41372LC	Serial Number: 6733
Description: Temperature/Relative Humidity Sensor	

Report of calibration comparison of test temperature sensor with National Institute of Standards and Technology traceable standard thermometers at three temperatures in the R.M. Young Company controlled temperature calibration bath facilities. Calibration accuracy $\pm 0.1^\circ$ Celsius.

Bath Temperature (degrees C)	Current Output (milliamperes)	Indicated (*) Temperature (degrees C)
-49.93	4.012	-49.93
0.03	12.005	0.03
49.99	~9.998	49.99

(1) Calculated from current output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology

Reference Instrument	Serial #	NIST Test Reference
Brooklyn Thermometer Model 43-FC	5006-118	204365
Brooklyn Thermometer Model 22332-D5-FC	25071	249763
Brooklyn Thermometer Model 2X400-D7-FC	77532	228060
Keithley Multimeter Model 191	15232	234027

Tested By: ECherryway

M E T E O R O L O G I C A L I N S T R U M E N T S
Tel: 231-946-3980 Fax: 231-946-4772 Email: metsales@youngusa.com Website: www.youngusa.com

PIR**THE EPPLEY LABORATORY, INC.**

12 Sheffield Ave., P.O. Box 419, Newport, RI 02840 USA

Telephone: 401-847-1020

Fax: 401-847-1031

Email: info@eppleylab.com

Internet: www.eppleylab.com



Scientific Instruments
for Precision Measurements
Since 1917

**STANDARDIZATION OF
EPPLEY PRECISION INFRARED RADIOMETER
Model PIR**

Serial Number: 33023F3

Resistance: 739 Ω at 23 °C

Temperature Compensation Range: -20° to +40° °C

This pyrgeometer has been compared against Eppley's Blackbody Calibration System under radiation intensities of approximately 200 watts meter⁻² and an average ambient temperature of 25 °C as measured by Standard Omega Temperature Probe, RTD#1.

As a result of a series of comparisons, it has been found to have a sensitivity of:

$$3.90 \times 10^{-6} \text{ volts/watts meter}^{-2}$$

The calculation of this constant is based on the fact that the relationship between radiation intensity and emf is rectilinear to intensities of 700 watts meter⁻². This radiometer is linear to within ±1.0% up to this intensity.

The calibration of this instrument is traceable to the International Practical Temperature Scale (IPTS) through a precision low-temperature blackbody.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".

Shipped to: Raytheon Technical Services Date of Test: June 13, 2008
National Science Foundation

Port Hueneme, CA

In Charge of Test: *R.T. Egeman*

S.O. Number: 61667

Date: June 20, 2008

Reviewed by: *Thomas J. Kueh*

Remarks:

PSP**THE EPPELEY LABORATORY, INC.**

12 Sheffield Ave., P.O. Box 419, Newport, RI 02840 USA
 Telephone: 401-847-1020 Fax: 401-847-1031
 Email: info@eppleylab.com Internet: www.eppleylab.com



Scientific Instruments
for Precision Measurements
Since 1917

**STANDARDIZATION OF
EPPELEY PRECISION SPECTRAL PYRANOMETER
Model PSP**

Serial Number: 33090F3

Resistance: 700 Ω at 23 °C
 Temperature Compensation Range: -20° to +40 °C

This radiometer has been compared with Standard Precision Spectral Pyranometer, Serial Number 21231F3 in Eppley's Integrating Hemisphere under radiation intensities of approximately 700 watts meter⁻² (roughly one half a solar constant).

As a result of a series of comparisons, it has been found to have a sensitivity of:

$$8.06 \times 10^{-6} \text{ volts/watts meter}^{-2}$$

The calculation of this constant is based on the fact that the relationship between radiation intensity and emf is rectilinear to intensities of 1400 watts meter⁻². This radiometer is linear to within ± 0.5% up to this intensity.

The calibration of this instrument is traceable to standard self-calibrating cavity pyrheliometers in terms of the Systems Internationale des Unites (SI units), which participated in the Tenth International Pyrheliometric Comparisons (IPC X) at Davos, Switzerland in September-October 2005.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".

Useful conversion facts: 1 cal cm⁻² min⁻¹ = 697.3 watts meter⁻²
 1 BTU/ft²·hr⁻¹ = 3.153 watts meter⁻²

Shipped to: Raytheon Technical Services Date of Test: June 11, 2008
National Science Foundation
 Port Hueneme, CA

In Charge of Test: *R.T. Egan*
 Reviewed by: *Thomas J. Kuk*

S.O. Number: 61666
 Date: June 20, 2008

Remarks:

PAR**Biospherical Instruments Inc.****CALIBRATION CERTIFICATE**

Calibration Date 8/8/2007
 Model Number QSR-240
 Serial Number 6356
 Operator TPC
 Standard Lamp HEC-1630(10/25/2006)
 Probe Excitation Voltage Range: 6 to 18 VDC(+)
 Output Polarity: Positive

Probe Conditions at Calibration(in air):

Calibration Voltage: 6 VDC(+)
 Probe Current: 1.2 mA

Probe Output Voltage:

Probe Illuminated	<u>93.1</u>	mV
Probe Dark	<u>-0.1</u>	mV
Probe Net Response	<u>93.1</u>	mV
RG780 Filter	<u>0.1</u>	mV

Corrected Lamp Output:

Output In Air (same condition as calibration):

9.56E+15 quanta/cm²sec
0.01588 uE/cm²sec

Calibration Scale Factor:

(To calculate irradiance, divide the net voltage reading in Volts by this value.)

Dry: 9.7383E-18 V/(quanta/cm²sec)
5.8644E+00 V/(uE/cm²sec)

Notes:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).
3. The collector should be cleaned frequently with alcohol.
4. Calibration was performed with customer cable, when available.

QSR240R 05/24/95

GUV

GUV-2511 Calibration Certificate											
System Serial Number Calibration database DASSN Microprocessor Tag Number				Date of Calibration Date of Certificate Standard of Spectral Irradiance Operator				3/18/08 3/18/2008 91537(10/25/06) TC			
Monochromatic Channels	Address	Wavelength [nm]	Responsivity [Amps per μW/(cm²·nm)]	ScaleSmall [Volts per μW/(cm²·nm)]	ScaleMedium [Volts per μW/(cm²·nm)]	ScaleLarge [Volts per μW/(cm²·nm)]	OffsetSmall [volts]	OffsetMedium [volts]	OffsetLarge [volts]	Measurement Units	
Ed0320	2	320	2.7010E-10	2.7551E-05	8.0491E-03	2.8497E+00	6.5000E-05	6.4000E-05	2.2300E-04	μW/(cm²·nm)	
Ed0340	6	340	2.2027E-10	2.2467E-05	6.5640E-03	2.3144E+00	3.7400E-04	3.7600E-04	1.1590E-03	μW/(cm²·nm)	
Ed0313	8	313	2.4100E-10	2.4539E-05	7.1692E-03	2.5594E+00	1.0000E-04	1.0000E-04	6.9100E-04	μW/(cm²·nm)	
Ed0305	10	305	1.6600E-11	1.6943E-06	4.9499E-04	1.5313E-01	4.1300E-04	4.0500E-04	-2.2080E-03	μW/(cm²·nm)	
Ed0380	12	380	7.0759E-11	7.2174E-06	2.1086E-03	7.0320E-01	1.1140E-03	1.1050E-03	-1.4780E-03	μW/(cm²·nm)	
Ed0395	18	395	3.4926E-10	3.5624E-05	1.0408E-02	3.3997E+00	1.3100E-04	1.3000E-04	2.2300E-04	μW/(cm²·nm)	
Broadband Channels	Address	Wavelength [nm]	Responsivity [Amps per μE/(cm²·s)]	ScaleSmall [Volts per μE/(cm²·s)]	ScaleMedium [Volts per μE/(cm²·s)]	ScaleLarge [Volts per μE/(cm²·s)]	OffsetSmall [volts]	OffsetMedium [volts]	OffsetLarge [volts]	Measurement Units	
Ed0PAR	13	400-700	1.9205E-05	1.9589E+00	5.7230E+02	2.0376E+05	8.8600E-04	8.7700E-04	-1.3030E-03	μE/(cm²sec)	
Auxiliary Channels	Address	Wavelength	Responsivity	ScaleS	ScaleM	ScaleL	OffsetS	OffsetM	OffsetL	Measurement Units	
Ed0Gnd	0	0	1	1	1	1	0	0	0	V	
Ed0Temp	22	0	1	0.01	0.01	0.01	0	0	0	C	
Ed0Vin	27	0	1	-0.25	-0.25	-0.25	0	0	0	V	

© Biospherical Instruments Inc., 5340 Riley Street, San Diego, California 92110 USA. Contact support@biospherical.com for more information.

Micro-TSG Calibration Files

Temperature Sensor (Micro-TSG) #0226

63

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0226
 CALIBRATION DATE: 12-Jan-08

SBE 45 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = -6.625307e-005
 a1 = 2.870318e-004
 a2 = -3.279891e-006
 a3 = 1.750784e-007


BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	744351.4	1.0000	-0.0000
4.4999	636501.4	4.5000	0.0001
15.0000	405773.8	14.9998	-0.0002
18.5000	351383.7	18.5001	0.0001
24.0000	281927.0	24.0001	0.0001
29.0000	232171.7	28.9998	-0.0002
32.5000	203331.2	32.5001	0.0001

$$\text{Temperature ITS-90} = 1/(a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]) - 273.15 \text{ } (\text{°C})$$

Residual = instrument temperature - bath temperature

Date, Delta T (mdeg C)

 12-Jan-08 -0.00



Conductivity Sensor (Micro-TSG) #0226**64**

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0226
 CALIBRATION DATE: 12-Jan-08

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

<i>g</i> = -1.014919e+000	<i>CPcor</i> = -9.5700e-008
<i>h</i> = 1.569437e-001	<i>CTcor</i> = 3.2500e-006
<i>i</i> = -4.469846e-004	<i>WBOTC</i> = 9.8072e-007
<i>j</i> = 6.064779e-005	

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2549.02	0.00000	0.00000
1.0000	34.9254	2.98440	5059.48	2.98442	0.00002
4.4999	34.9056	3.29231	5250.05	3.29229	-0.00001
15.0000	34.8618	4.27660	5816.78	4.27659	-0.00001
18.5000	34.8524	4.62264	6003.02	4.62263	-0.00001
24.0000	34.8419	5.18201	6292.11	5.18204	0.00003
29.0000	34.8360	5.70518	6550.56	5.70519	0.00001
32.5000	34.8334	6.07863	6728.76	6.07862	-0.00001


$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature[°C]; *p* = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction

 12-Jan-08 1.0000000


Temperature Sensor (Micro-TSG) #0242

63

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0242
 CALIBRATION DATE: 31-Mar-08

SBE 45 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a₀ = -3.912618e-006
 a₁ = 2.847375e-004
 a₂ = -3.193105e-006
 a₃ = 1.720429e-007


BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	649832.1	1.0000	0.0000
4.5000	554895.8	4.5000	-0.0000
15.0000	352330.7	15.0000	0.0000
18.5000	304721.7	18.5000	-0.0000
24.0000	244015.9	24.0000	0.0000
29.0000	200610.7	29.0000	-0.0000
32.5000	175490.5	32.5000	0.0000

$$\text{Temperature ITS-90} = 1/(a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]) - 273.15 \text{ (°C)}$$

Residual = instrument temperature - bath temperature

Date, Delta T (mdeg C)

31-Mar-08 0.00



Conductivity Sensor (Micro-TSG) #0242**64**

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0242
 CALIBRATION DATE: 31-Mar-08

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

<i>g</i> = -9.980367e-001	<i>CPcor</i> = -9.5700e-008
<i>h</i> = 1.523293e-001	<i>CTcor</i> = 3.2500e-006
<i>i</i> = -4.438334e-004	<i>WBOTC</i> = 0.0000e+000
<i>j</i> = 5.882995e-005	

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREO (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2565.99	0.00000	0.00000
1.0000	34.8739	2.98042	5122.83	2.98043	0.00002
4.5000	34.8542	3.28795	5316.56	3.28793	-0.00001
15.0000	34.8112	4.27105	5892.53	4.27104	-0.00001
18.5000	34.8020	4.61668	6081.77	4.61667	-0.00001
24.0000	34.7913	5.17532	6375.44	5.17535	0.00003
29.0000	34.7843	5.69767	6637.87	5.69765	-0.00002


$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

t = temperature [°C]; *p* = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

Date, Slope Correction



Underway Temperature Sensor (Primary Remote)

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0178
 CALIBRATION DATE: 07-Jan-08

SBE 38 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS


a0 = -4.331693e-005
 a1 = 2.811171e-004
 a2 = -2.677535e-006
 a3 = 1.661218e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.50010	750904.5	-1.50006	0.00004
0.99990	671271.1	0.99984	-0.00006
4.49990	575397.3	4.49989	-0.00001
7.99990	494813.7	7.99989	-0.00001
11.49990	426851.8	11.49995	0.00005
14.99990	369349.8	14.99992	0.00002
18.50000	320540.5	18.50001	0.00001
22.00000	278985.0	21.99995	-0.00005
25.49990	243497.2	25.49989	-0.00001
28.99990	213101.8	28.99992	0.00002
32.49990	186995.1	32.49990	0.00000

$$\text{Temperature ITS-90} = 1/\{a0 + a1[\ln(n)] + a2[\ln^2(n)] + a3[\ln^3(n)]\} - 273.15 \text{ (°C)}$$

Residual = instrument temperature - bath temperature

Date, Delta T (mdeg C)



Underway Temperature Sensor (Secondary Remote)

Underway fluorometer (Primary)

PO Box 518
620 Aoplegate St.
Philomath OR 97370



(541) 929-5650
Fax (541) 929-5277
<http://www.wetlabs.com>

Chlorophyll Fluorometer Characterization .

Date: 10/29/08
Serial #: AFLD-011
Job#: 0011007
Tech: K.C

Dark Counts 0.183 volts
CEV 2.902 volts
SF 8.6598

FSV 5.36 volts

Linearity: 0.999 R² (0-1.5 volts)
0.995 R² (0- 5.45 volts)

Notes:

Dark Counts: Signal output of the meter in clean water with black tape over detector.

CEV is the chlorophyll equivalent voltage. This value is the signal output of the fluorometer when using a fluorescent proxy that has been determined to be approximately equivalent to 25 µg/l of a *Thalassiosira weissflogii* phytoplankton culture.

SF is the scale factor used to derive chlorophyll concentration from the signal voltage output of the fluorometer. The scale factor is determine by using the following equation:
SF = (25) / (CEV - dark) e.g.(25 / (2.865 - 0.238) = 9.516)

FSV is the maximum signal voltage output that the fluorometer is capable of.

Chlorophyll concentration expressed in µg/l (mg/m³) can be derived by using the following equation: (µg/l) = (Vmeasured -dark) * SF

The relationship between fluorescence and chlorophyll-a concentrations in-situ is high variable. The scale factor listed on this document was determined by using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer you must perform secondary measurements on the populations of interest. This is typically done using extraction based measurement techniques on discrete samples. For additional information on determination of chlorophyll concentration see [Standard Methods For The Examination Of Water And Wastewater] part 10200 H published jointly by: American Public Health Association, American Water Works Association and Water Environment Federation.

Underway Transmissometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5660
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

Date	November 14, 2007	S/N#	CST-430DR	Pathlength	25 cm
------	-------------------	------	-----------	------------	-------

V_d	Analog meter
V_{air}	0.060 V
V_{ref}	4.794 V
	4.663 V

Temperature of calibration water	23.2 °C
Ambient temperature during calibration	25.9 °C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.
 V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.

Page intentionally left blank.

CTD Pressure Sensor (Fish)**SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0328
CALIBRATION DATE: 01-Aug-07SBE9plus PRESSURE CALIBRATION DATA
10000 psia S/N 53980

DIGIQUARTZ COEFFICIENTS:

C1 = -5.847002e+004
 C2 = 6.910390e-001
 C3 = 1.753360e-002
 D1 = 4.241600e-002
 D2 = 0.000000e+000
 T1 = 3.026040e+001
 T2 = -1.938830e-004
 T3 = 4.330190e-006
 T4 = 2.020250e-009
 T5 = 0.000000e+000


AD590M, AD590B, SLOPE AND OFFSET:
 AD590M = 1.13300e-002
 AD590B = -8.47592e+000
 Slope = 0.99999
 Offset = -0.8347 (dbars)

PRESSURE (PSIA)	INST OUTPUT(Hz)	INST TEMP(C)	INST OUTPUT (PSIA)	CORRECTED INST OUTPUT (PSIA)	RESIDUAL (PSIA)
14.647	33053.30	22.4	15.793	14.583	-0.064
2001.665	33609.50	22.7	2002.748	2001.525	-0.140
3988.943	34155.20	22.7	3990.098	3988.862	-0.081
5976.002	34690.70	22.8	5977.067	5975.820	-0.182
7962.994	35216.70	22.8	7964.504	7963.244	0.250
9950.680	35733.40	22.9	9951.508	9950.236	-0.444
7962.873	35216.70	23.0	7964.485	7963.225	0.352
5975.724	34690.70	23.0	5977.040	5975.792	0.068
3988.622	34155.20	23.1	3990.073	3988.837	0.215
2001.487	33609.50	23.2	2002.738	2001.515	0.028
14.639	33053.30	23.9	15.848	14.638	-0.001

Residual = corrected instrument pressure - reference pressure

Date, Avg Offset(ps):

01-Aug-07 -0.



CTD Primary Pump**NBP-Bird Electronics, Inc**

53° 10'S, 70° 50'W, Punta Arenas, Chile

Phone: (808)434-9769, (808)659-5076 Fax: 011-870-336-661-014

Email: et@NBP.usap.gov

SBE 5+ Pump Service & Calibration

4/17/08

S/N 051646

- Check all O-rings, lubricate and install new O-rings if required
- Check Thrust Washers, replace if necessary
- Remove and clean impeller – reinstall
- Tune R11 as necessary for RPM adjustment
- Seal and purge with Nitrogen

Technician: VS

Location: NBP

CTD Secondary Pump**NBP-Bird Electronics, Inc**

53° 10'S, 70° 50'W, Punta Arenas, Chile

Phone: (808)434-9769, (808)659-5076 Fax: 011-870-336-661-014

Email: et@NBP.usap.gov

SBE 5+ Pump Service & Calibration

4/17/08

S/N 051626

- Check all O-rings, lubricate and install new O-rings if required
- Check Thrust Washers, replace if necessary
- Remove and clean impeller – reinstall
- Tune R11 as necessary for RPM adjustment
- Seal and purge with Nitrogen

Technician: VS

Location: NBP

CTD Primary Conductivity Sensor

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1314
 CALIBRATION DATE: 16-Sep-08

SBE4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHII COEFFICIENTS

g = -4.07542614e+000
 h = 4.71146816e-001
 i = -7.66205187e-005
 j = 2.93708490e-005
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.90637348e-005
 b = 4.70900060e-001
 c = -4.07457414e+000
 d = -8.25580952e-005
 m = 4.1
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.94100	0.00000	0.00000
-0.9780	34.9143	2.81360	8.25651	2.81361	0.00001
1.0320	34.9147	2.98635	8.47421	2.98634	-0.00001
15.0000	34.9154	4.28248	9.95476	4.28246	-0.00002
18.4999	34.9151	4.63005	10.31509	4.63008	0.00003
28.9999	34.9140	5.71650	11.36622	5.71647	-0.00003
32.4999	34.9093	6.09036	11.70560	6.09037	0.00002


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4)/10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature [°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



CTD Primary Conductivity Sensor B

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1431
 CALIBRATION DATE: 30-Sep-08

SBE4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.25244500e+000
 h = 5.52066269e-001
 i = -2.29080850e-004
 j = 4.16367697e-005
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 9.87797409e-006
 b = 5.51421430e-001
 c = -4.25082224e+000
 d = -8.74651349e-005
 m = 4.4
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.77618	0.00000	0.00000
-1.0001	34.9150	2.81178	7.65249	2.81171	-0.00007
0.9999	34.9142	2.98353	7.85247	2.98360	0.00007
14.9999	34.9156	4.28249	9.22275	4.28253	0.00003
18.4999	34.9161	4.63017	9.55563	4.63012	-0.00004
28.9999	34.9138	5.71647	10.52698	5.71647	-0.00000
32.4999	34.9075	6.09008	10.84048	6.09009	0.00001


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature [°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



CTD Secondary Conductivity Sensor

SEA-BIRD ELECTRONICS, INC.
 1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1852
 CALIBRATION DATE: 16-Sep-08

SBE4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHJ COEFFICIENTS

g = -3.96294717e+000
 h = 5.06298522e-001
 i = -6.62958056e-004
 j = 5.90405647e-005
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 7.87569012e-007
 b = 5.03903177e-001
 c = -3.95486739e+000
 d = -7.98696909e-005
 m = 5.3
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.80159	0.00000	0.00000
-0.9780	34.9143	2.81360	7.97445	2.81362	0.00001
1.0320	34.9147	2.98635	8.18571	2.98634	-0.00002
15.0000	34.9154	4.28248	9.62195	4.28246	-0.00002
18.4999	34.9151	4.63005	9.97135	4.63008	0.00003
28.9999	34.9140	5.71650	10.99010	5.71649	-0.00002
32.4999	34.9093	6.09036	11.31887	6.09037	0.00001


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature [°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



CTD Primary Temperature Sensor

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2168
 CALIBRATION DATE: 09-Jul-08

SBE3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.35095567e-003
 h = 6.44079434e-004
 i = 2.35498457e-005
 j = 2.24441402e-006
 f₀ = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121313e-003
 b = 6.01277219e-004
 c = 1.63191127e-005
 d = 2.24598011e-006
 f₀ = 2938.723


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	2938.723	-1.5001	0.00004
1.0000	3107.762	0.9999	-0.00006
4.5000	3356.103	4.5000	-0.00001
8.0000	3618.421	8.0000	-0.00001
11.4999	3895.096	11.5000	0.00007
15.0000	4186.498	15.0000	-0.00003
18.5000	4492.984	18.5000	-0.00000
22.0000	4814.895	22.0000	0.00001
25.5000	5152.558	25.5000	-0.00003
29.0000	5506.304	29.0000	-0.00001
32.5000	5876.434	32.5000	0.00002

$$\text{Temperature ITS-90} = 1/(g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]) - 273.15 \text{ (}^\circ\text{C)}$$

$$\text{Temperature IPTS-68} = 1/(a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]) - 273.15 \text{ (}^\circ\text{C)}$$

Following the recommendation of JPOTS: T_{as} is assumed to be 1.00024 * T₉₀ (-2 to 35 °C)

Residual = instrument temperature - bath temperature



CTD Secondary Temperature Sensor

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2367
 CALIBRATION DATE: 12-Aug-08

SBE3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.37165772e-003
 h = 6.45579033e-004
 i = 2.32241034e-005
 j = 2.14307243e-006
 f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121325e-003
 b = 6.02113601e-004
 c = 1.61201412e-005
 d = 2.14461116e-006
 f0 = 3032.041


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	3032.041	-1.5001	0.00003
0.9999	3206.191	0.9999	-0.00002
4.4999	3462.003	4.4999	-0.00005
7.9999	3732.176	7.9999	-0.00002
11.4999	4017.093	11.4999	0.00002
14.9999	4317.132	15.0000	0.00010
18.4999	4632.637	18.4999	0.00000
21.9999	4963.977	21.9998	-0.00008
25.4999	5311.504	25.4999	-0.00001
28.9999	5675.525	28.9999	-0.00000
32.4999	6056.360	32.4999	0.00002

$$\text{Temperature ITS-90} = 1/g + h[\ln(f/f_0)] + i[\ln^2(f/f_0)] + j[\ln^3(f/f_0)] - 273.15 \text{ (°C)}$$

$$\text{Temperature IPTS-68} = 1/a + b[\ln(f/f_0)] + c[\ln^2(f/f_0)] + d[\ln^3(f/f_0)] - 273.15 \text{ (°C)}$$

Following the recommendation of JPOTS: $T_{\text{eff}} = 1.00024 * T_{\text{90}}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature



CTD Primary Dissolved Oxygen Sensor

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0082
 CALIBRATION DATE: 12-Sep-08p


SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS		A = -1.3323e-003	NOMINAL DYNAMIC COEFFICIENTS	
Soc = 0.2900		B = 1.4180e-004	D1 = 1.92634e-4	
Voffset = -0.6224		C = -2.5328e-006	D2 = -4.64803e-2	
Tau20 = 1.17		E nominal = 0.036	NOMINAL DYNAMIC COEFFICIENTS	
			D1 = 1.92634e-4	
BATH OX	BATH TEMP	BATH SAL	INSTRUMENT	INSTRUMENT
(ml/l)	ITS-90	PSU	OUTPUT(VOLTS)	OXYGEN(ml/l)
1.23	6.00	0.00	1.109	1.22
1.23	2.00	0.00	1.061	1.23
1.23	12.00	0.01	1.185	1.23
1.24	20.00	0.01	1.287	1.24
1.24	26.00	0.01	1.364	1.24
1.25	30.00	0.01	1.423	1.25
4.12	26.00	0.01	3.085	4.12
4.12	30.00	0.01	3.268	4.14
4.13	20.00	0.01	2.837	4.13
4.14	6.00	0.00	2.268	4.14
4.15	12.00	0.01	2.518	4.15
4.20	2.00	0.00	2.121	4.20
6.64	30.00	0.01	4.857	6.62
6.72	26.00	0.01	4.639	6.72
6.74	20.00	0.01	4.239	6.74
6.74	12.00	0.01	3.705	6.74
6.78	6.00	0.00	3.318	6.79
6.79	2.00	0.00	3.047	6.79

$$\text{Oxygen (ml/l)} = \text{Soc} * (\text{V} + \text{Voffset}) * (1.0 + \text{A} * \text{T} + \text{B} * \text{T}^2 + \text{C} * \text{T}^3) * \text{OxSol}(\text{T}, \text{S}) * \exp(\text{E} * \text{P} / \text{K})$$

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen



CTD Secondary Dissolved Oxygen Sensor

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0152
 CALIBRATION DATE: 19-Sep-08p


SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS		A = -5.2035e-003	NOMINAL DYNAMIC COEFFICIENTS	
Soc = 0.4654		B = 3.3169e-004	D1 = 1.92634e-4	
Voffset = -0.4995		C = -3.5756e-006	D2 = -4.64803e-2	
Tau20 = 0.97		E nominal = 0.036	NOMINAL DYNAMIC COEFFICIENTS	
			D1 = 1.92634e-4	
BATH OX	BATH TEMP	BATH SAL	INSTRUMENT	INSTRUMENT
(mM)	ITS-90	PSU	OUTPUT(VOLTS)	OXYGEN(mM)
1.25	2.00	0.00	0.779	1.25
1.25	6.00	0.00	0.813	1.25
1.26	12.00	0.01	0.865	1.26
1.26	20.00	0.01	0.927	1.27
1.27	26.00	0.01	0.969	1.27
1.27	30.00	0.02	0.996	1.28
4.16	30.00	0.02	2.114	4.16
4.16	20.00	0.01	1.902	4.15
4.16	26.00	0.01	2.036	4.17
4.17	12.00	0.01	1.713	4.17
4.18	6.00	0.00	1.551	4.18
4.19	2.00	0.00	1.437	4.18
6.73	20.00	0.01	2.770	6.72
6.73	30.00	0.02	3.112	6.73
6.74	12.00	0.01	2.464	6.75
6.75	26.00	0.01	2.989	6.75
6.78	6.00	0.00	2.205	6.78
6.79	2.00	0.00	2.022	6.80

$$\text{Oxygen (mM)} = \text{Soc} * (\text{V} + \text{Voffset}) * (1.0 + \text{A} * \text{T} + \text{B} * \text{T}^2 + \text{C} * \text{T}^3) * \text{OxSol}(\text{T}, \text{S}) * \exp(\text{E} * \text{P} / \text{K})$$

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [mM], P = pressure [dbar], Residual = instrument oxygen - bath oxygen



CTD Fluorometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 12/18/2008

S/N: FLRTD-397

Chlorophyll concentration expressed in µg/l can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

	Analog Range 1	Analog Range 2	Analog Range 4 (default)	Digital
Dark Counts	0.099	0.061	0.042 V	59 counts
Scale Factor (SF)	6	12	25 µg/W	0.0075 µg/l/count
Maximum Output	4.95	4.95	4.95 V	16400 counts
Resolution	0.5	0.5	0.5 mV	0.7 counts

Ambient temperature during characterization

21.5 °C

Analog Range: 1 (most sensitive, 0–4,000 counts), 2 (midrange, 0–8,000 counts), 4 (entire range, 0–16,000 counts).

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: SF = x ÷ (output - dark counts), where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations *in-situ* is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

CTD Transmissometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

Date April 9, 2008

S/N# CST-831

Pathlength 25 cm

Analog meter

V_d	0.061 V
V_{air}	4.838 V
V_{ref}	4.767 V

Temperature of calibration water	22.9 °C
Ambient temperature during calibration	26.0 °C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x \cdot \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.
 V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref}

Ambient temperature: meter temperature in air during the calibration

V_{sig} Measured signal output of meter.

Revision H

10/3/07

CTD PAR**Biospherical Instruments Inc**

CALIBRATION CERTIFICATE

UNDERWATER PAR SENSOR WITH LOG AMPLIFIER

Calibration Date:	07/28/08	Job No.:	R-9967																																																																																									
Model Number:	QSP-200L																																																																																											
Serial Number:	4471																																																																																											
Operator:	TPC																																																																																											
Standard Lamp:	91537(10/25/06)																																																																																											
Operating Voltage Range:	6	to	15 VDC (+)																																																																																									
Note: The QSP-200 uses a log amplifier to measure the detector signal current with $V = \log I$ (Amps) / I_{Ref}																																																																																												
To calculate irradiance, use this formula:																																																																																												
Irradiance = Calibration factor * (10^ALight Signal Voltage - 10^ADark Voltage)																																																																																												
With the appropriate (solar corrected) Irradiance Calibration Factor:																																																																																												
Dry Calibration Factor:	2.11E+12 quanta/cm ² ·sec/"amps"	3.51E-06 μ Einstens/cm ² ·sec/"amps"																																																																																										
Wet Calibration Factor:	3.56E+12 quanta/cm ² ·sec/"amps"	5.91E-06 μ Einstens/cm ² ·sec/"amps"																																																																																										
Sensor Test Data and Results⁴⁾																																																																																												
<table border="1"> <thead> <tr> <th>Sensor Supply Current (Dark):</th> <th>71.3 mA</th> <th colspan="3"></th> </tr> <tr> <th>Supply Voltage:</th> <th>6 Volts</th> <th colspan="3"></th> </tr> <tr> <th>Lamp Integrated PAR Irradiance:</th> <th>8.83E+15 quanta/cm²·sec</th> <th>0.01467 μEinstens/cm²·sec</th> <th colspan="2"></th> </tr> <tr> <th>SC3 Immersion Coefficient:</th> <th>0.594</th> <th>Scalar Correction:</th> <th>1</th> <th>PAR Solar Correction:</th> </tr> </thead> <tbody> <tr> <td>Nominal Filter OD</td> <td>Calibrated Trans.</td> <td>Measured Sensor Voltage</td> <td>Measured Trans.</td> <td>Estimated Measured Signal Output (Amps) (Volts)</td> <td>Calc. Signal Output (Amps) (Volts)</td> <td>Error (Volts)</td> <td>Error (%)</td> <td>Test Irrad. (quanta/cm²·sec)</td> </tr> <tr> <td>No Filter</td> <td>100.00%</td> <td>3.621</td> <td>100.00%</td> <td>4.18E-07</td> <td>4.18E-07</td> <td>3.621</td> <td>0.000</td> <td>0.0</td> <td>8.83E+15</td> </tr> <tr> <td>0.3</td> <td>36.10%</td> <td>3.182</td> <td>36.33%</td> <td>1.52E-07</td> <td>1.51E-07</td> <td>3.179</td> <td>-0.003</td> <td>-0.8</td> <td>3.21E+15</td> </tr> <tr> <td>0.5</td> <td>27.60%</td> <td>3.069</td> <td>27.98%</td> <td>1.17E-07</td> <td>1.15E-07</td> <td>3.063</td> <td>-0.006</td> <td>-1.4</td> <td>2.47E+15</td> </tr> <tr> <td>1</td> <td>9.27%</td> <td>2.611</td> <td>9.72%</td> <td>4.07E-08</td> <td>3.88E-08</td> <td>2.590</td> <td>-0.021</td> <td>-4.7</td> <td>8.59E+14</td> </tr> <tr> <td>2</td> <td>1.11%</td> <td>1.724</td> <td>1.23%</td> <td>5.15E-09</td> <td>4.84E-09</td> <td>1.680</td> <td>-0.044</td> <td>-9.9</td> <td>1.09E+14</td> </tr> <tr> <td>3</td> <td>0.05%</td> <td>0.711</td> <td>0.09%</td> <td>3.68E-10</td> <td>2.23E-10</td> <td>0.570</td> <td>-0.141</td> <td>-39.0</td> <td>7.73E+12</td> </tr> </tbody> </table>				Sensor Supply Current (Dark):	71.3 mA				Supply Voltage:	6 Volts				Lamp Integrated PAR Irradiance:	8.83E+15 quanta/cm ² ·sec	0.01467 μ Einstens/cm ² ·sec			SC3 Immersion Coefficient:	0.594	Scalar Correction:	1	PAR Solar Correction:	Nominal Filter OD	Calibrated Trans.	Measured Sensor Voltage	Measured Trans.	Estimated Measured Signal Output (Amps) (Volts)	Calc. Signal Output (Amps) (Volts)	Error (Volts)	Error (%)	Test Irrad. (quanta/cm ² ·sec)	No Filter	100.00%	3.621	100.00%	4.18E-07	4.18E-07	3.621	0.000	0.0	8.83E+15	0.3	36.10%	3.182	36.33%	1.52E-07	1.51E-07	3.179	-0.003	-0.8	3.21E+15	0.5	27.60%	3.069	27.98%	1.17E-07	1.15E-07	3.063	-0.006	-1.4	2.47E+15	1	9.27%	2.611	9.72%	4.07E-08	3.88E-08	2.590	-0.021	-4.7	8.59E+14	2	1.11%	1.724	1.23%	5.15E-09	4.84E-09	1.680	-0.044	-9.9	1.09E+14	3	0.05%	0.711	0.09%	3.68E-10	2.23E-10	0.570	-0.141	-39.0	7.73E+12
Sensor Supply Current (Dark):	71.3 mA																																																																																											
Supply Voltage:	6 Volts																																																																																											
Lamp Integrated PAR Irradiance:	8.83E+15 quanta/cm ² ·sec	0.01467 μ Einstens/cm ² ·sec																																																																																										
SC3 Immersion Coefficient:	0.594	Scalar Correction:	1	PAR Solar Correction:																																																																																								
Nominal Filter OD	Calibrated Trans.	Measured Sensor Voltage	Measured Trans.	Estimated Measured Signal Output (Amps) (Volts)	Calc. Signal Output (Amps) (Volts)	Error (Volts)	Error (%)	Test Irrad. (quanta/cm ² ·sec)																																																																																				
No Filter	100.00%	3.621	100.00%	4.18E-07	4.18E-07	3.621	0.000	0.0	8.83E+15																																																																																			
0.3	36.10%	3.182	36.33%	1.52E-07	1.51E-07	3.179	-0.003	-0.8	3.21E+15																																																																																			
0.5	27.60%	3.069	27.98%	1.17E-07	1.15E-07	3.063	-0.006	-1.4	2.47E+15																																																																																			
1	9.27%	2.611	9.72%	4.07E-08	3.88E-08	2.590	-0.021	-4.7	8.59E+14																																																																																			
2	1.11%	1.724	1.23%	5.15E-09	4.84E-09	1.680	-0.044	-9.9	1.09E+14																																																																																			
3	0.05%	0.711	0.09%	3.68E-10	2.23E-10	0.570	-0.141	-39.0	7.73E+12																																																																																			
Dark Before:	0.168 Volts	$I_{Ref} = 1.00E-10$ Amps																																																																																										
Light - No Filter Hldr.:	3.821 Volts	$I_{Dark} = 1.48E-10$ Amps																																																																																										
Dark After - NPH:	0.175 Volts					RG780	1.238																																																																																					
Average Dark	0.171 Volts	$10^{V_{Dark}} = 1.483372$ Amps																																																																																										
Notes:																																																																																												
1. Annual calibration is recommended.																																																																																												
2. There is increasing error associated with readings below zero.																																																																																												
3. The collector should be cleaned frequently with alcohol.																																																																																												
4) This section is for internal use and for more advanced analysis.																																																																																												

QSP-200L,QSP2300(2006-).xls

MOCNESS***Mocness 1 Temperature Sensor*****SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1497
 CALIBRATION DATE: 01-Mar-07

SBE3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.73706927e-003
 h = 6.67918416e-004
 i = 2.80235686e-005
 j = 2.53960597e-006
 f0 = 1000.0

ITS-68 COEFFICIENTS

a = 3.68121409e-003
 b = 5.95365624e-004
 c = 1.52403921e-005
 d = 2.54111460e-006
 f0 = 5372.853

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	5372.853	-1.5001	-0.00013
1.0000	5685.071	1.0001	0.00013
4.5000	6143.920	4.5001	0.00012
8.0000	6628.823	8.0000	0.00002
11.5000	7140.486	11.4999	-0.00011
15.0000	7679.612	14.9999	-0.00009
18.5000	8246.828	18.4999	-0.00008
22.0000	8842.773	22.0000	-0.00001
25.5000	9468.060	25.5002	0.00016
29.0000	10123.210	29.0001	0.00015
32.5000	10808.744	32.4998	-0.00016


$$\text{Temperature ITS-90} = 1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15 \text{ } (\text{°C})$$

$$\text{Temperature ITS-68} = 1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15 \text{ } (\text{°C})$$

Following the recommendation of JPOTS: T_{es} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

Date, Offset(mdeg C)



Mocness 10 Temperature Sensor

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4080
CALIBRATION DATE: 11-Jun-05

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPRATURE SCALE

ITS-90 COEFFICIENTS

$g = 4.38303148e-003$
 $h = 6.44579053e-004$
 $i = 2.20755252e-005$
 $j = 1.68991794e-006$
 $f_0 = 1000.0$

ITS-68 COEFFICIENTS

$a = 3.68121166e-003$
 $b = 6.01351307e-004$
 $c = 1.63810277e-005$
 $d = 1.69141269e-006$
 $f_0 = 3091.541$

BATH TEMP (ITS-90)	INSTRUMENT FREO (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.4999	3091.541	-1.5000	-0.00005
1.0001	3269.357	1.0002	0.00006
4.5001	3530.586	4.5002	0.00007
8.0001	3806.530	8.0000	-0.00005
11.5001	4097.616	11.5000	-0.00008
15.0000	4404.242	15.0000	0.00003
18.5000	4726.796	18.5000	0.00004
22.0001	5065.662	22.0001	-0.00002
25.5001	5421.205	25.5001	0.00001
29.0001	5793.784	29.0001	0.00001
32.5001	6183.743	32.5001	-0.00001


$$\text{Temperature ITS-90} = 1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15 \text{ } ^\circ\text{C}$$

$$\text{Temperature ITS-68} = 1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15 \text{ } ^\circ\text{C}$$

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 $^\circ\text{C}$)

Residual = instrument temperature - bath temperature

Date, Offset(mdeg C)



Mocness 1 Conductivity Sensor

SEA-BIRD ELECTRONICS, INC.
 1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2067
 CALIBRATION DATE: 23-Sep-08

SBE4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHII COEFFICIENTS

g = -1.03546725e+001
 h = 1.45983285e+000
 i = -4.63974043e-003
 j = 4.12854546e-004
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.70411528e-009
 b = 1.44531355e+000
 c = -1.03167895e+001
 d = -5.07463523e-005
 m = 9.0
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.67194	0.00000	0.00000
-1.0001	34.7327	2.79846	5.14757	2.79843	-0.00003
1.0050	34.7327	2.96993	5.26155	2.96996	0.00003
14.9999	34.7330	4.26246	6.05129	4.26250	0.00004
18.5000	34.7330	4.60851	6.24559	4.60846	-0.00005
28.9999	34.7300	5.68976	6.81677	5.68979	0.00002
32.5000	34.7261	6.06204	7.00244	6.06203	-0.00001


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^2 + bf^3 + c + df) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature [°C]; p = pressure[decibars]; δ = CTcor; ε = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



Mocness 10 Conductivity Sensor

SEA-BIRD ELECTRONICS, INC.
 1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2069
 CALIBRATION DATE: 23-Sep-08

SBE4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.02815084e+001
 h = 1.43547173e+000
 i = -4.08511708e-003
 j = 3.77463705e-004
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.38192625e-008
 b = 1.42301818e+000
 c = -1.02500806e+001
 d = -6.14592515e-005
 m = 8.0
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.68400	0.00000	0.00000
-1.0001	34.7327	2.79846	5.18302	2.79843	-0.00003
1.0050	34.7327	2.96993	5.29796	2.96997	0.00003
14.9999	34.7330	4.26246	6.09421	4.26250	0.00003
18.5000	34.7330	4.60851	6.29008	4.60846	-0.00005
28.9999	34.7300	5.68976	6.86586	5.68979	0.00003
32.5000	34.7261	6.06204	7.05300	6.06202	-0.00002


$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^2 + bf^3 + c + df) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature [°C]; p = pressure[decibars]; δ = CTcor; ε = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



Mocness Fluorometer Sensor

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 12/18/2008

S/N: FLRTD-855

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

	Analog Range 1	Analog Range 2	Analog Range 4 (default)	Digital
Dark Counts	0.106	0.066	0.048 V	63 counts
Scale Factor (SF)	6	13	25 $\mu\text{g/l}/\text{V}$	0.0077 $\mu\text{g/l}/\text{count}$
Maximum Output	4.97	4.97	4.97 V	16326 counts
Resolution	0.6	0.6	0.6 mV	0.6 counts

Ambient temperature during characterization 23.2 °C

Analog Range: 1 (most sensitive, 0–4,000 counts), 2 (midrange, 0–8,000 counts), 4 (entire range, 0–16,000 counts).

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $SF = x / (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.


Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations *in-situ* is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

Mocness Transmissometer Sensor

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wellabs.com

C-Star Calibration

Date July 27, 2007	Customer National Science Foundation	Work order 002
Job # 0507014	S/N# CST-889DR	Pathlength 25 cm

	Analog meter
V_d	0.060 V
V_{air}	4.837 V
V_{ref}	4.727 V

Temperature of calibration water	24.8 °C
Ambient temperature during calibration	26.1 °C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.
 V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.

Mocness Depth Sensor

from the rebuild of

MOCNESS MELTIC

7/15/01/COPY 1

NBP-01-02

DEPTH SENSOR CALIBRATION S/N 139.5 VI-28-2001

The pressure sensor used in the MOCNESS is a titanium strain gauge with an internal temperature sensor. The temperature of the sensor is measured and used to correct for the thermal offsets in the measurement of pressure. The MOCNESS measures the voltage across the pressure and temperature bridges of the sensor and reports these values in its output data stream. The MOCNESS pressure sensor is calibrated at several pressure points and at two temperatures. There are no adjustments in the MOCNESS hardware and all calibration is done in software in the surface control computer. The values sent up the wire in the MOCNESS data stream (the bridge voltages) are scaled to be sent as integers in the range of 0-99999 for pressure and 0-9999 for temperature. The calibration data is fit by the following equation-

$$Z = (C1 \cdot Vt + C0) \cdot Vp^2 + (B1 \cdot Vt + B0) \cdot Vp + (A1 \cdot Vt + A0)$$

where-

Z=pressure in decibars (1 decibar is approx 1m of water)

Vp=voltage reading in data stream from pressure sensor

Vt=voltage reading in data stream from temperature sensor

The following constants are for your MOCNESS underwater unit.

```

serial_number =
    139
C1 =
    -5.080432165373504e-14
C0 =
    1.526908907925555e-10
B1 =
    -3.666039764744632e-08
B0 =
    0.10373235478058
A1 =
    -7.451551675779305e-04
A0 =
    -1.936355531955248e+02

```

DEPTH SENSOR CALIBRATION
February 24, 2001

The pressure sensor used in the MOCNESS is a titanium strain gauge with an internal temperature sensor. The temperature of the sensor is measured and used to correct for the thermal offsets in the measurement of pressure. The MOCNESS measures the voltage across the pressure and temperature bridges of the sensor and reports these values in its output data stream. The MOCNESS pressure sensor is calibrated at several pressure points and at two temperatures. There are no adjustments in the MOCNESS hardware and all calibration is done in software in the surface control computer. The values sent up the wire in the MOCNESS data stream (the bridge voltages) are scaled to be sent as integers in the range of 0-99999 for pressure and 0-9999 for temperature. The calibration data is fit by the following equation-

$$Z = (C1 \cdot Vt + C0) \cdot Vp^2 + (B1 \cdot Vt + B0) \cdot Vp + (A1 \cdot Vt + A0)$$

where-

Z=pressure in decibars (1 decibar is approx 1m of water)

Vp=voltage reading in data stream from pressure sensor

Vt=voltage reading in data stream from temperature sensor

The following constants are for your MOCNESS underwater unit.

```
serial_number =  
149  
C1 =  
1.346556023582793e-13  
C0 =  
-9.659981195534541e-10  
B1 =  
-4.748135708603977e-08  
B0 =  
0.10487288807002  
A1 =  
0.00155922569240  
A0 =  
-1.728933890313200e+02
```