

# Data Report NBP1402

January 14, 2014 – March 16, 2014



Photo by David Gwyther

**RVIB Nathaniel B. Palmer**  
**United States Antarctic Program**  
**Antarctic Support Contractor**  
**Prepared by Scott Walker and Sean Drabant**

## Table of Contents

<b>INTRODUCTION</b>	<b>3</b>
<b>DISTRIBUTION CONTENTS AT A GLANCE</b>	<b>4</b>
EXTRACTING DATA	5
<b>DISTRIBUTION CONTENTS</b>	<b>6</b>
CRUISE INFORMATION	6
<i>Cruise Track</i>	6
<i>Satellite Images</i>	6
NBP DATA PRODUCTS	6
<i>MGD77</i>	8
SCIENCE OF OPPORTUNITY	9
<i>ADCP</i>	9
<i>pCO<sub>2</sub></i>	9
CRUISE SCIENCE	10
<i>XBT</i>	10
RVDAS	10
<i>Sensors and Instruments</i>	10
Underway Sensors	11
Meteorology and Radiometry	11
Geophysics	11
Oceanography	11
Navigational Instruments	12
<i>Data</i>	12
Underway Data /rvdas/uw	13
Sound Velocity Probe (svp1)	13
Meteorology (mwx1)	13
MET string	13
PUS string	14
SUS string	14
Knudsen (knud)	14
Fluorometer (flr1)	14
pCO <sub>2</sub> (pco2)	15
Micro-TSG (tsg1)	15
Micro-TSG #2 (tsg2)	15
Gravimeter (grv1)	15
Engineering (eng1)	16
Hydro-DAS (hdas)	16
GUV Data (pguv)	16
Remote Temperature (rtmp)	17
Oxygen Data (oxyg)	17
Winch Data (bwnc, twnc, cwnc)	17
Navigational Data /rvdas/nav	18
Seapath GPS (seap)	18
Trimble (P-Code) GPS (PCOD)	20
Gyro Compass (gyr1)	21
ADCP Course (adcp)	21
Processed Data /process/	22
pCO <sub>2</sub> -merged	22
Calculations	23
PAR	23
PSP	23
PIR	24
<b>ACQUISITION PROBLEMS AND EVENTS</b>	<b>25</b>
<b>APPENDIX: SENSORS AND CALIBRATIONS</b>	<b>26</b>

## Introduction

The NBP data acquisition systems continuously log data from the instruments used during the cruise. This document describes:

- The structure and organization of the data on the distribution media
- The format and contents of the data strings
- Formulas for calculating values
- Information about the specific instruments in use during the cruise
- A log of acquisition problems and events during the cruise that may affect the data
- Scanned calibration sheets for the instruments in use during the cruise.

The data is distributed on a DVD-R written in UDF format. It is readable by most modern computer platforms.

All the data has been compressed using Unix “gzip,” identified by the “.tz” extension. It has been copied to the distribution media in the Unix tar archive format, “.tar” extension. Tools are available on all platforms for uncompressing and de-archiving these formats: On Macintosh, one can use Stuffit Expander with DropStuff. On Windows operating systems, one can use WinZip or 7zip.

MultiBeam and raw ADCP data are distributed separately.

*IMPORTANT: Read the last section, “Acquisition Problems and Events,” for important information that may affect the processing of this data.*

## Distribution Contents at a Glance

### Volume 1 of 1: NBP1402

File	Description
/	Root level directory
NBP1402.trk	Text file of cruise track (lat,lon)
NBP1402.mgd	Full Cruise MGD77 data file
NBP1402.gmt	GMT binary file of MGD77 data
INSTCOEF.TXT	Instrument Coefficient File
1402DATA.doc	Data Report NBP1402 (MS Word)
1402DATA.pdf	Data Report NBP1402 (PDF format)
/cal-sheets	Calibration Sheets
NBP1402-Sensors.doc	Sensor Calibration Sheet Reference
NBP1402-CalSheets.zip	Sensor Calibration Sheet files
/plots	Cruise track plots
NBP1402-track.jpg	Cruise track plot (JPEG format)
NBP1402-track.png	Cruise track plot (PNG format)
/process	Processed data
1402JGOF.tz	JGOFs format data files
1402QC.tz	Daily RVDAS QC postscript plots
1402PCO2.tz	Merged pCO2 data files
1402MGD.tz	MGD Data
1402PROC.tz	Other processed data
/rvdas/nav	Navigation data
1402dcp.tz	ADCP Data Sets
1402gyr1.tz	Gyro raw data
1402PCOD.tz	Trimble P-code raw data
1402seap.tz	Seapath data
/rvdas/uw	Underway data
1402Abwnc.tz	Baltic winch data
1402Actdd.tz	CTD depth data
1402Aeng1.tz	Engineering data
1402Ahdas.tz	HydroDAS raw data
1402Aknud.tz	Knudsen raw data
1402Ambdp.tz	Multibeam depth data
1402Amwx1.tz	Meteorology raw data
1402Aoxyg.tz	Oxygen sensor
1402ApcO2.tz	pCO2 raw data
1402Apguv.tz	GUV raw data
1402Artmp.tz	Sound velocity probe (in ADCP well)
1402Atsg1.tz	Micro TSG data
1402Atsg2.tz	2 <sup>nd</sup> Micro TSG data
/Imagery	Satellite Imagery
1402Imagery.tz	Collection of Imagery Files
/ocean	Ocean data
1402ctd.tz	CTD Data
/MultiBeam	Raw multibeam data
[dates]	

## Extracting Data

The Unix tar command has many options. It is often useful to know exactly how an archive was produced when expanding its contents. All archives are gzipped tar files and were created using the command,

```
tar -czvf archive_filename files_to_archive
```

To create a list of the files in the archive, use the Unix command,

```
tar -tvf archive_filename > contents.list
```

where `contents.list` is the name of the file to create

To extract the files from the archive:

```
tar -xvf archive_filename file(s)_to_extract
```

G-zipped files will have a “.tz” extension on the filename. “.tz” stands for tared and gzipped. These files can be decompressed after de-archiving, using the Unix command,

```
gunzip filename.tz
```

## Distribution Contents

### Cruise Information

NBP1402 departed Hobart, Tasmania on November 19, 2013

Data logging was started on January 30, 2014 04:02 UTC

Data logging was ended on March 15, 2014 13:35 UTC

### Cruise Track

The distribution DVD includes a GMT cruise track file (NBP1402.trk). It contains the longitude and latitude of the ship's position at one-minute intervals extracted from the NBP1402.gmt file.

PNG and PDF cruise track files have been produced and placed in the /plots directory.

### Satellite Images

Satellite Images received for this cruise can be found in the file called /Imagery/1402Imagery.tar. Each type of image is contained in a .tz file within that file.

### NBP Data Products

The IT staff on the NBP creates two processed data products for every cruise: JGOFS and MGD77.

The data processing scripts used to produce JGOFS and MGD77 data sets create a lot of intermediate files. These files are included on the data distribution media in a file called /process/1402proc.tar. These files are not intended to be end-products. They are included to make re-processing easier in the event of an error, but no extensive detail of the formats is included in this document. If you have any questions, please contact [itvessel@usap.gov](mailto:itvessel@usap.gov).

**JGOFS**

The JGOFS data set can be found on the distribution media in the file /process/1402jgof.tar. The archive contains one file produced for each day named jgDDD.dat.tz, where DDD is the year-day the data was acquired. The ".tz" extension indicates that the individual files are compressed before archiving. Each daily file consists of 22 columnar fields in text format as described in the table below. The JGOFS data set is created from calibrated data decimated at one-minute intervals. Several fields are derived measurements from more than a single raw input. For example, Course Made Good (CMG) and Speed Over Ground (SOG) are calculated from gyro and GPS inputs. Daily plots during the cruise are produced from the JGOFS data set. Note: Null, unused, or unknown fields are indicated as "NAN" 9999 in the JGOFS data.

Field	Data	Units
01	UTC date	dd/mm/yy
02	UTC time	hh:mm:ss
03	SEAPATH latitude (negative is South)	tt.tttt
04	SEAPATH longitude (negative is West)	ggg.gggg
05	Speed over ground	Knots
06	GPS HDOP	-
07	Gyro Heading	Degrees (azimuth)
08	Course made good	Degrees (azimuth)
09	Mast PAR	$\mu$ Einsteins/meter <sup>2</sup> sec
10	Sea surface temperature (remote)	°C
11	Sea surface conductivity (TSG1)	siemens/meter
12	Sea surface salinity (TSG1)	PSU
13	Sea depth (uncorrected, calc. sw sound vel. 1500 m/s)	meters
14	True wind speed (max speed windbird)	meters/sec
15	True wind direction (max speed windbird)	degrees (azimuth)
16	Ambient air temperature	°C
17	Relative humidity	%
18	Barometric pressure	mBars
19	Sea surface fluorometry	$\mu$ g/l (mg/m <sup>3</sup> )
20	Transmissometer	%
21	PSP	W/m <sup>2</sup>
22	PIR	W/m <sup>2</sup>

**MGD77**

The MGD77 data set is contained in a single file for the entire cruise. It can be found in the top level of the distribution data structure as NBP1402.mgd. The file NBP1402.gmt is created from the MGD77 dataset using the "mgd77togmt" utility. NBP1402.gmt can be used with the GMT plotting package.

The data used to produce the NBP1402.mgd file can be found on the distribution media in the file /process/1402proc.tar. The data files in the archive contain a day's data and follow the naming convention Dddd.fnl.tz, where ddd is the year-day. These files follow a space-delimited columnar format that may be more accessible for some purposes. They contain data at one-second intervals rather than one minute and are individually "gzipped" to save space. Below is a detailed description of the MGD77 data set format. The other files in the archive contain interim processing files and are included to simplify possible reprocessing of the data using the RVDAS NBP processing scripts.

All decimal points are implied. Leading zeros and blanks are equivalent. Unknown or unused fields are filled with 9's. All "corrections", such as time zone, diurnal magnetics, and EOTVOS, are understood to be added.

Col	Len	Type	Contents	Description, Possible Values, Notes
1	1	Int	Data record type	Set to "5" for data record
2-9	8	Char	Survey identifier	
10-12	3	int	Time zone correction	Corrects time (in characters 13-27) to UTC when added; 0 = UTC
13-16	4	int	Year	4 digit year
17-18	2	int	Month	2 digit month
19-20	2	int	Day	2 digit day
21-22	2	int	Hour	2 digit hour
23-27	5	real	Minutes x 1000	
28-35	8	real	Latitude x 100000	+ = North - = South. (-9000000 to 9000000)
36-44	9	real	Longitude x 100000	+ = East - = West. (-18000000 to 18000000)
45	1	int	Position type code	1=Observed fix 3=Interpolated 9=Unspecified
46-51	6	real	Bathymetry, 2- way travel time	In 10,000th of seconds. Corrected for transducer depth and other such corrections
52-57	6	real	Bathymetry, corrected depth	In tenths of meters.
58-59	2	int	Bathymetric correction code	This code details the procedure used for determining the sound velocity correction to depth
60	1	int	Bathymetric type code	1 = Observed 3 = Interpolated (Header Seq. 12) 9 = Unspecified
61-66	6	real	Magnetics total field, 1 <sup>ST</sup> sensor	In tenths of nanoteslas (gammas)
67-72	6	real	Magnetics total field, 2 <sup>ND</sup> sensor	In tenths of nanoteslas (gammas), for trailing sensor
73-78	6	real	Magnetics residual field	In tenths of nanoteslas (gammas). The reference field used is in Header Seq. 13
79	1	int	Sensor for residual field	1 = 1 <sup>st</sup> or leading sensor 2 = 2 <sup>nd</sup> or trailing sensor 9 = Unspecified
80-84	5	real	Magnetics diurnal correction	In tenths of nanoteslas (gammas). (In nanoteslas) if 9-filled (i.e., set to "+9999"), total and residual fields are assumed to

Col	Len	Type	Contents	Description, Possible Values, Notes
				be uncorrected; if used, total and residuals are assumed to have been already corrected.
85-90	6	F6.0	Depth or altitude of magnetics sensor	(In meters) + = Below sea level 3 = Above sea level
91-97	7	real	Observed gravity	In 10 <sup>th</sup> of mgals. Corrected for Eotvos, drift, tares
98-103	6	real	EOTVOS correction	In 10 <sup>th</sup> of mgals. $E = 7.5 V \cos \phi \sin \alpha + 0.0042 V^*V$
104-108	5	real	Free-air anomaly	In 10 <sup>th</sup> of mgals G = observed G = theoretical
109-113	5	char	Seismic line number	Cross-reference for seismic data
114-119	6	char	Seismic shot-point number	
120	1	int	Quality code for navigation	5= Suspected, by the originating institution 6= Suspected, by the data center 9= No identifiable problem found

## Science of Opportunity

### ADCP

The shipboard ADCP system measures currents in a depth range from about 30 to 300 m -- in good weather. In bad weather or in ice, the range is reduced, and sometimes no valid measurements are made. ADCP data collection is the OPP-funded project of Eric Firing (University of Hawaii) and Teri Chereskin (Scripps Institution of Oceanography). Data is collected on both the LMG and the NBP for the benefit of scientists on individual cruises, and for the long-term goal of building a profile of current structure in the Southern Ocean.

A data feed is sent from the ADCP system to RVDAS whenever a reference layer is acquired. This feed contains east and north vectors for ship's speed, relative to the reference layer, and ship's heading. Collected files (one per day) are archived in 1402adcp.tar in the directory /rvdas/nav.

### pCO<sub>2</sub>

The NBP carries a pCO<sub>2</sub> measurement system from Lamont-Doherty Earth Observatory (LDEO). pCO<sub>2</sub> data is recorded by RVDAS and transmitted to LDEO at the end of each cruise. You will find pCO<sub>2</sub> data in a file named 1402pco2.tar in the /process directory, which contains the pCO<sub>2</sub> instrument's data merged with GPS, meteorological and other oceanographic measurements. For more information contact Colm Sweeney (csweeney@ldeo.columbia.edu).

## Cruise Science

### **XBT**

During the cruise, eXpendable BathyThermographs were used to obtain water column temperature profiles, providing corrections to the sound velocity profile for the multibeam system. The data files from these launches are included as 1402xbt.tar in the /ocean directory. No XBTs were collected on this cruise.

### **RVDAS**

The Research Vessel Data Acquisition System (RVDAS) was developed at Lamont-Doherty Earth Observatory of Columbia University and has been in use on its research ship for many years. It has been extensively adapted for use on the USAP research vessels.

Daily data processing of the RVDAS data is performed to calibrate and convert values into useable units and as a quality-control on operation of the DAS. Raw and processed data sets from RVDAS are included in the data distribution. The tables below provide detailed information on the sensors and data. Be sure to read the “Significant Acquisition Events” section for important information about data acquisition during this cruise.

### **Sensors and Instruments**

RVDAS data is divided into two general categories, *underway and navigation*. They can be found on the distribution media as subdirectories under the top level rvdas directory: /rvdas/uw, and /rvdas/nav. Processed oceanographic data is in the top level directory, /process. Each instrument or sensor produces a data file named with its channel ID. Each data file is g-zipped to save space on the distribution media. Not all data types are collected every day or on every cruise.

The naming convention for data files produced by the sensors and instruments is

NBP[CruiseID][ChannelID].dDDD

**Example:**     NBP1402mwx1.d025

- The CruiseID is the numeric name of the cruise, in this case, NBP1402.
- The ChannelID is a 4-character code representing the system being logged. An example is “mwx1,” the designation for meteorology.
- DDD is the day of year the data was collected

**Underway Sensors****Meteorology and Radiometry**

Measurement	Channel ID	Collect. Status	Rate	Instrument
Air Temperature	mwx1	continuous	1 sec	R.M. Young 41372LC
Relative Humidity	mwx1	continuous	1 sec	R.M. Young 41372LC
Wind Speed/Direction	mwx1	continuous	1 sec	Gill 1390-PK-062/R
Barometer	mwx1	continuous	1 sec	R.M. Young 61201
PIR (LW radiation)	mwx1	continuous	1 sec	Eppley PIR
PSP (SW radiation)	mwx1	continuous	1 sec	Eppley PSP
PAR	mwx1	continuous	1 sec	BSI QSR-240
GUV	pguv	continuous	2 sec	BSI PUV-2511
PUV	pguv	not collected		BSI PUG-2500

**Geophysics**

Measurement	Channel ID	Collect. Status	Rate	Instrument
Gravimeter	grv1	continuous	1 sec	BGM-3
Magnetometer	mag1	continuous	15 sec	EG&G G-866
Bathymetry	knud	continuous	Varies	Knudsen 320B/R Knudsen 3260

**Oceanography**

Measurement	Channel ID	Collect. Status	Rate	Instrument
Conductivity	mtsg	Continuous	6 sec	SeaBird SBE-45
Salinity	mtsg	Continuous	6 sec	Calc. from pri. temp
Sea Surface Temp	mtsg	Continuous	6 sec	SeaBird SBE 38
Fluorometry	hdas	Continuous	2 sec	WET Lab AFL
Transmissometry	hdas	Continuous	2 sec	WET Lab C-Star
pCO <sub>2</sub>	pco2	Continuous	70 sec	(LDEO)
ADCP	adcp	Continuous	varies	RD Instruments
Oxygen	oxyg	Continuous	10 sec	Oxygen Optode 3835

**Navigational Instruments**

Measurement	Channel ID	Collect. Status	Rate	Instrument
Trimble GPS	PCOD	Continuous	1 sec	Trimble 20636-00SM
Gyro	gyr1	Continuous	0.2 sec	Yokogawa Gyro
SeaPath	seap	Continuous	1 sec	SeaPath 330

***Data***

Data is received from the RVDAS system via RS-232 serial connections. A time tag is added at the beginning of each line of data in the form,

```
yy+dd:hh:mm:ss.sss [data stream from instrument]
```

where

```
yy      = two-digit year
ddd     = day of year
hh      = 2 digit hour of the day
mm      = 2 digit minute
ss.sss  = seconds
```

All times are reported in UTC.

The delimiters that separate fields in the raw data files are often spaces and commas but can be other characters such as : = @. Occasionally no delimiter is present. Care should be taken when reprocessing the data that the field's separations are clearly understood.

In the sections below a sample data string is shown, followed by a table that lists the data contained in the string.

**Underway Data** /rvdas/uw

Each section below describes a type of data file (file name extension in parentheses) followed by a typical line of data in the file. In the table(s) for each section is a description of the fields within each line of data. Note: most data files listed below will be included with each cruise's data distribution; however some types of files may be omitted if the instrument was not operating during the cruise. The available data files can be found in the /rvdas/uw directory on the distribution disc.

**Sound Velocity Probe (svp1)**

08+330:00:00:49.011 1519.35

Field	Data	Units
1	RVDAS Time tag	
2	Sound velocity in ADCP sonar well	m/s

**Meteorology (mwx1)**

There are 3 different data strings in the mwx1 data file:

MET

```
08+330:23:59:57.725 MET,12.1,-54,6.64,88.7,111.3374,0.02414567,-
0.4827508,282.9581,281.8823,1005.119
```

PUS

```
08+330:23:59:58.546 PUS,A,020,008.53,M,+337.12,+009.00,00,0F
```

SUS

```
08+330:23:59:58.779 SUS,A,017,008.76,M,+335.53,+006.35,00,02
```

**MET string**

Field	Data	Units
1	RVDAS time tag	
2	MET (string flag)	
3	Power Supply Voltage	V
4	Enclosure Relative Humidity (not currently implemented)	%
5	Air temperature	°C
6	Air Relative Humidity	%
7	PAR (photosynthetically available radiation)*	mV
8	PSP (short wave radiation)*	mV
9	PIR Thermopile (long wave radiation)*	mV
10	PIR Case Temperature	°Kelvin
11	PIR Dome Temperature	°Kelvin
12	Barometer	mBar

\*See page 21 for calculations.

## PUS string

Field	Data	Units
1	RVDAS time tag	
2	PUS (string flag)	
3	A (unit identification)	
4	Port Wind direction relative	deg
5	Port Wind speed relative	m/s
6	Units	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Unit Status (00 or 60 are good, any other value indicates fault)	
10	Check Sum	

## SUS string

Field	Data	Units
1	RVDAS time tag	
2	SUS (string flag)	
3	A (unit identification)	
4	Starboard Wind direction relative	deg
5	Starboard Wind speed relative	m/s
6	Units	
7	Sound Speed	m/s
8	Sonic Temperature	°C
9	Unit Status (00 or 60 are good, any other value indicates fault)	
10	Check Sum	

## Knudsen (knud)

99+099:00:18:19.775 HF,305.2,LF,304.3

Field	Data	Units
1	RVDAS time tag	
2	HF = High frequency flag (12 kHz)	
3	High frequency depth	meters
4	LF = Low frequency flag (3.5 kHz)	
5	Low frequency depth	meters

## Fluorometer (flr1)

This Fluorometer is not in use. The current Fluorometer goes to the hdas string.

00+019:23:59:58.061 0 0818 :: 1/19/00 17:23:17 = 0.983 (RAW) 1.2 (C)

Field	Data	Units
1	RVDAS time tag	
2	Marker 0 to 8	
3	4-digit index	
4	Date	mm/dd/yy
5	Time	hh:mm:ss
6	Signal	
7	Signal units of measurement	
8	Cell temperature (if temperature compensation package is installed)	
9	Temperature units (if temperature compensation package is installed)	

**pCO<sub>2</sub> (pco2)**

00+021:23:59:43.190 2000021.99920 2382.4 984.2 30.73 50.8 345.9 334.1 -1.70 -  
68.046 -144.446 Equil

Field	Data	Units
1	RVDAS time tag	
2	pCO <sub>2</sub> time tag (decimal is fractional time of day)	yyyyddd.ttt
3	Raw voltage (IR)	mV
4	Cell temperature	°C
5	Barometer	MBar
6	Concentration	ppm
7	Equilibrated temperature	°C
8	pCO <sub>2</sub> pressure	microAtm
9	Flow rate	ml / min
10	Source ID #	1 or 2 digits
11	Valve position	1 or 2 digits
12	Flow source (Equil = pCO <sub>2</sub> measurement)	text

**Micro-TSG (tsg1)**

08+330:23:59:40.894 5.9322, 3.34685, 34.0550, 1473.281

Field	Data	Units
1	RVDAS time tag	
2	Internal Temperature	°C
3	Conductivity	s/m
4	Salinity	PSU
5	Sound velocity	m/s

**Micro-TSG #2 (tsg2)**

08+330:23:59:40.894 5.9322, 3.34685, 34.0550, 1473.281

Field	Data	Units
1	RVDAS time tag	
2	Internal Temperature	°C
3	Conductivity	s/m
4	Salinity	PSU
5	Sound velocity	m/s

**Gravimeter (grv1)**

14+050:00:01:32.363 01:025415 00

Field	Data	Conversion	Units
1	RVDAS time tag		
2	01:		
3	Gravity count	mgal = count x 4.99407552 + bias	count
4	Error Flag		

**Engineering (eng1)**

13+079:10:22:16.035 12.26 19.68 507.4 0.3 173.3 -751.9 0 0 NAN NAN 43.2 85.7

Field	Data	Units
1	RVDAS time tag	
2	Power Supply Voltage	V
3	Internal Case Temperature	°C
4	Pump #1 flow rate (aquarium room)	L/min
5	Pump #2 flow rate (helo deck)	L/min
6	Pump #3 flow rate (hydro-lab)	L/min
7	Seismic air pressure	Lbs/sq-in
8	PIR case resistance (not currently hooked up, data is irrelevant)	Kohm
9	PIR case ratiometric output (not currently hooked up, data is irrelevant)	mV
10	Freezer #1 temperature	°C
11	Freezer #2 temperature	°C
12	Altimeter, OIS benthic (yoyo) camera; distance from the seafloor	m
13	Transmissometer, OIS benthic (yoyo) camera	%

\*See page 24 for PIR calculations.

**Hydro-DAS (hdas)**

08+330:23:59:41.877 12.15836 14.22853 368.9655 4060.69 -1 65.5 65.5 80 57

Field	Data	Units
1	RVDAS time tag	
2	Supply voltage	V
3	Panel temperature	°C
4	Fluorometer	mV
5	Transmissometer	mV
6	Sea Water Valve (-1 = stern thruster valve, 0 = moon pool valve)	
7	Flow meter 1 frequency	Hz
8	Flow meter 2 frequency	Hz
9	Flow meter 3 frequency	Hz
10	Flow meter 4 frequency	Hz

**GUV Data (pguv)**08+330:23:59:40.328 112508 235940 .000197 1.856E-1 1.116E0 4.987E-2 -1.959E-4  
1.637E0 4.153E-3 1.76E0 42.296 17.844

Field	Data	Units
1	RVDAS time tag	
2	Date	mmddy
3	Time (UTC)	hhmmss
4	Ed0Gnd	V
5	Ed0320	uW (cm <sup>2</sup> nm)
6	Ed0340	uW (cm <sup>2</sup> nm)
7	Ed0313	uW (cm <sup>2</sup> nm)
8	Ed0305	uW (cm <sup>2</sup> nm)
9	Ed0380	uW (cm <sup>2</sup> nm)
10	Ed0PAR	uE (cm <sup>2</sup> nm)
11	Ed0395	uW (cm <sup>2</sup> nm)
12	Ed0Temp	°C
13	Ed0Vin	V

**Remote Temperature (rtmp)**

07+272:00:00:15.960 -1.7870

Field	Data	Units
1	RVDAS time tag	
2	Temperature at seawater intake	°C

**Oxygen Data (oxyg)**

Internal reference salinity is set to 34 ppt. For further information on this data, contact Sharon Stammerjohn, [sstammer@ucsc.edu](mailto:ssammer@ucsc.edu).

11+011:00:21:48.109 MEASUREMENT 3835 1424 Oxygen: 334.01 Saturation:  
 90.71 Temperature: -0.78 DPhase: 37.65 BPhase: 35.95  
 RPhase: 0.00 BAmp: 212.13 BPot: 30.00 RAmp: 0.00  
 RawTem.: 788.05

Field	Data	Units
1	RVDAS time tag	
2-4	Measurement ID, Model Number, Serial Number	alphanumeric
5	Oxygen heading	text
6	Oxygen Reading	µM
7	Saturation heading	text
8	Saturation Reading	%
9	Temperature heading	text
10	Water Temperature	°C
11	Dphase heading	text
12	Dphase	Raw numeric
13	Rphase heading	Text
14	Rphase	Raw numeric
15	Bamp heading	Text
16	Bamp	Raw numeric
17	Bpot heading	Text
18	Bpot	Raw numeric
19	Ramp heading	Text
20	Ramp	Raw numeric
21	RawTem heading	Text
22	RawTemp	V

**Winch Data (bwnc, twnc, cwnc)**

13+157:04:20:20.976 ^^A03RD, 2013-06-06T04:20:29.352, BALTIC, 00000236, -  
 00000.0, -00009.3, 3306

Field	Data	Units
1	RVDAS time tag	alphanumeric
2	LAN ID	alphanumeric
3	LCI-90i Date and Time	alphanumeric
4	Winch Name	alphanumeric
5	Tension	lbs
6	Speed	m/min
7	Pay-out	m
8	Checksum	numeric

**Navigational Data** /rvdas/nav**Seapath GPS (seap)**

The Seapath GPS outputs the following data strings, four in NMEA format and two in proprietary PSXN format:

- GPZDA
- GPGGA
- GPVTG
- GPHDT
- PSXN, 20
- PSXN, 22
- PSXN, 23

**GPZDA**

```
02+253:00:00:00.772 $GPZDA,235947.70,09,09,2002,,*7F
```

Field	Data	Units
1	RVDAS time tag	
2	\$GPZDA	
3	time	hhmmss.ss
4	Day	dd
5	Month	mm
6	Year	yyyy
7	(empty field)	
8	Checksum	

**GPGGA**

```
02+253:00:00:00.938
```

```
GPGGA,235947.70,6629.239059,S,06827.668899,W,1,07,1.0,11.81,M,,M,,*6F
```

Field	Data	Units
1	RVDAS time tag	
2	\$GPGGA	
3	time	hhmmss.ss
4	Latitude	ddmm.mmmmmm
5	N or S for north or south latitude	
6	Longitude	ddmm.mmmmmm
7	E or W for east or west longitude	
8	GPS quality indicator, 0=invalid, 1=GPS SPS, 2=DGPS, 3=PPS, 4=RTK, 5=float RTK, 6=dead reckoning	
9	number of satellites in use (00-99)	
10	HDOP	x.x
9	height above ellipsoid in meters	m.mm
11	M	
12	(empty field)	
13	M	
14	age of DGPS corrections in seconds	s.s
15	DGPS reference station ID (0000-1023)	
16	Checksum	

## GPVGT

02+253:00:00:00.940 \$INVTG,19.96,T,,M,4.9,N,,K,A\*39

Field	Data	Units
1	RVDAS time tag	
2	\$GPVGT	
3	course over ground, degrees true	d.dd
4	T	
5	,	
6	M	
7	speed over ground in knots	k.k
8	N	
9	,	
10	K	
11	Mode	
12	Checksum	

## GPHDT

02+253:00:00:00.941 \$GPHDT,20.62,T\*23

Field	Data	Units
1	RVDAS time tag	
2	\$GPHDT	
3	Heading, degrees true	d.dd
4	T	
5	Checksum	

## PSXN,20

02+253:00:00:00.942 \$PSXN,20,0.43,0.43\*39

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	20	
4	Horizontal position & velocity quality: 0=normal, 1=reduced performance, 2=invalid data	
5	Height & vertical velocity quality: 0=normal, 1=reduced performance, 2=invalid data	
6	Heading quality: 0=normal, 1=reduced performance, 2=invalid data	
7	Roll & pitch quality: 0=normal, 1=reduced performance, 2=invalid data	
8	Checksum	

## PSXN,22

02+253:00:00:00.942 \$PSXN,22,0.43,0.43\*39

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	22	
4	gyro calibration value since system start-up in degrees	d.dd
5	short term gyro offset in degrees	d.dd
6	Checksum	

**PSXN,23**

02+253:00:00:02.933 \$PSXN,23,0.47,0.57,20.62,0.03\*0C

Field	Data	Units
1	RVDAS time tag	
2	\$PSXN	
3	23	
4	roll in degrees, positive with port side up	d.dd
5	pitch in degrees, positive with bow up	d.dd
6	Heading, degrees true	d.dd
7	heave in meters, positive down	m.mm
8	Checksum	

**Trimble (P-Code) GPS (PCOD)**

The Trimble GPS, which formerly output Precise Position (*P-Code*) strings, but now only outputs Standard Position (*Civilian*) strings, outputs three NMEA standard data strings:

- Position fix (GGA)
- Latitude / longitude (GLL),
- Track and ground speed (VTG)

**GGA: GPS Position Fix – Geoid/Ellipsoid**

01+319:00:04:11.193 \$GPGGA,000410.312,6227.8068,S,06043.6738,W,1,06,1.0,  
031.9,M,-017.4,M,,\*49

Field	Data	Units
1	RVDAS Time tag	
2	\$GPGGA	
3	UTC time at position	hhmmss.sss
4	Latitude	ddmm.mmm
5	North (N) or South (S)	
6	Longitude	ddmm.mmm
7	East (E) or West (W)	
8	GPS quality: 0 = Fix not available or invalid 1 = GPS, SPS mode, fix valid 2 = DGPS (differential GPS), SPS mode, fix valid 3 = P-CODE PPS mode, fix valid	
9	Number of GPS satellites used	
10	HDOP (horizontal dilution of precision)	
11	Antenna height	meters
12	M for meters	
13	Geoidal height	meters
14	M for meters	
15	Age of differential GPS data (no data in the sample string)	
16	Differential reference station ID (no data in the sample string)	
17	Checksum (no delimiter before this field)	

**GLL: GPS Latitude/Longitude**

01+319:00:04:11.272 \$GPGLL,6227.8068,S,06043.6738,W,000410.312,A\*32

Field	Data	Units
1	RVDAS Time tag	
2	\$GPGLL	
3	Latitude	degrees
4	North or South	
5	Longitude	degrees
6	East or West	
7	UTC of position	hhmmss.sss
8	Status of data (A = valid)	
9	Checksum	

**VTG: GPS Track and Ground Speed**

01+319:00:04:11.273 \$GPVTG,138.8,T,126.0,M,000.0,N,000.0,K\*49

Field	Data	Units
1	RVDAS time tag	
2	\$GPVTG	
3	Heading	degrees
4	Degrees true (T)	
5	Heading	degrees
6	Degrees magnetic (M)	
7	Ship speed	knots
8	N = knots	
9	Speed	km/hr
10	K = km per hour	
11	Checksum	

**Gyro Compass (gyr1)**

00+019:23:59:59.952 \$HEHDT 25034,-020\*73

Field	Data	Units
1	RVDAS time tag	
2	\$HEHDT	
3	Heading, Degrees True	degrees
5	Checksum	

**ADCP Course (adcp)**

00+019:23:59:59.099 \$PUHAW,UVH,-1.48,-0.51,250.6

Field	Data	Units
1	RVDAS time tag	
2	\$PUHAW	
3	UVH (E-W, N-S, Heading)	
4	Ship Speed relative to reference layer, east vector	knots
5	Ship Speed relative to reference layer, north vector	knots
6	Ship heading	degrees

**Processed Data** /process/**pCO2-merged**

00+346:23:58:20.672 2000346.9991 2398.4 1008.4 0.01 45.4 350.3 342.6 15.77 Equil -  
 43.6826 173.1997 15.51 33.90 0.33 5.28 9.05 1007.57 40.0 14.87 182.44 -1

Field	Data	Units
1	RVDAS time tag	
2	pCO <sub>2</sub> time tag (decimal is fractional time of day)	yyyddd.ttt
3	Raw voltage (IR)	mV
4	Cell temperature	°C
5	Barometer	MBar
6	Flow rate	ml / min
7	Concentration	ppm
8	pCO <sub>2</sub> pressure	microAtm
9	Equilibrated temperature	°C
10	Sea Water Temp	1 or 2 digits
11	Valve position	°C
12	Flow source (Equil = pCO <sub>2</sub> measurement)	text
13	RVDAS latitude	degrees
14	RVDAS longitude	degrees
15	TSG external temperature	°C
16	TSG 1 salinity	PSU
17	Fluorometer	V
18	RVDAS true wind speed	m/s
19	RVDAS true wind direction	degrees
20	Barometric Pressure	mBars
21	Uncontaminated seawater pump flow rate	l/min
22	Speed over ground	knots
23	Course made good	degrees
24	Oxygen	µM
25	TSG 2 internal temperature	°C
26	TSG 2 salinity	PSU
27	TSG 1 internal temperature	°C
28	H2O Input Source	-1 stern thruster 0 moonpool

## Calculations

The file `instrument.coeff` located in the `/` directory contains the calibration factors for shipboard instruments. This was the file used by the RVDAS processing software.

## PAR

Coefficients `parc1` and `parcv` for this cruise can be found in the `instrument.coeff` file as the variable labeled PAR, respectively. Variable `par` is the raw data in mV, as described in the “`mwx1`” file description. The calibration scale and probe offset dark are values taken from the PAR Cal Sheet.

```
par = raw data mV
calibration scale = 5.8644 V/(μEinstiens/cm2sec)
parc1 = 1 / scale = .17
probe offset dark = -.1 mV
parcv = dark x 1000 mV/V = -0.0001 V
((par / 1000 mV/V) - parcv) x parc1 x 10000 cm2/m2 = μEinstiens/m2sec
```

Calculations (extracted from the C code):

```
/* Convert from mV to V */
par /= 1000;
/* (par V - vdark V) / Calibration Scale Factor V/uE/cm2sec */
parCalc = (par - parcv) * parc1 * 10000;
```

## PSP

Coefficient `pspCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled PSP1. Variable `psp` is the raw data in mV, as described in the “`mwx1`” file description.

```
psp = raw data mV
calibration scale = pspCoeff x 10^-6 V/(W/m2)
psp / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* Convert from mV to W/m^2 */
pspCalc = (psp * 1000 / pspCoeff);
```

## PIR

Coefficient `pirCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled `PIR1`. Variable `pir_thermo` is the raw data in mV, `pir_case` is the PIR case temperature in Kelvins and `pir_dome` is the PIR dome temperature in Kelvins, as described in the “`mw1`” file description. Hard-coded “C” coefficients are shown below:

```
Dome constant = 3.5
```

```
Sigma = 5.6704e-8
```

```
pir_thermo = raw data mV
```

```
calibration scale = pirCoeff x 10-6 V/(W/m2)
```

```
pir_thermo / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* convert mV to W/m2 */  
pirCalc = (pir_thermo * 1000 / pirCoeff)  
/* correct for case temperature */  
pirCalc += sigma * pow(pir_case,4)  
/* correct for dome temperature */  
pirCalc -= 3.5 * sigma * (pow(pir_dome, 4) - pow(pir_case, 4))
```

## Acquisition Problems and Events

This section lists problems with acquisition noted during this cruise including instrument failures, data acquisition system failures and any other factor affecting this data set. The format is ddd:hh:mm (ddd is year-day, hh is hour, and mm is minute). Times are reported in UTC.

Start	End	Description
030:04:02		Data collection started at EEZ -46.475, 147.549
	073:00:35	Data collection cut off for EEZ -46.772, 144.415
051:03:41		Oxygen Sensor was turned off on day 051 due to unreadable strings passed to RVDAS
048:05:37		Switched from the Knudsen 3260 to the Knudsen 320B/R Deck boxes
052:00:01		Switched from the Knudsen 320B/R to the Knudsen 3260 Deck boxes
066:06:33	066:07:16	short power outages
071:09:19		During a rough sea state during the transit a Gyroscope in the Gravimeter failed. This will be repaired once the NBP gets to port.

**Appendix: Sensors and Calibrations**

<i>Sensor</i>	<i>Serial Number</i>	<i>Last Cal.</i>	<i>Comments</i>
<b>Meteorology &amp; Radiometers</b>			
Stbd Anemometer (Gill US)	847014	9/29/2010	Installed 11/17/2010
Port Anemometer (Gill US)	924057	11/18/09	Installed 3/5/2010
Barometer	BP00872	11/29/2012	Installed 1/28/2014
Humidity/Wet Temp	06135	11/29/2012	Installed 9/11/2013
PIR	32845F3	7/17/2013	Installed 1/28/2014
PSP	32850F3	8/15/2013	Installed 1/28/2014
Mast PAR	6357	12/27/2012	Installed 9/11/2013
GUV (Mast)	25110203114	12/18/2012	Installed 9/11/2013
<b>Underway</b>			
Micro-TSG #1 (until 3/4/13)	4546167-0242	12/29/2012	Installed 5/9/2013
Micro-TSG #2	4566350-0389	10/20/2011	Installed 9/7/2012
Digital Remote Temp	3849120-0178	9/21/2012	Installed 5/9/2013
Oxygen Optode	3835-1424	10/21/2010	Installed 12/30/2010
Fluorometer	AFL-016D	8/22/2012	Installed 9/11/2013
Transmissometer	CST-557DR	8/28/2013	Installed 1/28/2014
<b>CTD</b>			
CTD Fish	91480	12/18/2012	Installed 1/28/2014
CTD Fish Pressure	53952	12/18/2012	Installed 1/28/2014
CTD Deck Unit	11P19858-0768	N/A	Installed 1/28/2014
Slip-Ring Assembly	1.406	N/A	Installed 1/28/2014
Carousel Water Sampler	3214153-0140	N/A	Installed 1/28/2014
Pump (primary)	051627 3.0K	12/23/2012	Installed 1/28/2014
Pump (secondary)	051626 3.0K	12/23/2012	Installed 1/28/2014
Temperature (primary)	03P2308	6/28/2013	Installed 1/28/2014
Temperature (secondary)	03P2299	6/12/2013	Installed 1/28/2014
Conductivity (primary)	042513	2/26/2013	Installed 1/28/2014
Conductivity (secondary)	041798	6/21/2013	Installed 1/28/2014
Dissolved Oxygen (primary)	430161	6/12/2013	Installed 1/28/2014
Dissolved Oxygen (primary)	430080	2/13/2013	Installed 1/28/2014
Fluorometer	AFLD-0011	7/17/2013	Installed 1/28/2014
Transmissometer	CST-0889	9/5/13	Installed 1/28/2014
Altimeter	49432	N/A	Installed 1/28/2014

**Mast Barometer**

**R.M. Young Company**  
2801 Aero Park Drive  
Traverse City, Michigan 49686 USA



**CALIBRATION REPORT**  
**Barometric Pressure**

Customer: *Lockheed Martin Maritime Systems & Sensors*

Test Number: 2060-01B

Customer PO: 4900027957

Test Date: 29 November 2012

Sales Order: 2973

Test Sensor:

Model: 61201      Serial Number: *BP00872*  
Description: Barometric Pressure Sensor

Report of calibration comparison of test barometric pressure sensor with National Institute of Standards and Technology traceable standard pressure calibrator at five pressures in the R.M. Young Company controlled pressure facility. Calibration accuracy  $\pm 1.0$  hPa.

Reference Pressure (hPa)	Voltage Output (millivolts)	Indicated (1) Pressure (hPa)
800.0	-1	800.0
875.0	1251	875.0
950.0	2501	950.0
1025.0	3749	1024.9
1100.0	4996	1099.7

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument  
Druck Pressure Controller Model DPI515  
Fluke Multimeter Model 8060A

Serial # NIST Test Reference  
51500497      UKAS Lab 0221  
4865407      234027

Tested By: *ECherny*

M E T E O R O L O G I C A L I N S T R U M E N T S  
Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: youngusa.com  
ISO 9001:2008 CERTIFIED



R.M. Young Company  
2801 Aero Park Drive  
Traverse City, Michigan 49686 USA



**CALIBRATION REPORT**  
**Relative Humidity**

Customer: *Lockheed Martin Maritime Systems & Sensors*

Test Number: 2044-02R  
Test Date: 29 November 2012

Customer PO: 4900027957  
Sales Order: 2973

Test Sensor:  
Model: 41372LC      Serial Number: 7306135  
Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test relative humidity sensor with National Institute of Standards and Technology traceable standard relative humidity sensor at five humidity levels in the R.M. Young Company controlled humidity chamber facility. Calibration accuracy  $\pm 2.0$  %.

Reference Humidity (%)	Current Output (milliamps)	Indicated (1) Humidity (%)
10.0	5.9	12.1
30.0	9.0	31.2
50.0	12.4	52.3
69.9	15.4	71.0
89.9	18.1	88.1

(1) Calculated from voltage output.

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument  
Vaisala Humidity Sensor Model 35AC  
Fluke Multimeter Model 6080A

Serial # NIST Test Reference  
N475040      TN 2661E2  
4855407      234027

Tested By: *E. Channing*

**METEOROLOGICAL INSTRUMENTS**  
Tel: 231-948-3880 Fax: 231-946-4772 Email: [met@youngusa.com](mailto:met@youngusa.com) Website: [youngusa.com](http://youngusa.com)  
ISO 9001:2008 CERTIFIED

**Mast Temperature Sensor**

R.M. Young Company  
2801 Aero Park Drive  
Traverse City, Michigan 49686 USA



**CALIBRATION REPORT**  
**Temperature**

Customer: *Lockheed Martin Maritime Systems & Sensors*

Test Number: 2044-021

Customer PO: 4900027957

Test Date: 29 November 2012

Sales Order: 2973

Test Sensor:

Model: 41372LC      Serial Number: TS03135  
Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test temperature sensor with National Institute of Standards and Technology traceable standard thermometers at three temperatures in the R.M. Young Company controlled temperature calibration bath facilities. Calibration accuracy  $\pm 0.1^\circ$  Celsius.

Bath Temperature (degrees C)	Current Output (milliamps)	Indicated (1) Temperature (degrees C)
-49.86	4.023	-49.56
0.03	12.008	0.05
50.18	20.029	50.16

(1) Calculated from current output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument	Serial #	NIST Test Reference
Brooklyn Thermometer Model 43-FC	8006-118	204355
Brooklyn Thermometer Model 22332-D5-FC	25071	249753
Brooklyn Thermometer Model 2X400-D7-FC	77532	228030
Keithley Multimeter Model 191	15232	234027

Tested By: E. Channing

**METEOROLOGICAL INSTRUMENTS**  
Tel: 231-946-3060 Fax: 231-946-4772 Email: metesales@youngusa.com Website: youngusa.com  
ISO 9001:2008 CERTIFIED

**Mast PIR**



**THE EPPLEY LABORATORY, INC.**

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840  
Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

**STANDARDIZATION OF  
EPPLEY PRECISION INFRARED RADIOMETER  
Model PIR**

Serial Number: 32845F3

Resistance: 712  $\Omega$  at 23°C

Temperature Compensation Range: -20° to +40°C

This pyrgometer has been compared against Eppley's Blackbody Calibration System under radiation intensities of approximately 200 watts meter<sup>-2</sup> and an average ambient temperature of 30°C as measured by Standard Omega Temperature Probe, RTD#1.

As a result of a series of comparisons, it has been found to have a sensitivity of:

$$4.08 \times 10^{-6} \text{ volts/watts meter}^{-2}$$

The calculation of this constant is based on the fact that the relationship between radiation intensity and emf is rectilinear to intensities of 700 watts meter<sup>-2</sup>. This radiometer is linear to within  $\pm 1.0\%$  up to this intensity.

The calibration of this instrument is traceable to the International Practical Temperature Scale (IPTS) through a precision low-temperature blackbody.

Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy. Unless otherwise stated in the remarks section below or on the Sales Order, the results are "AS FOUND / AS LEFT".

Shipped to: LMP4 ISGS N.S.F.  
Port Hueneme, CA

Date of Test: July 17, 2013

S.O. Number: 63850  
Date: July 18, 2013

In Charge of Test: *Oliver J. Slattery*  
Reviewed by: *Thomas J. Kirk*

Remarks:



## THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840  
Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

### Calibration Certificate

---

**Instrument:** Precision Spectral Pyranometer, Model PSP, Serial Number 32850F3

**Procedure:** This pyranometer was compared in Eppley's Integrating Hemisphere according to procedures described in *ISO 9847 Section 5.3.1* and Technical Procedure, TP01 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

**Transfer Standard:** Eppley Precision Spectral Pyranometer, Model PSP, Serial Number 21231F3

**Results:**  
**Sensitivity:**  $S = 7.68 \mu V / Wm^{-2}$   
**Uncertainty:**  $U_{95} = \pm 0.91\%$  (95% confidence level, k=2)  
**Resistance:** 706  $\Omega$  at 23°C

**Date of Test:** August 7, 2013

**Traceability:** This calibration is traceable to the World Radiation Reference (WRR) through comparisons with Eppley's AHF standard self-calibrating cavity pyrheliometers which participated in the Eleventh International Pyrheliometric Comparisons (IPC XI) at Davos, Switzerland in September-October 2010. Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

**Due Date:** Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

**Customer:** LMP4 ISGS  
Port Hueneme, CA

**Signatures:**  
D. GIENTY In Charge of Test: Thomas J. Kuh Reviewed by:

**Eppley SO** 63884

**Date of Certificate** August 15, 2013

**Remarks:**

**Biospherical Instruments Inc.**

CALIBRATION CERTIFICATE

Calibration Date: 12/27/2012  
Model Number: QSR-240  
Serial Number: 6357  
Operator: TPC  
Standard Lamp: V-031(3/7/12)  
Probe Excitation Voltage Range: 6 to 18 VDC(+)  
Output Polarity: Positive

Probe Conditions at Calibration(in air):

Calibration Voltage: 6 VDC(+)  
Probe Current: 7.2 mA

Probe Output Voltage:

Probe Illuminated: 98.3 mV  
Probe Dark: 1.0 mV  
Probe Net Response: 97.3 mV  
RG780: 1.0 mV

Corrected Lamp Output:

Output In Air (same condition as calibration):

1.044E+16 quanta/cm<sup>2</sup>sec  
0.01733 uE/cm<sup>2</sup>sec

Calibration Scale Factor:

*(To calculate irradiance, divide the net voltage reading in Volts by this value.)*

Dry: 9.3240E-18 V/(quanta/cm<sup>2</sup>sec)  
5.6149E+00 V/(uE/cm<sup>2</sup>sec)

Notes:

1. Annual calibration is recommended
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).
3. The collector should be cleaned frequently with alcohol.
4. Calibration was performed with customer cable, when available.

QSR240R.05/24/95





Biospherical Instruments Inc.

**System Calibration Certificate**

THE INSTRUMENTS REFERENCED BELOW WERE FACTORY TESTED AND CALIBRATED BY

**BIOSPHERICAL INSTRUMENTS INC.**  
 5340 Riley Street  
 San Diego, California 92110 USA

Instruments: GUV-2511 No 25110203114

**Optical Calibrations:**

**NIST Traceability.** For wavelengths longer than 313 nm, the specific instruments cited here were calibrated using a 1000W FEL #V-031(3/7/12) following procedures and standards traceable to NIST Standard of Spectral Irradiance F616. Traceability paths and all procedures for all calibrated lamps and associated apparatus (shunts, power supplies, DMMs, etc) are maintained following calibration methodologies per National Bureau of Standards (US) (NBS) Special Publication 250-20 Spectral Irradiance Calibrations (1987) and NBS Publication 594-13 Optical Radiation Measurements: The 1973 Scale of Spectral Irradiance (1977).

**Solar Calibrations.** Lamp calibrations are problematic for solar UV measurements (wavelengths below 320 nm) because the solar spectrum is radically different from the lamp spectrum and changes greatly as a function of wavelength. Solar calibrations are achieved through direct comparison with measurements of a high resolution scanning spectroradiometer in San Diego (SUV-100), which is part of the National Science Foundation's UV Monitoring Network. The SUV-100 instrument has a bandwidth of 1 nm. Calibrated filter radiometer data therefore report spectral irradiance at the channel's nominal wavelengths with a bandwidth of 1 nm. Solar calibrations are typically accurate to within ±10% for solar zenith angles smaller than 75°. At larger solar zenith angles, UV channels have a greater uncertainty due to the rapid change of the solar UV spectrum.

Note that this certificate contains a subset of the information delivered in the calibration database 25110203114v7.mdb. This database is required for operation of this system using Biospherical Instruments Inc.'s Logger® software.



Biospherical Instruments Inc.

**GUV-2511 Calibration Certificate**

System Serial Number		25110203114		Date of Calibration		12/18/2012	
Calibration database		25110203114v7.mdb		Date of Certificate		12/18/2012	
DAS SN		0069		Standard of Spectral Irradiance		V-031(3/7/12)	
Microprocessor Tag Number		4		Operator		TC	

Monochromatic Channels	Address	Wavelength [nm]	Responsivity	ScaleSmall	ScaleMedium	ScaleLarge	OffsetSmall	OffsetMedium	OffsetLarge	Measurement Units
			[Amps per $\mu\text{W}/(\text{cm}^2\text{-nm})$ ]	[Volts per $\mu\text{W}/(\text{cm}^2\text{-nm})$ ]	[Volts per $\mu\text{W}/(\text{cm}^2\text{-nm})$ ]	[Volts per $\mu\text{W}/(\text{cm}^2\text{-nm})$ ]	[Volts]	[Volts]	[Volts]	
E00320	2	320	2.3101E-10	2.3563E-05	6.8841E-03	2.1728E+00	6.8000E-05	7.1000E-05	5.8600E-04	$\mu\text{W}/(\text{cm}^2\text{-nm})$
E00340	6	340	1.8298E-10	1.8664E-05	5.4528E-03	1.8705E+00	7.8000E-05	9.0000E-05	8.2900E-04	$\mu\text{W}/(\text{cm}^2\text{-nm})$
E00313	8	313	2.3000E-10	2.3428E-05	6.8446E-03	2.4027E+00	9.2400E-04	9.2000E-04	-1.3310E-03	$\mu\text{W}/(\text{cm}^2\text{-nm})$
E00305	10	305	1.0300E-11	1.0554E-06	3.0836E-04	1.0599E-01	3.6800E-04	3.7000E-04	1.1140E-03	$\mu\text{W}/(\text{cm}^2\text{-nm})$
E00380	12	380	8.0945E-11	8.2564E-06	2.4122E-03	7.7368E-01	2.7800E-04	2.7600E-04	-1.0800E-04	$\mu\text{W}/(\text{cm}^2\text{-nm})$
E00395	18	395	2.8798E-10	2.9374E-05	8.5818E-03	2.7127E+00	3.8900E-04	3.9300E-04	1.4470E-03	$\mu\text{W}/(\text{cm}^2\text{-nm})$

Broadband Channels	Address	Wavelength [nm]	Responsivity	ScaleSmall	ScaleMedium	ScaleLarge	OffsetSmall	OffsetMedium	OffsetLarge	Measurement Units
			[Amps per $\mu\text{E}/(\text{cm}^2\text{-s})$ ]	[Volts per $\mu\text{E}/(\text{cm}^2\text{-s})$ ]	[Volts per $\mu\text{E}/(\text{cm}^2\text{-s})$ ]	[Volts per $\mu\text{E}/(\text{cm}^2\text{-s})$ ]	[Volts]	[Volts]	[Volts]	
Ed0PAR	13	400-700	1.7033E-05	1.7374E+00	5.0759E+02	1.7939E+05	5.7300E-04	5.7100E-04	-4.7800E-04	$\mu\text{E}/(\text{cm}^2\text{-sec})$

Auxiliary Channels	Address	Wavelength [nm]	Responsivity	ScaleS	ScaleM	ScaleL	OffsetS	OffsetM	OffsetL	Measurement Units
			[Amps per $\mu\text{E}/(\text{cm}^2\text{-s})$ ]	[Volts]	[Volts]	[Volts]	[Volts]	[Volts]	[Volts]	
Ed0Temp	22	0	1.0000E+00	1.0000E-02	1.0000E-02	1.0000E-02	0.0000E+00	0.0000E+00	0.0000E+00	C
Ed0Vin	27	0	1.0000E+00	-2.5000E-01	-2.5000E-01	-2.5000E-01	0.0000E+00	0.0000E+00	0.0000E+00	V

© Biospherical Instruments Inc., 5340 Riley Street, San Diego, California 92110 USA. Contact [support@biospherical.com](mailto:support@biospherical.com) for more information.

## Underway Oxygen Sensor



AANDERAA DATA INSTRUMENTS

## CALIBRATION CERTIFICATE

Form No. 622, Dec 2005  
Page 1 of 2Sensing Foil Batch No: 5009  
Certificate No:Product: Oxygen Optode 3835  
Serial No: 1424  
Calibration Date: 21 October 2010

This is to certify that this product has been calibrated using the following instruments:

Calibration Bath model FNT  
ASL Digital Thermometer model F250321-1-40  
Serial: 6792/06**Parameter: Internal Temperature:****Calibration points and readings:**

Temperature (°C)	1.17	12.12	24.11	36.08
Reading (mV)	730.09	383.95	-11.29	-379.10

**Giving these coefficients**

Index	0	1	2	3
TempCoef	2.37613E01	-3.08128E-02	2.84735E-06	-4.15311E-09

**Parameter: Oxygen:**

	O2 Concentration	Air Saturation
Range:	0-500 $\mu\text{M}$ <sup>1)</sup>	0 - 120%
Accuracy <sup>1)</sup> :	< $\pm 8\mu\text{M}$ or $\pm 5\%$ (whichever is greater)	$\pm 5\%$
Resolution:	< 1 $\mu\text{M}$	< 0.4%
Settling Time (63%):	< 25 seconds	

**Calibration points and readings<sup>2)</sup>:**

	Air Saturated Water	Zero Solution (Na <sub>2</sub> SO <sub>3</sub> )
Phase reading (°)	3.27669E+01	6.65595E+01
Temperature reading (°C)	9.90918E+00	2.04774E+01
Air Pressure (hPa)	9.76884E+02	

**Giving these coefficients**

Index	0	1	2	3
PhaseCoef	-4.44928E00	1.17131E00	0.00000E00	0.00000E00

<sup>1)</sup> Valid for 0 to 2000m (6562ft) depth, salinity 33 - 37ppt<sup>2)</sup> The calibration is performed in fresh water and the salinity setting is set to: 0

AANDERAA DATA INSTRUMENTS AS

0351 BERGEN, NORWAY Tel. +47 55 60 46 00 Fax. +47 55 60 48 01 E-mail: info@aadi.no Web: http://www.aadi.no



AANDERAA DATA INSTRUMENTS

# CALIBRATION CERTIFICATE

Form No. 622, Dec 2005  
Page 2 of 2Sensing Foil Batch No: 5009  
Certificate No:Product: Oxygen Optode 3835  
Serial No: 1424  
Calibration Date: 21 October 2010

## SR10 Scaling Coefficients:

At the SR10 output the Oxygen Optode 3830 can give either absolute oxygen concentration in  $\mu\text{M}$  or air saturation in %. The setting of the internal property "Output"<sup>3)</sup>, controls the selection of the unit. The coefficients for converting SR10 raw data to engineering units are fixed.

Output = -1	Output = -2
A = 0	A = 0
B = 4.883E-01	B = 1.465E-01
C = 0	C = 0
D = 0	D = 0
Oxygen ( $\mu\text{M}$ ) = A + BN + CN2 + DN3	Oxygen (%) = A + BN + CN2 + DN3

<sup>3)</sup>The default output setting is set to -1

Date: 22 October 2010

Sign:

Tor Ole Kvaloy, Calibration Engineer

AANDERAA DATA INSTRUMENTS AS

5351 BERGEN, NORWAY Tel. +47 55 60 46 00 Fax. +47 55 60 46 01 E-mail: info@aadi.no Web: http://www.aadi.no



# CALIBRATION CERTIFICATE

Form No. 621, Dec 2005

AANDERAA DATA INSTRUMENTS

Certificate No: 3853\_5009\_40331  
Batch No: 5009

Product: O2 Sensing Foil PSt3 3853  
Calibration Date: 2 June 2010

### Calibration points and phase readings (degrees)

Temperature (°C)		3.97	10.93	20.15	29.32	38.39
Pressure (hPa)		977.00	977.00	977.00	977.00	977.00
O2 in % of O2+N2	0.00	73.18	72.63	71.62	70.72	69.77
	1.00	68.01	67.02	65.42	63.92	62.31
	2.00	64.39	63.16	61.20	59.44	57.57
	5.00	55.80	54.16	51.76	49.56	47.45
	10.00	46.27	44.47	41.97	39.75	37.69
	20.90	35.09	33.38	31.14	29.24	27.56
	30.00	29.85	28.30	26.31	24.64	23.19

Giving these coefficients <sup>1)</sup>

Index	0	1	2	3
C0 Coefficient	4.53793E+03	-1.62595E+02	3.29574E+00	-2.79285E-02
C1 Coefficient	-2.50953E+02	8.02322E+00	-1.58398E-01	1.31141E-03
C2 Coefficient	5.66417E+00	-1.59647E-01	3.07910E-03	-2.46265E-05
C3 Coefficient	-5.99449E-02	1.48326E-03	-2.82110E-05	2.15156E-07
C4 Coefficient	2.43614E-04	-5.26759E-06	1.00064E-07	-7.14320E-10

<sup>1)</sup> Ask for Form No 621S when this O2 Sensing Foil is used in Oxygen Sensor 3830 with Serial Numbers lower than 184.

Date: 11/4/2010

Sign:

Tor Odd Kvaloy - Calibration Engineer

AANDERAA DATA INSTRUMENTS AS

5351 BERGEN, NORWAY

Tel. +47 55 60 46 00

Fax. +47 55 60 46 01

E-mail: info@aadi.no

Web: http://www.aadi.no

**Underway Micro-TSG number 1**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0747  
 CALIBRATION DATE: 29-Dec-12

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**COEFFICIENTS:**

g = -9.992296e-001  
 h = 1.524743e-001  
 i = -4.722991e-004  
 j = 6.065458e-005

CPcor = -9.5700e-008  
 CTcor = 3.2500e-006  
 WBOTC = -0.0000e+000


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2566.82	0.00000	0.00000
1.0000	34.8118	2.97562	5119.70	2.97561	-0.00001
4.5000	34.7917	3.28263	5313.24	3.28264	0.00001
15.0000	34.7487	4.26420	5888.60	4.26420	0.00000
18.5000	34.7394	4.60927	6077.64	4.60927	0.00001
24.0000	34.7293	5.16711	6371.04	5.16711	-0.00001
29.0000	34.7238	5.68887	6633.34	5.68886	-0.00001
32.5000	34.7207	6.06120	6814.13	6.06121	0.00001

$f = \text{INST FREQ} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p)$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

Residual = instrument conductivity - bath conductivity



### Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0747  
 CALIBRATION DATE: 29-Dec-12


SBE 45 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS  
 a0 = 4.555848e-005  
 a1 = 2.733778e-004  
 a2 = -2.324224e-006  
 a3 = 1.499077e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	649816.0	1.0000	0.0000
4.5000	554883.5	4.5000	-0.0000
15.0000	352327.7	15.0000	-0.0000
18.5000	304717.7	18.5000	0.0000
24.0000	244011.0	24.0000	0.0000
29.0000	200602.2	29.0000	-0.0000
32.5000	175478.8	32.5000	0.0000

$$\text{Temperature ITS-90} = 1/[a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]] - 273.15 \text{ (}^\circ\text{C)}$$

$$\text{Residual} = \text{instrument temperature} - \text{bath temperature}$$



**Underway Micro-TSG number 2**

8

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA  
 Phone: (+1) 425-643-8388 Fax: (+1) 425-643-9954 Email: seabird@seabird.com

*Micro TSG*

SENSOR SERIAL NUMBER: 0589  
 CALIBRATION DATE: 20-Oct-11

SBB-45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(3S, 5,0) = 1.2014 Siemens/meter

COEFFICIENTS:

a = -9.366211e-001  
 b = 1.453843e-001  
 c = -6.022911e-004  
 j = 3.189313e-002

CPcor = 3.5700e-006  
 CTcor = 3.2500e-008  
 KDCTC = 1.2700e-007

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2004.51	0.00000	0.00000
1.0000	24.9210	2.98400	5241.92	2.98436	0.00000
4.5000	24.9000	3.79188	5443.39	3.79188	-0.00000
14.3990	24.8864	4.27000	6033.34	4.27000	0.00000
18.4990	24.8867	4.62155	6237.28	4.62155	-0.00000
27.0000	24.8804	5.18115	6528.11	5.18115	-0.00000
29.0000	24.8277	5.70357	6793.07	5.70357	-0.00000
32.5000	24.8211	6.07070	6962.11	6.07070	0.00000

$\delta = \text{INST FREQ}^2 \cdot \text{sq}(\text{a}) + \text{WBOTC} \cdot \text{p} / 10000$


Conductivity =  $(g + hf^2 + \alpha^2 + jf^4) / (1 - \delta + \epsilon \text{p})$  Siemens/meter

$\epsilon = (\text{temperature}[^{\circ}\text{C}])$ ;  $\text{p} = \text{pressure}[\text{decibars}]$ ;  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$

Residual = Instrument conductivity - bath conductivity

Date, Slope Correction

20-Oct-11 1.0000000



### Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-843-9866 Fax: (+1) 425-843-0654 Email: seabird@seabird.com

*Micro TSG*

SENSOR SERIAL NUMBER: 0589  
CALIBRATION DATE: 20 Oct 11

SBE 45 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE


**ITS-90 COEFFICIENTS**

a0 = 3.173863e-003  
a1 = 2.856137e-004  
a2 = -1.828637e-006  
a3 = 1.127174e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	828314.3	0.0000	0.0001
4.5000	706709.4	4.4999	-0.0002
14.8999	447305.3	15.0000	0.0001
18.4999	386307.8	18.5000	0.0001
24.0000	308342.0	23.9999	-0.0001
28.0000	253727.3	28.0000	-0.0000
32.0000	221749.3	32.0000	0.0000

Temperature ITS-90 =  $1/(a0 + a1(bath) + a2(bath)^2 + a3(bath)^3) - 273.15$  (°C)

Residual = instrument temperature - bath temperature



**Underway Digital Remote Temperature**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0178  
 CALIBRATION DATE: 21-Sep-12


SBE 38 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS  
 a0 = -4.740793e-005  
 a1 = 2.820902e-004  
 a2 = -2.754939e-006  
 a3 = 1.681819e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.50000	750879.8	-1.49997	0.00003
1.00000	671250.6	0.99996	-0.00004
4.50000	575382.9	4.49998	-0.00002
8.00000	494802.5	7.99999	-0.00001
11.50000	426843.9	11.50002	0.00002
15.00000	369343.4	15.00002	0.00002
18.50000	320537.2	18.49998	-0.00002
21.99990	278981.8	21.99999	0.00009
25.50000	243494.9	25.49993	-0.00007
28.99990	213101.3	28.99982	-0.00008
32.49990	186993.9	32.49996	0.00006

Temperature ITS-90 =  $1/(a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]) - 273.15$  (°C)

Residual = instrument temperature - bath temperature



PO Box 518  
620 Applegate St  
Philomath OR 97370



(541) 929-5650  
Fax (541) 929-5277  
<http://www.wetlabs.com>

### Chlorophyll Fluorometer Characterization in Uranine liquid Proxy (new method)

Date: 08/22/12  
Serial #: AFL-016D  
Tech: dcm

Dark Counts 0.152 volts  
CEV 1.195 volts  
SF 25.311

FSV 4.61 volts

Linearity: 0.999 R<sup>2</sup> (0–1.5 volts)  
0.995 R<sup>2</sup> (0– 5.45 volts)

#### Notes:

**Dark Counts:** Signal output of the meter in clean water with black tape over detector.

**CEV** is the chlorophyll equivalent voltage. This value is the signal output of the fluorometer when using a Uranine dye fluorescent proxy that has been determined to be approximately equivalent to 26.4 µg/l of a *Thalassiosira weissflogii* phytoplankton culture.

**SF** is the scale factor used to derive chlorophyll concentration from the signal voltage output of the fluorometer. The scale factor is determined by using the following equation:  
 $SF = (26.4) / (CEV - \text{dark})$ .

**FSV** is the maximum signal voltage output that the fluorometer is capable of.

Chlorophyll concentration expressed in µg/l (mg/m<sup>3</sup>) can be derived by using the following equation: (µg/l) = (V<sub>measured</sub> - dark) \* SF

The relationship between fluorescence and chlorophyll-*a* concentrations in-situ is high variable. The scale factor listed on this document was determined by using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer you must perform secondary measurements on the populations of interest. This is typically done using extraction based measurement techniques on discrete samples. For additional information on determination of chlorophyll concentration see [ Standard Methods For The Examination Of Water And Wastewater] part 10200 H published jointly by: American Public Health Association, American Water Works Association and Water Environment Federation.

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## C-Star Calibration

Date	August 28, 2013	S/N#	CST-557DR	Pathlength	25cm
		<b>Analog output</b>	<b>Digital output</b>		
V <sub>d</sub>		0.009 V	0 counts		
V <sub>air</sub>		4.760 V	15596 counts		
V <sub>ref</sub>		4.700 V	15399 counts		
Temperature of calibration water					21.2 °C
Ambient temperature during calibration					21.8 °C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x, in meters):  $Tr = e^{-cx}$

To determine beam transmittance:  $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient:  $c = -1/x * \ln (Tr)$

V<sub>d</sub> Meter output with the beam blocked. This is the offset.

V<sub>air</sub> Meter output in air with a clear beam path.

V<sub>ref</sub> Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V<sub>ref</sub>.

Ambient temperature: meter temperature in air during the calibration.

V<sub>sig</sub> Measured signal output of meter.

**CTD Fish and Pressure Sensor**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1480  
CALIBRATION DATE: 18-Dec-12

SBE9plus PRESSURE CALIBRATION DATA  
10000 psia S/N 53952

DIGIQUARTZ COEFFICIENTS:

C1 = -5.561704e+004  
C2 = 4.302402e-001  
C3 = 1.582810e-002  
D1 = 4.708200e-002  
D2 = 0.000000e+000  
T1 = 3.029296e+001  
T2 = -2.122954e-004  
T3 = 4.352450e-006  
T4 = 2.626550e-009  
T5 = 0.000000e+000


AD590M, AD590B, SLOPE AND OFFSET:

AD590M = 1.16300e-002  
AD590B = -8.63457e+000  
Slope = 0.99999  
Offset = -3.0213 (dbars)

PRESSURE (PSIA)	INST OUTPUT(Hz)	INST TEMP(C)	INST OUTPUT (PSIA)	CORRECTED INST OUTPUT (PSIA)	RESIDUAL (PSIA)
14.547	33019.50	21.4	19.466	15.084	0.537
2014.689	33606.67	21.7	2018.592	2014.196	-0.493
4014.621	34182.17	21.8	4018.885	4014.476	-0.145
6014.640	34746.23	21.9	6019.053	6014.631	-0.009
8014.742	35299.59	21.9	8019.715	8015.280	0.537
10014.990	35842.18	22.0	10018.718	10014.268	-0.722
8014.780	35299.62	22.1	8019.806	8015.370	0.590
6014.719	34746.31	22.2	6019.301	6014.878	0.159
4014.689	34182.23	22.2	4019.027	4014.618	-0.070
2014.710	33606.71	22.3	2018.677	2014.281	-0.429
14.555	33019.38	22.4	18.981	14.598	0.043

Residual = corrected instrument pressure - reference pressure

Date, Avg Offset (psia)



**CTD Temperature (Primary)**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2308  
CALIBRATION DATE: 28-Jun-13

SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.34531719e-003  
h = 6.44991551e-004  
i = 2.35185807e-005  
j = 2.23479362e-006  
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121230e-003  
b = 6.02583850e-004  
c = 1.63930551e-005  
d = 2.23636632e-006  
f0 = 2906.476


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	2906.476	-1.5000	0.00000
1.0000	3073.288	1.0000	0.00001
4.5000	3318.316	4.5000	-0.00001
8.0000	3577.096	8.0000	-0.00004
11.5000	3850.006	11.5000	0.00003
15.0000	4137.394	15.0001	0.00005
18.5000	4439.604	18.5000	-0.00001
22.0000	4756.983	22.0000	-0.00003
25.5000	5089.855	25.5000	-0.00003
28.9999	5438.527	28.9999	0.00005
32.5000	5803.307	32.5000	-0.00001

Temperature ITS-90 =  $1/[g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]] - 273.15$  (°C)

Temperature IPTS-68 =  $1/[a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]] - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature



**CTD Temperature (Secondary)**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2299  
CALIBRATION DATE: 12-Jun-13

SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.33219965e-003  
h = 6.44461471e-004  
i = 2.41492147e-005  
j = 2.44706389e-006  
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121247e-003  
b = 6.02091743e-004  
c = 1.64917777e-005  
d = 2.44867224e-006  
f0 = 2848.641


BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	2848.641	-1.5000	-0.00001
1.0000	3012.273	1.0000	0.00000
4.4999	3252.647	4.5000	0.00007
8.0000	3506.532	7.9999	-0.00005
11.5000	3774.292	11.5000	-0.00003
15.0000	4056.268	14.9999	-0.00007
18.4999	4352.809	18.4999	0.00004
22.0000	4664.250	22.0001	0.00008
25.5000	4990.903	25.5003	0.00032
29.0000	5332.948	28.9994	-0.00057
32.5000	5690.953	32.5002	0.00023

Temperature ITS-90 =  $1 / \{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$  (°C)

Temperature IPTS-68 =  $1 / \{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature



**CTD Conductivity (Primary)**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2513  
 CALIBRATION DATE: 26-Jun-13

SBE4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**GHIJ COEFFICIENTS**

g = -1.05846412e+001  
 h = 1.63289463e+000  
 i = -1.60820062e-003  
 j = 2.36014503e-004  
 CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

**ABCDM COEFFICIENTS**

a = 7.40772717e-006  
 b = 1.62923614e+000  
 c = -1.05785259e+001  
 d = -8.60807664e-005  
 m = 5.2  
 CPcor = -9.5700e-008 (nominal)


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.54801	0.00000	0.00000
-1.0000	34.7933	2.80290	4.86617	2.80288	-0.00001
1.0000	34.7936	2.97421	4.97286	2.97422	0.00001
15.0000	34.7943	4.26920	5.71473	4.26921	0.00001
18.5000	34.7942	4.61575	5.89731	4.61574	-0.00002
29.0000	34.7933	5.69898	6.43437	5.69898	0.00001
32.5000	34.7892	6.07180	6.60900	6.07180	-0.00000

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients



**CTD Conductivity (Secondary)**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1798  
CALIBRATION DATE: 21-Jun-13

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

**GHIJ COEFFICIENTS**

g = -3.92941949e+000  
h = 4.59841645e-001  
i = -7.88790971e-004  
j = 6.42017186e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

**ABCDM COEFFICIENTS**

a = 5.86987503e-007  
b = 4.56772457e-001  
c = -3.91757440e+000  
d = -7.11998198e-005  
m = 5.4  
CPcor = -9.5700e-008 (nominal)


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.92882	0.00000	0.00000
-1.0000	34.7942	2.80296	8.35585	2.80297	0.00001
1.0000	34.7951	2.97433	8.57635	2.97433	0.00001
15.0000	34.7956	4.26934	10.08504	4.26931	-0.00003
18.5000	34.7955	4.61591	10.45102	4.61590	-0.00001
29.0001	34.7944	5.69915	11.51780	5.69922	0.00008
32.5000	34.7889	6.07175	11.86157	6.07170	-0.00005

Conductivity =  $(g + hf^2 + if^3 + jf^4) / (10(1 + \delta t + \epsilon p))$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients



**CTD Dissolved Oxygen Sensor (primary)**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com


SENSOR SERIAL NUMBER: 0161  
CALIBRATION DATE: 12-Jun-13

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS		A = -2.3123e-003		NOMINAL DYNAMIC COEFFICIENTS	
Soc = 0.5019	B = 1.0028e-004	D1 = 1.92634e-4	H1 = -3.30000e-2		
Voffset = -0.5162	C = -2.1649e-006	D2 = -4.64803e-2	H2 = 5.00000e+3		
Tau20 = 1.26	E nominal = 0.036	H3 = 1.45000e+3			

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.25	2.00	0.00	0.775	1.25	0.00
1.26	12.00	0.00	0.856	1.26	0.00
1.27	6.00	0.00	0.810	1.27	0.00
1.30	20.00	0.00	0.931	1.29	-0.00
1.31	26.00	0.00	0.990	1.31	-0.00
1.32	30.00	0.00	1.031	1.32	0.00
3.97	2.00	0.00	1.337	3.97	-0.00
4.00	6.00	0.00	1.442	4.00	0.00
4.03	12.00	0.00	1.600	4.03	-0.00
4.05	20.00	0.00	1.814	4.05	0.00
4.05	26.00	0.00	1.983	4.05	0.00
4.07	30.00	0.00	2.111	4.07	-0.00
6.78	2.00	0.00	1.917	6.78	0.00
6.78	12.00	0.00	2.340	6.79	0.00
6.79	6.00	0.00	2.087	6.79	-0.00
6.82	20.00	0.00	2.703	6.82	-0.00
6.84	26.00	0.00	2.994	6.84	0.00
6.85	30.00	0.00	3.196	6.85	-0.00

Oxygen (ml/l) = Soc \* (V + Voffset) \* (1.0 + A \* T + B \* T<sup>2</sup> + C \* T<sup>3</sup>) \* OxSol(T,S) \* exp(E \* P / K)  
 V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU], K = temperature [Kelvin]  
 OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen



**CTD Dissolved Oxygen Sensor (secondary)**

**Sea-Bird Electronics, Inc.**

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com


SENSOR SERIAL NUMBER: 0080  
 CALIBRATION DATE: 13-Feb-13

SBE 43 OXYGEN CALIBRATION DATA

<b>COEFFICIENTS</b>	A = -3.0719e-003	<b>NOMINAL DYNAMIC COEFFICIENTS</b>
Soc = 0.4885	B = 1.5019e-004	D1 = 1.92634e-4 H1 = -3.30000e-2
Voffset = -0.5049	C = -2.7921e-006	D2 = -4.64803e-2 H2 = 5.00000e+3
Tau20 = 1.79	E nominal = 0.036	H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.18	2.00	0.07	0.756	1.18	-0.00
1.19	6.00	0.07	0.788	1.19	-0.00
1.20	12.00	0.06	0.839	1.20	0.00
1.23	20.00	0.06	0.909	1.23	0.00
1.27	26.00	0.06	0.977	1.27	0.01
1.28	30.00	0.06	1.018	1.28	0.01
4.01	6.00	0.07	1.461	4.01	-0.00
4.04	12.00	0.06	1.626	4.04	-0.00
4.08	20.00	0.06	1.849	4.08	0.00
4.10	2.00	0.07	1.376	4.09	-0.01
4.11	26.00	0.06	2.028	4.11	0.00
4.14	30.00	0.06	2.162	4.14	0.00
6.82	30.00	0.06	3.231	6.81	-0.00
6.95	26.00	0.06	3.084	6.95	0.00
7.02	20.00	0.06	2.817	7.01	-0.01
7.17	12.00	0.06	2.493	7.17	0.00
7.33	6.00	0.07	2.251	7.33	0.00
7.43	2.00	0.07	2.087	7.43	0.00

Oxygen (ml/l) = Soc \* (V + Voffset) \* (1.0 + A \* T + B \* T<sup>2</sup> + C \* T<sup>3</sup>) \* OxSol(T,S) \* exp(E \* P / K)  
 V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU], K = temperature [Kelvin]  
 OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen



**Fluorometer**

PO Box 518  
620 Applegate St.  
Philomath OR 97370



(541) 929-5650  
Fax (541) 929-5277  
<http://www.wetlabs.com>

**Chlorophyll Fluorometer Characterization in Uranine liquid Proxy (new method)**

**Date:** 07/17/13  
**Serial #:** AFLD-011  
**Tech:** K.C.

**Dark Counts** 0.117volts  
**CEV** .682 volts  
**SF** 32.743

**FSV** 4.61 volts

**Linearity:** 0.999 R<sup>2</sup> (0–1.5 volts)  
0.995 R<sup>2</sup> (0– 5.45 volts)

Notes:

**Dark Counts:** Signal output of the meter in clean water with black tape over detector.

**CEV** is the chlorophyll equivalent voltage. This value is the signal output of the fluorometer when using a Uranine dye fluorescent proxy that has been determined to be approximately equivalent to **26.4 µg/l** of a *Thalassiosira weissflogii* phytoplankton culture.

**SF** is the scale factor used to derive chlorophyll concentration from the signal voltage output of the fluorometer. The scale factor is determined by using the following equation:  
 $SF = (18.3) / (CEV - \text{dark})$ .

**FSV** is the maximum signal voltage output that the fluorometer is capable of.

Chlorophyll concentration expressed in µg/l (mg/m<sup>3</sup>) can be derived by using the following equation: (µg/l)  
= (V<sub>measured</sub> - dark) \* SF

The relationship between fluorescence and chlorophyll-*a* concentrations in-situ is highly variable. The scale factor listed on this document was determined by using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer you must perform secondary measurements on the populations of interest. This is typically done using extraction based measurement techniques on discrete samples. For additional information on determination of chlorophyll concentration see [ Standard Methods For The Examination Of Water And Wastewater] part 10200 H published jointly by: American Public Health Association, American Water Works Association and Water Environment Federation.

PO Box 518  
620 Applegate St.  
Philomath OR 97370



(541) 929-5650  
Fax (541) 929-5277  
<http://www.wetlabs.com>

### Chlorophyll Fluorometer Characterization in Reflective Solid Proxy (old method)

**Date:** 07/17/13  
**Serial #:** AFLD-011  
**Tech:** K.C.

**Dark Counts** 0.117 volts  
**CEV** 1.594 volts  
**SF** 14.962

**FSV** 4.61 volts

**Linearity:** 0.999 R<sup>2</sup> (0–1.5 volts)  
0.995 R<sup>2</sup> (0– 5.45 volts)

Notes:

**Dark Counts:** Signal output of the meter in clean water with black tape over detector.

**CEV** is the chlorophyll equivalent voltage. This value is the signal output of the fluorometer when using a Uranine dye fluorescent proxy that has been determined to be approximately equivalent to **21.6 µg/l** of a *Thalassiosira weissflogii* phytoplankton culture.

**SF** is the scale factor used to derive chlorophyll concentration from the signal voltage output of the fluorometer. The scale factor is determined by using the following equation:  
 $SF = (21.6) / (CEV - \text{dark})$ .

**FSV** is the maximum signal voltage output that the fluorometer is capable of.

Chlorophyll concentration expressed in µg/l (mg/m<sup>3</sup>) can be derived by using the following equation: (µg/l) = (V<sub>measured</sub> - dark) \* SF

The relationship between fluorescence and chlorophyll-*a* concentrations in-situ is highly variable. The scale factor listed on this document was determined by using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer you must perform secondary measurements on the populations of interest. This is typically done using extraction based measurement techniques on discrete samples. For additional information on determination of chlorophyll concentration see [ Standard Methods For The Examination Of Water And Wastewater] part 10200 H published jointly by: American Public Health Association, American Water Works Association and Water Environment Federation.

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## C-Star Calibration

---

Date **September 5, 2013** S/N# **CST-889DR** Pathlength **25cm**

---

### Analog output

**V<sub>d</sub>** **0.060 V**  
**V<sub>air</sub>** **4.726 V**  
**V<sub>ref</sub>** **4.624 V**

Temperature of calibration water **23.1 °C**  
Ambient temperature during calibration **21.2 °C**

---

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x, in meters): **Tr = e<sup>-cx</sup>**

To determine beam transmittance: **Tr = (V<sub>sig</sub> - V<sub>dark</sub>) / (V<sub>ref</sub> - V<sub>dark</sub>)**

To determine beam attenuation coefficient: **c = -1/x \* ln (Tr)**

**V<sub>d</sub>** Meter output with the beam blocked. This is the offset.

**V<sub>air</sub>** Meter output in air with a clear beam path.

**V<sub>ref</sub>** Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V<sub>ref</sub>.

Ambient temperature: meter temperature in air during the calibration.

**V<sub>sig</sub>** Measured signal output of meter.

Revision M

7/26/11

## Customer Alert: July, 2011

### CHLa Scale Factors Shift

WET Labs calibration testing has revealed that our CHLa solid proxy used to calibrate our ECO and Wetstar fluorometers allows a large amount of instrument to instrument variability. Also, we have differences in scaling between Wetstar CHLa fluorometers and ECO CHLa Fluorometers because of differences in the solid proxy used to characterize these instruments. A new methodology using a liquid proxy has been implemented to assure stable calibrations between instruments and to match up the ECO FL and Wetstar FL corrected data outputs.

#### Instruments affected:

All CHLa ECO fluorometers built or calibrated before January 2011.

All CHLa Wetstar fluorometers built or calibrated before July 2011.

#### WET Labs' Actions:

##### New Instruments:

WET Labs has instituted a new calibration standard solution preparation methodology. All new ECO/Wetstar CHLa fluorometers delivered from this date forward will have range characteristics as per current specifications and scale factors.

##### Instruments returned for service and calibration:

Instruments returned for service and calibration will be calibrated using the new methodology. We are tuning all service instruments to this new liquid proxy to decrease instrument to instrument variability.

In some cases, we will not be able to achieve the previously stated range of an instrument. In these cases, we will strive for the highest resolution with the highest signal to noise ratio possible.

WET Labs service technicians will incorporate these improvements during service when practical. WET Labs' term for this service is 'retuning.' Accordingly, a serviced instrument may well have a better performance after retuning than when it was first built.

For instruments that are retuned, benefiting in either resolution or signal to noise ratio, WET Labs can provide pre calibration data to allow you to link your data sets prior to service with your data sets after the instrument is returned to you.

##### Recommended Customer Actions:

If you calibrate your instruments then you do not need to take any action. Continue to use your calibration.

If you report scaled or raw data, you should adjust your reported values.

For instruments returned for service, you will use the ratio between the previous scale factor and pre-service scale factor. This ratio will cover both the change in the methodology and any change in your instrument between the previous calibration and this servicing.

Use the post-service scale factor going forward.