

OOI Southern Ocean

Data Report NBP17-09

RVIB Nathaniel B. Palmer

11/23/2017 – 12/09/2017

Data Report Prepared By:
Chris Linden

Table of Contents

| | |
|---|-----------|
| Introduction | 1 |
| Distribution Contents at a Glance..... | 2 |
| Extracting Data..... | 3 |
| Distribution Contents..... | 4 |
| Cruise Track..... | 4 |
| Satellite Images..... | 4 |
| NBP Data Products..... | 4 |
| JGOFS | 5 |
| MGD77 | 6 |
| Science of Opportunity..... | 8 |
| ADCP | 8 |
| pCO ₂ | 8 |
| Cruise Science | 9 |
| CTD | 9 |
| XBT | 9 |
| RVDAS | 9 |
| Sensors and Instruments | 9 |
| Underway Sensors | 10 |
| <i>Meteorology and Radiometry</i> | <i>10</i> |
| <i>Geophysics</i> | <i>10</i> |
| <i>Oceanography.....</i> | <i>10</i> |
| <i>Navigational Instruments</i> | <i>10</i> |
| Data | 11 |
| Underway Data | 12 |
| <i>Meteorology (mwx1).....</i> | <i>12</i> |
| <i>Knudsen (knud)</i> | <i>13</i> |
| <i>Gravimeter (grv1).....</i> | <i>13</i> |
| <i>pCO₂ (pco2)</i> | <i>14</i> |
| <i>Micro TSG (tsg1, tsg2).....</i> | <i>14</i> |
| <i>Remote Temperature (rtmp).....</i> | <i>14</i> |
| <i>GUV (pguv).....</i> | <i>15</i> |
| <i>Engineering (eng1).....</i> | <i>15</i> |
| <i>Hydro DAS (hdas)</i> | <i>16</i> |
| <i>Winch (bwnc, cwnc, twnc)</i> | <i>16</i> |
| <i>Multibeam (mbdp).....</i> | <i>16</i> |
| <i>Fluorometer (flr1).....</i> | <i>17</i> |
| Navigational Data..... | 18 |
| <i>GPS (s330, seap, PCOD).....</i> | <i>18</i> |
| <i>Gyro Compass (gyr1).....</i> | <i>23</i> |
| <i>ADCP Course (adcp).....</i> | <i>23</i> |

| | |
|--|-----------|
| Processed Data | 24 |
| <i>pCO₂ – Merged</i> | 24 |
| Calculations | 25 |
| <i>PAR</i> | 25 |
| <i>PSP</i> | 25 |
| <i>PIR</i> | 26 |
| Acquisition Problems and Events | 27 |
| Appendix A: Sensors | 28 |
| NBP17-09 Sensors | 28 |
| Appendix B: Calibration Sheets | 29 |
| Gravity | 29 |
| Meteorology | 30 |
| <i>Anemometers</i> | 30 |
| <i>Barometer 1</i> | 31 |
| <i>Barometer 2</i> | 32 |
| <i>Temperature / Humidity</i> | 33 |
| <i>PIR</i> | 35 |
| <i>PSP</i> | 36 |
| <i>PAR (Mast)</i> | 37 |
| <i>GUV (Mast)</i> | 38 |
| Underway Seawater Sampling System | 40 |
| <i>Micro-TSG 1</i> | 40 |
| <i>Micro-TSG2</i> | 42 |
| <i>Digital Remote Temp</i> | 44 |
| <i>Transmissometer</i> | 45 |
| <i>Fluorometer</i> | 46 |
| CTD | 47 |
| <i>Conductivity (primary)</i> | 47 |
| <i>Conductivity (secondary)</i> | 48 |
| <i>Fluorometer</i> | 49 |
| <i>Pressure Sensor</i> | 50 |
| <i>Dissolved Oxygen (primary)</i> | 51 |
| <i>Dissolved Oxygen (secondary)</i> | 52 |
| <i>Pump (primary)</i> | 53 |
| <i>Pump (secondary)</i> | 54 |
| <i>Temperature (primary)</i> | 55 |
| <i>Temperature (Secondary)</i> | 56 |
| <i>Transmissometer</i> | 57 |

Introduction

The NBP data acquisition systems continuously log data from the instruments used during the cruise. This document describes:

- The structure and organization of the data on the distribution media
- The format and contents of the data strings
- Formulas for calculating values
- Information about the specific instruments in use during the cruise
- A log of acquisition problems and events during the cruise that may affect the data
- Scanned calibration sheets for the instruments in use during the cruise.

The data is distributed on a DVD+R or CD-ROM written in ISO9660 level-1 format. It is readable by virtually every computing platform.

All the data has been archived using 'tar' and compressed using 'gzip', identified by the '.tz' extension. Tools are available on all platforms for uncompressing and de-archiving these formats: On Macintosh use the built-in Archive Utility, or tar in the terminal. On Windows operating systems use WinZip or 7Zip.

MultiBeam and Bathymetry data, if collected, are distributed separately.

IMPORTANT: Read the last section, "Acquisition Problems and Events," for important information that may affect the processing of this data.

Distribution Contents at a Glance

Volume 1 of 1: NBP17-09

File

```

/
  NBP17-09.gmt
  NBP17-09.mgd
  NBP17-09.trk
  NBP1709DATA.docx
  NBP1709DATA.pdf
  INSTCOEF.TXT
  NBP1709map.pdf
  NBP1709-study-area.pdf
/process
  NBP1709JGOF.tz
  NBP1709MGD.tz
  NBP1709PCO2.tz
  NBP1709PROC.tz
  NBP1709QC.tz
/rvdas/nav
  NBP1709adcp.tz
  NBP1709gyrl.tz
  NBP1709s330.tz
  NBP1709seap.tz
/rvdas/uw
  NBP1709bwnc.tz
  NBP1709ctdd.tz
  NBP1709engl.tz
  NBP1709grv1.tz
  NBP1709hdas.tz
  NBP1709knud.tz
  NBP1709mbdp.tz
  NBP1709mwx1.tz
  NBP1709pco2.tz
  NBP1709pguv.tz
  NBP1709rtmp.tz
  NBP1709tsg1.tz
  NBP1709tsg2.tz
  NBP1709twnc.tz
/ocean
  NBP1709ctd.tz

```

Description

```

Root level directory
  GMT binary file of MGD77 data
  Full Cruise MGD77 data file
  Text file of cruise track
  Data Report NBP17-09 (MS Word)
  Data Report NBP17-09 (PDF format)
  Instrument Coefficient File
  Cruise Track Map
  Cruise Track of the Study Area

Processed data
  JGOFs format data files
  MGD Data
  Merged pCO2 data files
  Other processed data
  Daily RVDAS QC postscript plots

Navigation data
  ADCP Data Sets
  Gyro raw data
  Seapath 330 data
  Seapath 200 data

Underway data
  Baltic winch data
  CTD depth data
  Engineering data
  Gravimeter data
  HydroDAS raw data
  Knudsen raw data
  Multi-beam depth
  Meteorology raw data
  pCO2 raw data
  GUV raw data
  Remote Temperature data
  Micro TSG1 data
  Micro TSG2 data
  Trawl winch data

Ocean data
  CTD Data

```

Extracting Data

The data files will have a “.tz” extension on the filename. The “.tz” extension is for files whose contents have been archived using the “tar” utility and compressed with the “gzip” utility.

An example of creating a compressed archive file:

```
tar -czvf archive_filename files_to_archive
```

An example of listing the files in an archive:

```
tar -tzvf archive_filename
```

An example redirecting the list output to a file, where `contents.list` is the name of the file to create:

```
tar -tzvf archive_filename > contents.list
```

An example extracting all files from the archive:

```
tar -xzf archive_filename
```

An example extracting specific files from the archive:

```
tar -xzf archive_filename list_of_files_to_extract
```

Distribution Contents

Cruise Track

The distribution DVD includes a GMT cruise track file (NBP17-09.trk). It contains the longitude and latitude at one-minute intervals extracted from the NBP17-09.gmt file.

Satellite Images

Satellite Images processed for this cruise can be found in the directory, /Imagery in two subdirectories, ice and wx (weather). Files are named using the convention, IdDDDYA.jpg where:

Id = image type (ice = ice, wx = weather)

DDD = year-day

YY = year

A = allows for multiple images of one type for one day

NBP Data Products

Two datasets are created on each cruise: JGOFS and MGD77.

The data processing scripts used to produce JGOFS and MGD77 data sets create a lot of intermediate files. These files are included on the data distribution media in a file called NBP1709proc.TAR. They are included to make re-processing easier in the event of an error, but no extensive detail of the formats is included in this document. If you have any questions, please contact itvessel@usap.gov.

JGOFS

The JGOFS data set can be found on the distribution media in the file /process/NBP17-09JGOF.tar. The archive contains one file produced for each day named jgDDD.dat.gz, where DDD is the year-day the data was acquired. The “.gz” extension indicates that the individual files are compressed before archiving. Each daily file consists of 22 columnar fields in text format as described in the table below. The JGOFS data set is created from calibrated data decimated at one-minute intervals. Several fields are derived measurements from more than a single raw input. For example, Course Made Good (CMG) and Speed Over Ground (SOG) are calculated from gyro and GPS inputs. Daily plots during the cruise are produced from the JGOFS data set. Note: Null, unused, or unknown fields are indicated as “NAN” as 9999 in the JGOFS data.

| Field | Data | Units |
|-------|---|--------------------------------|
| 01 | UTC Date | dd/mm/yy |
| 02 | UTC Time | hh:mm:ss |
| 03 | Seapath Latitude (negative is South) | tt.tttt |
| 04 | Seapath Longitude (negative is West) | ggg.gggg |
| 05 | Speed Over Ground | knots |
| 06 | GPS HDOP | - |
| 07 | Gyro Heading | Degrees (azimuth) |
| 08 | Course Made Good | Degrees (azimuth) |
| 09 | Mast PAR | μEinstein's/meter ² |
| 10 | Sea Surface Temperature | °C |
| 11 | Sea Surface Conductivity | siemens/meter |
| 12 | Sea Surface Salinity | PSU |
| 13 | Sea Depth (uncorrected, calc. sw sound vel. 1500 m/s) | meters |
| 14 | True Wind Speed (max speed windbird) | meters/sec |
| 15 | True Wind Direction (max speed windbird) | degrees (azimuth) |
| 16 | Ambient Air Temperature | °C |
| 17 | Relative Humidity | % |
| 18 | Barometric Pressure | mBars |
| 19 | Sea Surface Fluorometry | volts (0-5 FSO) |
| 20 | Transmissometry | % |
| 21 | PSP | W/m ² |
| 22 | PIR | W/m ² |

MGD77

The MGD77 data set is contained in a single file for the entire cruise. It can be found in the top level of the distribution data structure as NBP17-09.mgd. The file NBP17-09.gmt is created from the MGD77 dataset using the “mgd77togmt” utility. NBP17-09.gmt can be used with the GMT plotting package.

The data used to produce the NBP17-09.mgd file can be found on the distribution media in the file /process/NBP1709proc.tar. The data files in the archive contain a day’s data and follow the naming convention Dddd.fnl.tz, where ddd is the year-day. These files follow a space-delimited columnar format that may be more accessible for some purposes. They contain data at one-second intervals rather than one minute and are individually “gzipped” to save space. Below is a detailed description of the MGD77 data set format. The other files in the archive contain interim processing files and are included to simplify possible reprocessing of the data using the RVDAS NBP processing scripts.

All decimal points are implied. Leading zeros and blanks are equivalent. Unknown or unused fields are filled with 9’s. All “corrections”, such as time zone, diurnal magnetics, and EOTVOS, are understood to be added.

| Col | Len | Type | Contents | Description, Possible Values, Notes |
|---------|-----|------|---|--|
| 1 | 1 | int | Data record type | Set to "5" for data record |
| 2-9 | 8 | char | Survey identifier | |
| 10-12 | 3 | int | Time zone correction | corrects time (in chars 13-27) to UTC when added; 0=UTC |
| 13-16 | 4 | int | Year | 4 digit year |
| 17-18 | 2 | int | Month | 2 digit month |
| 19-20 | 2 | int | Day | 2 digit day |
| 21-22 | 2 | int | Hour | 2 digit hour |
| 23-27 | 5 | real | Minutes x 1000 | |
| 28-35 | 8 | real | Latitude x 100000 | Positive = North, Negative = South. (-9000000 to 9000000) |
| 36-44 | 9 | real | Longitude x 100000 | Positive = East, Negative = West. (-18000000 to 18000000) |
| 45 | 1 | int | Position type code | 1 = Observed fix, 3 = Interpolated, 9 = Unspecified |
| 46-51 | 6 | real | Bathymetry, 2-way travel time | In 10,000th of seconds. Corrected for transducer depth and other such corrections. |
| 52-57 | 6 | real | Bathymetric, corrected depth | In tenths of meters |
| 58-59 | 2 | int | Bathymetric correction code | This code details the procedure used for determining the sound velocity correction to depth |
| 60 | 1 | int | Bathymetric type code | 1 = Observed, 3 = Interpolated (Header Seq. 12), 9 = Unspecified |
| 61-66 | 6 | real | Magnetics total field, 1 st sensor | In tenths of nanoteslas (gammas) |
| 67-72 | 6 | real | Magnetics total field, 2 nd sensor | In tenths of nanoteslas (gammas), for trailing sensor |
| 73-78 | 6 | real | Magnetics residual field | In tenths of nanoteslas (gammas). The reference field used is in Header Seq. 13 |
| 79 | 1 | int | Sensor for residual field | 1 = 1 st or leading sensor, 2 = 2 nd or trailing sensor, 9 = Unspecified |
| 80-84 | 5 | real | Magnetics diurnal correction | In tenths of nanoteslas (gammas). (In nanoteslas) if 9-filled (i.e., set to "+9999"), total and residual fields are assumed to be uncorrected; if used, total and residual are assumed to have been already corrected. |
| 85-90 | 6 | F6.0 | Depth or altitude of magnetics sensor | (In meters). Positive = Below sea level, 3 = Above sea level |
| 91-97 | 7 | real | Observed gravity | In 10 th of mgals. Corrected for Eotvos, drift, tares |
| 98-103 | 6 | real | EOTVOS correction | In 10 th of mgals. $E = 7.5 V \cos \phi \sin \alpha + 0.0042 V^*V$ |
| 104-108 | 5 | real | Free-air anomaly | In 10 th of mgals, G = observed, G = theoretical |
| 109-113 | 5 | char | Seismic line number | Cross reference for seismic data |
| 114-119 | 6 | char | Seismic shot-point number | |
| 120 | 1 | int | Quality code for navigation | 5 = Suspected, by the originating institution 6 = Suspected, by the data center 9 = No identifiable problem found |

Science of Opportunity

ADCP

The shipboard ADCP system measures currents in a depth range from about 30 to 300 m -- in good weather. In bad weather or in ice, the range is reduced, and sometimes no valid measurements are made. ADCP data collection is the OPP-funded project of Eric Firing (University of Hawaii) and Teri Chereskin (Scripps Institution of Oceanography). Data is collected on both the LMG and the NBP for the benefit of scientists on individual cruises, and for the long-term goal of building a profile of current structure in the Southern Ocean.

A data feed is sent from the ADCP system to RVDAS whenever a reference layer is acquired. This feed contains east and north vectors for ship's speed, relative to the reference layer, and ship's heading. Collected files (one per day) are archived in NBP1709adcp.tar in the directory /rvdas/nav.

pCO₂

The NBP carries a pCO₂ measurement system from Lamont-Doherty Earth Observatory (LDEO). pCO₂ data is recorded by RVDAS and transmitted to LDEO at the end of each cruise. You will find pCO₂ data in a file named NBP1709pco2.tar in the /process directory, which contains the pCO₂ instrument's data merged with GPS, meteorological and other oceanographic measurements. For more information contact Colm Sweeney (csweeney@ldeo.columbia.edu).

Cruise Science

CTD

The CTD data has been placed in the tar file /ocean/NBP17-09ctd.tar. The archive contains tar files NBP17-09proc.tar.

XBT

During a cruise, eXpendable BathyThermographs (XBTs) may have been used to obtain water column temperature profiles, providing corrections to the sound velocity profile for the multibeam system. The data files from those launches would be included as NBP1709xbt.tar in the /ocean directory. **No XBTs were collected on this cruise.**

RVDAS

The Research Vessel Data Acquisition System (RVDAS) was developed at Lamont-Doherty Earth Observatory of Columbia University and has been in use on its research ship for many years. It has been extensively adapted for use on the USAP research vessels.

Daily data processing of the RVDAS data is performed to calibrate and convert values into useable units and as a quality-control on operation of the DAS. Raw and processed data sets from RVDAS are included in the data distribution. The tables below provide detailed information on the sensors and data. Be sure to read the “Significant Acquisition Events” section for important information about data acquisition during this cruise.

Sensors and Instruments

RVDAS data is divided into two general categories, *underway and navigation*. They can be found on the distribution media as subdirectories under the top level rvdas directory: /rvdas/uw, and /rvdas/nav. Processed oceanographic data is in the top level directory, /process. Each instrument or sensor produces a data file named with its channel ID. Each data file is g-zipped to save space on the distribution media. Not all data types are collected every day or on every cruise.

The naming convention for data files produced by the sensors and instruments is

NBP[CruiseID][ChannelID].dDDD

Example: NBP17-09mwx1.d025

- The CruiseID is the numeric name of the cruise, in this case, NBP17-09.
- The ChannelID is a 4-character code representing the system being logged. An example is “mwx1,” the designation for meteorology.
- DDD is the day of year the data was collected

Underway Sensors

Meteorology and Radiometry

| Measurement | String ID | Collection Status | Rate | Instrument |
|------------------------|----------------|-------------------|-------|-----------------------------------|
| Air Temperature | mwx1 (met) | Continuous | 1/sec | RM Young 41372LC |
| Relative Humidity | mwx1 (met) | Continuous | 1/sec | RM Young 41372LC |
| Wind Speed / Direction | mwx1 (pus,sus) | Continuous | 1/sec | Gill Instruments 1390-PK-062 |
| Barometer | mwx1 (met) | Continuous | 1/sec | RM Young 61201 |
| PAR | mwx1 (met) | Continuous | 1/sec | Biospherical Instruments QSR-240 |
| PIR | mwx1 (met) | Continuous | 1/sec | Eppley PIR |
| PSP | mwx1 (met) | Continuous | 1/sec | Eppley PSP |
| GUV | pguv | Continuous | 2/sec | Biospherical Instruments GUV-2511 |

Geophysics

| Measurement | String ID | Collection Status | Rate | Instrument |
|-------------|-----------|-------------------|--------|-----------------|
| Gravimeter | grv1 | Continuous | 1/sec | BGM3/210 |
| Bathymetry | knud | Continuous | varies | Knudsen Chirp |
| Bathymetry | mbdp | Continuous | varies | Kongsberg EM122 |

Oceanography

| Measurement | String ID | Collection Status | Rate | Instrument |
|---------------------------|-----------|-------------------|-----------|----------------------|
| Conductivity | tsg1,tsg2 | Continuous | 0.5/sec | Sea-Bird SBE 45 |
| Ocean Surface Temperature | rtmp | Continuous | 1.2/sec | Sea-Bird SBE 38 |
| Transmissometer | hdas | Continuous | 0.5/sec | WetLabs C-Star |
| Fluorometer | hdas | Continuous | 0.5/sec | WetLabs AFLT |
| pCO ₂ | pco2 | Continuous | 0.017/sec | LDEO instrumentation |
| ADCP | adcp | Continuous | 1/sec | UHDAS |
| Bathymetry | sim1 | Continuous | varies | Simrad EK60 Sonar |

Navigational Instruments

| Measurement | String ID | Collection Status | Rate | Instrument |
|--|-----------|-------------------|---------|------------------|
| Heading, Speed, Course, GPS, Heave, Roll and Pitch | s330 | Continuous | 1/sec | Seapath 330 GPS |
| Heading, Speed, Course, GPS, Heave, Roll and Pitch | seap | Continuous | 1/sec | Seapath 200 GPS |
| Heading, Speed, Course, and GPS | PCOD | Continuous | 1/sec | Furuno GP-330B |
| Heading | gyr1 | Continuous | 0.2/sec | Yokogawa Compass |

Data

Data is received from the RVDAS system via RS-232 serial connections. A time tag is added at the beginning of each line of data in the form,

```
yy+dd:hh:mm:ss.sss [data stream from instrument]
```

where

| | |
|--------|---------------------------|
| yy | = two-digit year |
| ddd | = day of year |
| hh | = 2 digit hour of the day |
| mm | = 2 digit minute |
| ss.sss | = seconds |

All times are reported in UTC.

The delimiters that separate fields in the raw data files are often spaces and commas but can be other characters such as : = @. Occasionally no delimiter is present. Care should be taken when reprocessing the data that the field's separations are clearly understood.

In the sections below a sample data string is shown, followed by a table that lists the data contained in the string.

Each section on the next page describes a type of data file (file name extension in parentheses) followed by a typical line of data in the file. In the table(s) for each section is a description of the fields within each line of data. Note: most data files listed below will be included with each cruise's data distribution; however some types of files may be omitted if the instrument was not operating during the cruise. The available data files can be found in the /rvdas/uw and /rvdas/nav directories on the distribution disc.

Underway Data /rvdas/uw**Meteorology (mwx1)****MET**

15+055:20:27:24.636 MET,12.1,-39,-6.07,77.4,178.0729,0.809536,-0.1235019,268.1754,267.9648,970.7878

| Field | Data | Format | Unit |
|-------|---|---------------------|------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | MET Flag | | |
| 3 | Power Supply Voltage | vv.v | V |
| 4 | Enclosure Relative Humidity (not implemented) | xx.x | % |
| 5 | Air Temperature, Celsius | xx.x | C |
| 6 | Air Relative Humidity | xx.x | % |
| 7 | PAR (Photosynthetically Available Radiation) | xxx.xxxx | mV |
| 8 | PSP (Shortwave Radiation) | x.xxxxxx | mV |
| 9 | PIR Thermopile (Longwave Radiation) | x.xxxxxx | mV |
| 10 | PIR Case Temperature | xxx.xxxx | K |
| 11 | PIR Dome Temperature | xxx.xxxx | K |
| 12 | Barometer | xxx.xxxx | mBar |

PUS

15+055:21:47:42.452 PUS,A,037,014.36,M,+325.38,-010.29,60,0F

| Field | Data | Format | Unit |
|-------|------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | PUS Flag | | |
| 3 | A | x | A |
| 4 | Port Wind Relative Direction | xxx | degrees |
| 5 | Port Wind Relative Speed | xxx.xx | m/s |
| 6 | M = Meters (for previous) | x | M |
| 7 | Sound Speed | xxx.xx | m/s |
| 8 | Sonic Temperature | xxx.xx | C |
| 9 | Unit Status* | xx | numeric |
| 10 | Checksum | xx | alphanumeric |

Status

00 = Good, 60 = Good. Any other value indicates fault

SUS

15+055:21:50:48.409 SUS,A,338,012.63,M,+326.15,-009.05,60,0F

| Field | Data | Format | Unit |
|-------|-----------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | SUS Flag | | |
| 3 | A | x | A |
| 4 | Starboard Wind Relative Direction | xxx | degrees |
| 5 | Starboard Wind Relative Speed | xxx.xx | m/s |
| 6 | M = Meters (for previous) | x | M |
| 7 | Sound Speed | xxx.xx | m/s |
| 8 | Sonic Temperature | xxx.xx | C |
| 9 | Unit Status* | xx | numeric |
| 10 | Checksum | xx | alphanumeric |

Status

00 = Good, 60 = Good. Any other value indicates fault

Knudsen (knud)

| Field | Data | Format | Unit |
|-------|---------------------------------|---------------------|---------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | 3.5kHz = Low frequency in use | x.xxxx | 3.5kHz |
| 3 | Low Frequency Depth | xxxx.xx | m |
| 4 | Valid Flag | x | 0 |
| 5 | 12.0kHz = High frequency in use | xx.xxxx | 12.0kHz |
| 6 | High Frequency Depth | xxxx.xx | m |
| 7 | Valid Flag | x | 0 |
| 8 | Sound Speed Velocity | xxxx | m/s |
| 9 | Latitude | xx.xxxxxx | degrees |
| 10 | Longitude | xx.xxxxxx | degrees |

Gravimeter (grv1)

15+056:14:21:21.153 01:025268 00

| Field | Data | Format | Unit |
|-------|----------------|---------------------|------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | 01: | xx: | 01 |
| 3 | Gravity Count* | xxxxxx | Flit Count |
| 4 | Error Flag | xx | numeric |

Error Flag

00 = All well, 01 = CPS malfunction, 02 = Sensor Malfunction, 03 = CPS and sensor Malfunction

A gravity tie is taken at the start of the cruise and applied throughout the cruise. There is no accounting for drift after the pre-cruise gravity time. The post cruise gravity tie is available by requesting it from ethq@usap.gov.

pCO₂ (pco2)

15+056:14:41:10.392 2015056.60236 2608.36 30.14 977.91 48.25 368.76 353.92 -1.18 -1.26 0.00 Equil

| Field | Data | Format | Unit |
|-------|---------------------------------|---------------------|----------------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | pCO ₂ time tag* | yyyyddd.ttt | UTC |
| 3 | Raw Voltage (IR) | xxxx.xx | mV |
| 4 | Cell Temperature | xx.xx | C |
| 5 | Equilibration Pressure | xxx.xx | mBar |
| 6 | Flowrate | xxx.xx | cm ³ /min |
| 7 | pCO ₂ Pressure | xxx.xx | μAtm |
| 8 | VCO ₂ Concentration | xx.xx | ppm |
| 9 | Equilibrator Temperature, RTD | xx.xx | C |
| 10 | Equilibrator Temperature, SBE38 | xx.xx | C |
| 11 | Valve Position | xx | numeric |
| 12 | Flow Source* | | text |

pCO₂ time tag

ttt = fractional time of day

Flow SourceEquil = pCO₂ Measurement**Micro TSG (tsg1, tsg2)**

15+056:15:06:06.644 -1.1809, 2.73404, 34.0574, 1442.367

| Field | Data | Format | Unit |
|-------|----------------|---------------------|------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | Temperature | xx.xxxx | C |
| 3 | Conductivity | xx.xxxx | s/m |
| 4 | Salinity | xx.xxxx | PSU |
| 5 | Sound Velocity | xxxx.xxx | m/s |

Remote Temperature (rtmp)

15+056:15:10:38.244 -1.4644

| Field | Data | Format | Unit |
|-------|------------------------------|---------------------|------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | Temperature, Seawater Intake | xx.xxxx | C |

GUV (pguv)

15+057:14:51:33.808 022615 065133 .000132 .010878 .047479 .004407 -.002799 .014652 .027558 .094395
 .417814 -4.466095

| Field | Data | Format | Unit |
|-------|---|---------------------|-------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | Date | mmddyy | UTC-4 |
| 3 | Time | hhmmss | UTC-4 |
| 4 | Ed0GND (sensor ground voltage) | xxxxxx | V |
| 5 | Ed0320 (downwelling 320nm irradiance) | xxxxxx | μW |
| 6 | Ed0340 (downwelling 340nm irradiance) | xxxxxx | μW |
| 7 | Ed0313 (downwelling 313nm irradiance) | xxxxxx | μW |
| 8 | Ed0305 (downwelling 305nm irradiance) | xxxxxx | μW |
| 9 | Ed0380 (downwelling 380nm irradiance) | xxxxxx | μW |
| 10 | Ed0PAR (downwelling 400-700nm irradiance) | xxxxxx | μE |
| 11 | Ed0395 (downwelling 395nm irradiance) | xxxxxx | μW |
| 12 | Ed0Temp (sensor array temperature) | xxxxxx | C |
| 13 | Ed0Vin (input voltage) | x.xxxxxx | V |

Engineering (eng1)

15+057:16:41:24.536 12.25 23.21 507.8 0.6 162.6 -751.9 0 0 NAN NAN -10.3 7.2

| Field | Data | Format | Unit |
|-------|--|---------------------|---------------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | Supply Voltage | xx.xx | V |
| 3 | Case Temperature | xx.xx | C |
| 4 | Seawater Flow, Aquarium Room | xxx.x | l / min |
| 5 | Seawater Flow, Helo-deck | x.x | l / min |
| 6 | Seawater Flow, Hydro-lab | xxx.x | l / min |
| 7 | Seismic Air Pressure | xxx.x | lbf/in ² |
| 8 | Not Currently Hooked Up | x | 0 or NAN |
| 9 | Not Currently Hooked Up | x | 0 or NAN |
| 10 | Not Currently Hooked Up | x | 0 or NAN |
| 11 | Not Currently Hooked Up | x | 0 or NAN |
| 12 | Altimeter for Yo-Yo Camera - Rarely used* | xx.xx | m |
| 13 | Transmissometer for Yo-Yo camera - Rarely used* | xxx.x | % |

Altimeter

This is rarely used, and only provides real data when connected. When not connected, provides a value approx = -10.

Transmissometer

This is rarely used, and only provides real data when connected. When not connected, provides a value range of approx = 0 to 10.

Hydro DAS (hdas)

15+057:16:07:09.456 12.15038 12.39402 336.5517 4431.724 -1 20.5 64 33.5 43.5

| Field | Data | Format | Unit |
|-------|------------------------|---------------------|---------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | Supply Voltage | xx.xxxxx | V |
| 3 | Case Temperature | xx.xxxxx | C |
| 4 | Fluorometer | xxx.xxxx | mV |
| 5 | Transmissometer | xxxx.xxx | mV |
| 6 | Sea Water Valve* | x | -1 or 0 |
| 7 | Flow Meter 1 Frequency | xx.x | Hz |
| 8 | Flow Meter 2 Frequency | xx.x | Hz |
| 9 | Flow Meter 3 Frequency | xx.x | Hz |
| 10 | Flow Meter 4 Frequency | xx.x | Hz |

Sea Water Valve

-1 = Stern Thruster Valve, 0 = Moon Pool Valve

Winch (bwnc, cwnc, twnc)

15+057:14:12:24.405 02RD,2015-02-26T14:55:32.051,STBD TRAWL,00000064,-00000.0,-00023.2,3594

| Field | Data | Format | Unit |
|-------|-----------------------|-------------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | LAN ID | | alphanumeric |
| 3 | LCI-90i Date and Time | yyyy-mm-ddThh:mm:ss.sss | |
| 4 | Winch Name | | alphabetical |
| 5 | Tension | xxxxxxxxx | lbs |
| 6 | Speed | xxxxx.x | m/min |
| 7 | Payout | xxxxx.x | m |
| 8 | Checksum | x.xxxx | numeric |

Multibeam (mbdp)

15+058:22:04:52.826 \$KIDPT,594.68,7.67,12000.0*43

| Field | Data | Format | Unit |
|-------|---------------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | KIDPT | x.x | m |
| 3 | Depth at Transducer | x.x | m |
| 4 | Distance to Waterline from Transducer | x.x | m |
| 5 | Maximum Range in Use | x.x | alphanumeric |
| 6 | Checksum | xx | UTC |

Fluorometer (flr1)

17+241:00:00:22.630 99/99/99 99:99:99 695 83 538

| Field | Data | Format | Unit |
|-------|--------------------------------|---------------------|------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | Ignore | x.x | n/a |
| 3 | Ignore | x.x | n/a |
| 4 | Wavelength (Not used) | xxx | nanometers |
| 5 | Chlorophyll Counts | xxxx | Counts |
| 6 | Internal Thermistor (Not used) | xxx | Therm |

Navigational Data /rvdas/nav**GPS (s330, seap, PCOD)**

1. ***Seapath 330***
 - a. NMEA 0183 strings
 - i. GPZDA
 - ii. GPGGA
 - iii. GPVTG
 - iv. GPHDT
 - v. GPRMC
 - b. Proprietary Strings
 - i. PSXN 20
 - ii. PSXN 22
 - iii. PSXN 23
2. ***Seapath 200***
 - a. NMEA 0182 strings
 - i. GPZDA
 - ii. GPGGA
 - iii. GPVTG
 - iv. GPHDT
 - b. Proprietary Strings
 - i. PSXN 20
 - ii. PSXN 22
 - iii. PSXN 23
3. ***Furuno GP-330B***
 - a. NMEA 0183 strings
 - i. GPZDA
 - ii. GPGGA
 - iii. GPVTG
 - iv. GPRMC
 - v. GPGLL
 - vi. GPDTM

GPZDA

15+051:21:02:04.507 \$GPZDA,210204.39,20,02,2015,,*6F

| Field | Data | Format | Unit |
|-------|----------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPZDA | | |
| 3 | Time | hhmmss.ss | UTC |
| 4 | Day | dd | UTC |
| 5 | Month | mm | UTC |
| 6 | Year | yyyy | UTC |
| 7 | (empty field) | x | Blank or 0 |
| 8 | Checksum | xx | alphanumeric |

GPGBA

15+051:21:02:02.507 \$GPGBA,210202.38,7712.979244,S,16741.040258,W,1,12,0.7,-5.04,M,-55.90,M,,*6F

| Field | Data | Format | Unit |
|-------|-------------------------------------|---------------------|-------------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPGBA | | |
| 3 | Time | hhmmss.ss | UTC |
| 4 | Latitude | ddmm.mmmmmm | degrees |
| 5 | North or South (for previous) | x | N or S |
| 6 | Longitude | ddmm.mmmmmm | degrees |
| 7 | East or West (for previous) | x | E or W |
| 8 | GPS quality indicator* | x | 0,1,2,3,4,5, or 6 |
| 9 | Number of satellites in use (00-99) | xx | 00-99 |
| 10 | HDOP | x.x | |
| 11 | Antenna height | x.xx | m |
| 12 | M = Meters (for previous) | x | M |
| 13 | Geoidal height | x.xx | m |
| 14 | M = Meters (for previous) | x | M |
| 15 | Age of DGPS corrections (seconds) | x.x | seconds |
| 16 | Station ID of DGPS (if used) | x | numeric |
| 17 | Checksum | xx | alphanumeric |

Quality

0 = invalid, 1 = GPS SPS, 2 = DGPS, 3 = PPS, 4 = RTK, 5 = float RTK, 6 = dead reckoning

GPVTG

15+051:16:47:06.625 \$GPVTG,357.84,T,251.99,M,9.5,N,17.7,K,A*15

| Field | Data | Format | Unit |
|-------|---|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPVTG | | |
| 3 | Heading | x.xx | degrees |
| 4 | T = True (for previous) | x | T |
| 5 | Heading | x.xx | degrees |
| 6 | M = Magnetic (for previous) | x | M |
| 7 | Speed over Ground (knots) | x.x | knots |
| 8 | N = knots (for previous) | x | N |
| 9 | Speed over Ground (kilometers per hour) | x.x | km/h |
| 10 | K = km per hour (for previous) | x | K |
| 11 | Mode* | X | A,D,E, or N |
| 12 | Checksum | xx | alphanumeric |

Modes

A = GPS used, D = DGPS used, E = Dead reckoning used, N = Invalid position / velocity

GPRMC

15+051:21:02:04.741 \$GPRMC,210204.38,A,7712.979182,S,16741.063669,W,9.4,270.82,200215,105.6,E,A*06

| Field | Data | Format | Unit |
|-------|-------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPRMC | | |
| 3 | Time | hhmmss.sss | UTC |
| 4 | Status* | x | A or N |
| 5 | Latitude | ddmm.mmmmmm | degrees |
| 6 | North or South (for previous) | x | N or S |
| 7 | Longitude | ddmm.mmmmmm | degrees |
| 8 | East or West (for previous) | x | E or W |
| 9 | Speed over Ground, True | x.x | knots |
| 10 | Course over Ground True | x.xx | degrees |
| 11 | Date | ddmmyy | UTC |
| 12 | Magnetic Variation | x.x | degrees |
| 13 | East or West (for previous) | x | E or W |
| 14 | Mode* | x | alphanumeric |
| 15 | Checksum | xx | UTC |

GPHT

15+051:21:02:04.741 \$GPHT,268.87,T*06

| Field | Data | Format | Unit |
|-------|-------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPHT | | |
| 3 | Heading, True | x.xx | degrees |
| 4 | T = True (for previous) | x | T |
| 5 | Checksum | xx | alphanumeric |

GPGLL

16+077:00:00:00.725 \$GPGLL,6356.6505,S,05716.0002,W,000000,A,A*4F

| Field | Data | Format | Unit |
|-------|---------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPGLL | | |
| 3 | Latitude | ddmm.mmmmmm | degrees |
| 4 | North or South (for previous) | x | N or S |
| 5 | Longitude | ddmm.mmmmmm | degrees |
| 6 | East or West (for previous) | x | E or W |
| 7 | Time of Position (not received) | hhmmss.ss | UTC |
| 8 | Status* | x | A or V |
| 9 | Mode* | x | alphanumeric |
| 10 | Checksum | xx | alphanumeric |

Status

A = Data Valid, V = Data not valid

Modes

A = GPS used, D = DGPS used, E = Dead reckoning used, M = Manual input mode, S = Simulator Mode, N = Invalid position / velocity

GPDTM

16+077:00:00:02.527 \$GPDTM,W84,,0000.0000,N,00000.0000,E,0.0,W84*5F

| Field | Data | Format | Unit |
|-------|-------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$GPDTM | | |
| 3 | Local Datum Code* | XXX | alphanumeric |
| 4 | Local datum subdivision code | x | numeric |
| 6 | Lat offset | x | alphanumeric |
| 7 | North or South (for previous) | x | N or S |
| 8 | Lon offset | x | alphanumeric |
| 9 | East or West (for previous) | x | E or W |
| 10 | Altitude offset, meters | x,x | numeric |
| 11 | Reference datum code* | xxx | alphanumeric |
| 12 | Checksum | xx | alphanumeric |

Datum Codes

W84 = WGS84, W72 = WGS72, S85 = SGS85, P90 = PE90, 999 = User defined

PSXN 20

15+051:22:20:58.740 \$PSXN,20,1,0,0,0*3A

| Field | Data | Format | Unit |
|-------|---|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$PSXN | | |
| 3 | 20 | | |
| 4 | Horizontal position and velocity quality* | x | 0,1,2 |
| 5 | Height and vertical velocity quality* | x | 0,1,2 |
| 6 | Heading quality* | x | 0,1,2 |
| 7 | Roll and pitch quality* | x | 0,1,2 |
| 8 | Checksum | xx | alphanumeric |

Qualities

0 = Normal, 1 = Reduced Performance, 2 = Invalid data

PSXN 22

15+051:22:20:59.019 \$PSXN,22,0.43,0.50*3B

| Field | Data | Format | Unit |
|-------|---|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$PSXN | | |
| 3 | 22 | | |
| 4 | Gyro calibration value since system startup | x.xx | degrees |
| 5 | Short-term gyro offset | x.xx | degrees |
| 6 | Checksum | xx | alphanumeric |

PSXN 23

15+051:22:20:58.748 \$PSXN,23,-0.20,-0.09,279.85,0.24*34

| Field | Data | Format | Unit |
|-------|--------------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$PSXN | | |
| 3 | 23 | | |
| 4 | Roll, port side up is positive | x.xx | degrees |
| 5 | Pitch, bow up is positive | x.xx | degrees |
| 6 | Heading, True | x.xx | degrees |
| 7 | Heave, positive is down | x.xx | m |
| 8 | Checksum | xx | alphanumeric |

Gyro Compass (gyr1)

15+055:20:27:23.653 \$HEHDT,087.31,T*12

| Field | Data | Format | Unit |
|-------|-------------------------|---------------------|--------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | \$HEHDT | | |
| 3 | Heading | x.xx | degrees |
| 4 | T = True (for previous) | x | T |
| 5 | Checksum | xx | alphanumeric |

ADCP Course (adcp)

15+049:20:20:57.327 \$PUHAW,UVH,-0.07,-4.59,179.3

| Field | Data | Format | Unit |
|-------|--|--------|---------|
| 1 | RVDAS time tag | | |
| 2 | \$PUHAW | | |
| 3 | UVH | | |
| 4 | Ship Speed relative to reference layer, east vector | x.xx | knots |
| 5 | Ship Speed relative to reference layer, north vector | x.xx | knots |
| 6 | Ship heading | x.xx | degrees |

Processed Data /process**pCO₂ – Merged**

15+055:11:24:43.960 2015055.46596 2534.72 32.41 975.33 48.86 356.94 341.67 -1.20 -1.27 0.00 Equil -
 75.9209 178.9696 -1.435 33.852 2.26 7.86 137.38 975.34 163.80 9.31 253.75 NaN -1.27 33.84 -1.14 -
 1.0

| Field | Data | Format | Unit |
|-------|--------------------------------------|---------------------|----------------------|
| 1 | RVDAS time tag | yy+ddd:hh:mm:ss.sss | UTC |
| 2 | pCO ₂ time tag* | yyyyddd.ttt | UTC |
| 3 | Raw Voltage (IR) | xxxx.xx | mV |
| 4 | Cell Temperature | xx.xx | C |
| 5 | Equilibration Pressure | xxx.xx | mBar |
| 6 | Flowrate | xxx.xx | cm ³ /min |
| 7 | pCO ₂ Pressure | xxx.xx | μAtm |
| 8 | VCO ₂ Concentration | xx.xx | ppm |
| 9 | Equilibrator Temperature, RTD | xx.xx | C |
| 10 | Equilibrator Temperature, SBE38 | xx.xx | C |
| 11 | Valve Position | xx | numeric |
| 12 | Flow Source* | | text |
| 13 | Latitude | xx.xxxxx | degrees |
| 14 | Longitude | xxx.xxxxx | degrees |
| 15 | Sea Water Intake Temperature | xx.xxx | C |
| 16 | Sea Surface Salinity | xx.xxx | PSU |
| 17 | Sea Surface Fluorometry | x.xxx | mg/m ³ |
| 18 | True Wind Speed | x.xx | m/s |
| 19 | True Wind Direction | x.xx | degrees |
| 20 | Barometric Pressure | xxx.xx | mBar |
| 21 | Hydro-Lab H ₂ O Flow Rate | xxx.x | l / min |
| 22 | Speed over Ground | x.xx | knots |
| 23 | Course Made Good | xx.xx | degrees |
| 24 | Unused | | |
| 25 | TSG2 Temperature | x.xx | C |
| 26 | TSG2 Salinity | xx.xx | PSU |
| 27 | TSG1 Temperature | x.xx | C |
| 28 | Sea Water Valve* | x | -1 or 0 |

pCO₂ time tag

ttt = fractional time of day

Flow Source

Equil = pCO₂ Measurement

Sea Water Valve

-1 = Stern Thruster Valve, 0 = Moon Pool Valve

Calculations

PAR

Coefficients `parc1` and `parcv` for this cruise can be found in the `instrument.coeff` file as the variable labeled PAR, respectively. Variable `par` is the raw data in mV, as described in the “mw1” file description. The calibration scale and probe offset dark are values taken from the PAR Cal Sheet.

```
par = raw data mV
calibration scale = 5.8644 V/(μEinstiens/cm2sec)
parc1 = 1 / scale = .17
probe offset dark = -.1 mV
parcv = dark x 1000 mV/V = -0.0001 V
((par / 1000 mV/V) - parcv) x parc1 x 10000 cm2/m2 = μEinstiens/m2sec
```

Calculations (extracted from the C code):

```
/* Convert from mV to V */
par /= 1000;
/* (par V - vdark V) / Calibration Scale Factor V/uE/cm2sec */
parCalc = (par - parcv) * parc1 * 10000;
```

PSP

Coefficient `pspCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled PSP1. Variable `psp` is the raw data in mV, as described in the “mw1” file description.

```
psp = raw data mV
calibration scale = pspCoeff x 10^-6 V/(W/m2)
psp / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* Convert from mV to W/m^2 */
pspCalc = (psp * 1000 / pspCoeff);
```

PIR

Coefficient `pirCoeff` for this cruise can be found in the `instrument.coeff` file as the variable labeled `PIR1`. Variable `pir_thermo` is the raw data in mV, `pir_case` is the PIR case temperature in Kelvins and `pir_dome` is the PIR dome temperature in Kelvins, as described in the “`mwx1`” file description. Hard-coded “C” coefficients are shown below:

```
Dome constant = 3.5
```

```
Sigma = 5.6704e-8
```

```
pir_thermo = raw data mV
```

```
calibration scale = pirCoeff x 10-6 V/(W/m2)
```

```
pir_thermo / (scale x 1000 mV/V) = W/m2
```

Calculations (extracted from the C code):

```
/* convert mV to W/m^2 */  
pirCalc = (pir_thermo * 1000 / pirCoeff)  
/* correct for case temperature */  
pirCalc += sigma * pow(pir_case,4)  
/* correct for dome temperature */  
pirCalc -= 3.5 * sigma * (pow(pir_dome, 4) - pow(pir_case, 4))
```

This section lists problems with acquisition noted during this cruise including instrument failures, data acquisition system failures and any other factor affecting this data set. The format is ddd:hh:mm (ddd is year-day, hh is hour, and mm is minute). Times are reported in UTC.

[illegible]

Appendix A: Sensors

NBP17-09 Sensors

| Sensor | Manufacturer | Model | Serial No. | Last Cal Date | Comments |
|-----------------------------------|------------------|-------------|-------------|---------------|----------------|
| Meteorology and Radiometry | | | | | |
| Anemometer | RM Young | 5106 | WM128975 | 10/27/2011 | ECO/Bridge |
| Anemometer, U/S | Gill Instruments | 1390-PK-062 | 847014 | 9/29/2010 | Starboard Side |
| Anemometer, U/S | Gill Instruments | 1390-PK-062 | 924057 | 11/18/2009 | Port Side |
| Barometer | RM Young | 61201 | 01705 | 1/22/2016 | |
| Barometer | Vaisala | PTB210B | M2750443 | 7/13/2016 | |
| Humidity/Wet Temp | RM Young | 41382LC | 06733 | 10/6/2016 | |
| PAR for Mast | Biosph. Inst. | QSR-2200 | 20531 | 3/8/2017 | |
| PIR | Eppley | PIR | 33023F3 | 10/19/2016 | |
| PSP | Eppley | PSP | 33090F3 | 10/18/2016 | |
| GUV (Mast) | Biosph. Inst. | GUV-2511 | 25110203113 | 5/31/2017 | |

| Sensor | Manufacturer | Model | Serial No. | Last Cal Date | Comments |
|---------------------|----------------|--------------|---------------|---------------|-----------|
| CTD | | | | | |
| SBE 11+ Deck Unit | Sea-Bird | 11+ | 11P47914-0768 | NA | |
| Conductivity | Sea-Bird | 4 - 02/O | 041314 | 12/1/2016 | Primary |
| Conductivity | Sea-Bird | 4C 6800m | 041798 | 3/22/2016 | Secondary |
| CTD Fish | Sea-Bird | SBE 9+ | 91480 | 1/19/2017 | |
| CTD Pressure Sensor | Paroscientific | 410K-105 | 53952 | 1/19/2017 | |
| Dissolved Oxygen | Sea-Bird | SBE 43 | 0150 | 11/18/2016 | Primary |
| Dissolved Oxygen | Sea-Bird | SBE 43 | 0155 | 11/18/2016 | Secondary |
| CTD Pump | Sea-Bird | 5T, PN 90543 | 055641 3.0K | 1/3/2017 | Primary |
| CTD Pump | Sea-Bird | 5T, PN 90543 | 055643 3.0K | 1/3/2017 | Secondary |
| Fluorometer | WetLabs | FLRTD | FLRTD-1482 | 3/19/2016 | |
| Temperature | Sea-Bird | 3-02/F | 031649 | 3/14/2016 | Primary |
| Temperature | Sea-Bird | 3plus 6800m | 03P5730 | 2/26/2016 | Secondary |
| Transmissometer | WET Labs | C-Star | CST-439DR | 7/29/2016 | |

| Sensor | Manufacturer | Model | Serial No. | Last Cal Date | Comments |
|--|--------------|--------|--------------|---------------|-----------|
| Underway Seawater Sampling System | | | | | |
| Fluorometer | WetLabs | FLRTD | FLRTD-855 | 1/27/2016 | |
| Transmissometer | WET Labs | C-Star | CST-1581DR | 3/30/2016 | |
| Micro-TSG | Sea-Bird | SBE 45 | 4549120-0226 | 2/19/2017 | Primary |
| Micro-TSG | Sea-Bird | SBE 45 | 4566350-0389 | 10/5/2016 | Secondary |
| Digital Remote Temp | Sea-Bird | SBE 38 | 3846730-0323 | 1/23/2016 | |

Appendix B: Calibration Sheets

Gravity

BGM3 ship-to-shore gravity tie report

AUKON/CHIN, vessel: R/V Palmer

Release Date: 2017/06/17 03:11:38 UTC

Sensor: S210

Software version: 1.2

Port/Pier/Berth: LYTTLETON PIER 7

| | |
|----------------------------|---|
| Gravity station number | 486732A CHRISTCHURCH |
| Station name | NEW ZEALAND NATIONAL GRAVITY BASE STATION |
| mGal at pier | 980505.71 |
| Tie start time UTC | 2017/06/17 02:10:19.792 |
| Samples used | 3600 |
| Land tie used | Yes |
| Water height to pier 1 | 12 ft 3 in |
| Water height to pier 2 | 12 ft 8 in |
| Water height to pier 3 | 12 ft 11 in |
| Average of filtered counts | 25034.809182694 |
| Filter length | 181 |
| Scale factor | 4.994070552 |
| NEW BIAS | 855481.29 |

Table 1: Gravity tie information

Meteorology

Anemometers

Cal sheet not required

Barometer 1



R.M. Young Company
2801 Aero Park Drive
Traverse City, Michigan 49586 USA

CALIBRATION REPORT Barometric Pressure

Customer: *Lockheed Martin*

Test Number: 6122-018
Test Date: 22 January 2016

Customer PO: 4102344091
Sales Order: 5254

| | |
|---|------------------------|
| <u>Test Sensor:</u> | |
| Model: 61201 | Serial Number: BP01705 |
| Description: Barometric Pressure Sensor | |

Report of calibration comparison of test barometric pressure sensor with National Institute of Standards and Technology traceable standard pressure calibrator at five pressures in the R.M. Young Company controlled pressure facility. Calibration accuracy ± 1.0 hPa.

| Reference Pressure (hPa) | Voltage Output (millivolts) | Indicated (1) Pressure (hPa) |
|--------------------------|-----------------------------|------------------------------|
| 800.0 | 0 | 800.0 |
| 875.0 | 1250 | 875.0 |
| 950.0 | 2501 | 950.0 |
| 1025.0 | 3750 | 1025.0 |
| 1100.0 | 4998 | 1099.9 |

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument
Druck Pressure Controller Model DPI515
Fluke Multimeter Model 8060A

Serial # NIST Test Reference
51500497 0046591
4865407 8804897

Tested By: *R. Rullman*

METEOROLOGICAL INSTRUMENTS
Tel: 231-946-3980 Fax: 231-546-4772 Email: met.sales@youngusa.com Website: youngusa.com
ISO 9001:2008 CERTIFIED

Barometer 2

VAISALA

1 (1)
Certificate report no. H44-15280021

CALIBRATION CERTIFICATE

Instrument PTB210B Digital Barometer
Serial number M2750443
Manufacturer Vaisala Oyj, Finland
Calibration date 13th July 2016

The above instrument was calibrated by comparing the readings of the instrument to the factory working standard of Vaisala.

The pressure readings of the factory working standard have been calibrated at an ISO/IEC 17025 accredited calibration laboratory (FINAS), Vaisala Measurement Standards Laboratory (MSL), by using MSL working standards traceable to NIST.

Calibration results

| Reference hPa | Observed hPa | Correction* hPa | Acceptance limit hPa |
|------------------|-----------------|--------------------|-------------------------|
| 510.0 | 510.0 | 0.0 | ± 0.2 |
| 610.0 | 610.0 | 0.0 | ± 0.2 |
| 700.0 | 700.0 | 0.0 | ± 0.2 |
| 810.0 | 810.0 | 0.0 | ± 0.2 |
| 910.0 | 910.0 | 0.0 | ± 0.2 |
| 950.0 | 950.0 | 0.0 | ± 0.2 |
| 1000.0 | 1000.0 | 0.0 | ± 0.2 |
| 1098.0 | 1098.0 | 0.0 | ± 0.2 |


*To obtain the true pressure, add the correction to the barometer reading.
Interpolated corrections may be used at intermediate readings of the scale of the barometer.

Equipment used in calibration

| Type | Serial number | Calibration date | Certificate number |
|------|---------------|------------------|--------------------|
| PPC4 | 670 | 2015-08-21 | K008-Y01989 |

Uncertainty (95 % confidence level, k=2)
Pressure ± 0.15 hPa

Ambient Conditions
Humidity 39 %RH ± 5 %RH
Temperature 21 °C ± 1 °C
Pressure 1002 hPa ± 1 hPa


Technician

This report shall not be reproduced except in full, without the written approval of Vaisala.

doc223087-R

Vaisala Oyj | PO Box 26, FI-00042 Vaisala, Finland

Temperature / Humidity



R.M. Young Company
2801 Aero Park Drive
Traverse City, Michigan 49686 USA

CALIBRATION REPORT**Temperature**

Customer *Lockheed Martin*

Test Number: 6300-01T
Test Date: 6 October 2016

Customer PO: 4102385716
Sales Order: 5771

Test Sensor:

Model: 41382LC2 Serial Number: TS06733
Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test temperature sensor with National Institute of Standards and Technology traceable standard thermometer at three temperatures in the R.M. Young Company controlled temperature calibration facilities. Calibration accuracy $\pm 0.2^\circ$ Celsius.

| Reference Temperature (degrees C) | Current Output (mA) | Indicated (1) Temperature (degrees C) | Calculated (2) Temperature (degrees C) |
|---|---------------------------|---|--|
| 45.29 | 19.27 | 45.47 | 45.28 |
| 19.95 | 15.24 | 20.27 | 19.97 |
| -10.38 | 10.41 | -9.94 | -10.38 |

(1) Calculated from current output.
(2) Calculated values using derived formula: $T = -75.745 + \text{mA} \times 6.27899$

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

| Reference Instrument: | Serial # | NIST Test Reference |
|--|----------|---------------------|
| Brooklyn Thermometer Model 43-FC | 8006-118 | W204690 |
| Brooklyn Thermometer Model 22332-D5-FC | 26371 | W204691 |
| Brooklyn Thermometer Model 2X400-D7-FC | 77532 | W204692 |
| Keithley Multimeter Model 191 | 15232 | 7124815 |

Tested By: *S. Sage*

METEOROLOGICAL INSTRUMENTS
Tel: 231-946-3880 Fax: 231-946-4772 E-mail: met.sales@youngusa.com Website: youngusa.com
ISO 9001:2008 CERTIFIED



R.M. Young Company
2801 Aero Park Drive
Traverse City, Michigan 49686 USA

CALIBRATION REPORT
Relative Humidity

Customer: *Lockheed Martin*

Test Number: 6006-01R
Test Date: 6 October 2016

Customer PO: 4102685716
Sales Order: 5771

Test Sensor:

Model: 41362LC2 Serial Number: TS06733
Description: Temperature/Relative Humidity Sensor

Report of calibration comparison of test relative humidity sensor with National Institute of Standards and Technology traceable standard relative humidity sensor at five humidity levels in the R.M. Young Company controlled humidity chamber facility. Calibration accuracy = 2.0 %.

| Reference Humidity (%) | Current Output: (milliamps) | Indicated (1) Humidity (%) |
|------------------------|-----------------------------|----------------------------|
| 10.1 | 5.7 | 10.8 |
| 30.1 | 8.9 | 30.9 |
| 50.0 | 12.3 | 51.6 |
| 70.0 | 15.4 | 71.3 |
| 90.1 | 18.6 | 91.1 |

(1) Calculated from voltage output

All reference equipment used in this calibration procedure have been tested by comparison to traceable standards certified by the National Institute of Standards and Technology.

Reference Instrument
Vaisala Humidity Sensor Model 35A/C
Fluke Multimeter Model 8060A

Serial # NIST Test Reference
N475040 TN 266152
4865407 8604897

Tested By: *S. Sage*

METEOROLOGICAL INSTRUMENTS
Tel: 231-946-3960 Fax: 231-946-4772 Email: met.sales@youngusa.com Website: youngusa.com
ISO 9001:2008 CERTIFIED

PIR


THE EPPLEY LABORATORY, INC.

17 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
 Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

Calibration Certificate

Instrument: Precision Infrared Radiometer, Model PIR, Serial Number 33023F3

Procedure: This pyrogeometer was compared against Eppley's Blackbody Calibration System under radiation intensities of approximately 350 Wm^{-2} with an average ambient temperature of 23°C according to procedures described in Technical Procedure, TP05 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

Transfer Standard: Eppley Precision Infrared Radiometer, Model PIR, Serial Number 32227F3

Results: **Sensitivity:** $S = 3.91 \mu\text{V} / \text{Wm}^{-2}$
Uncertainty: $U_{95} = \pm 1.7\%$ (95% confidence level, $k=2$)
Resistance: 739Ω at 23°C

Date of Test: October 19, 2016

Traceability: This calibration is traceable to the International Practical Temperature Scale (IPTS). Additionally, transfer standard PIR #32227F3 provides traceability to the World Infrared Standard Group (WISG) of pyrogeometers housed at the Infrared Radiometry Section of the World Radiation Centre (WRC-IRS). Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: Leidos Innovations Corp
 Port Huene, CA

Signatures: D GIENTY
 In Charge of Test:

Thomas D. Kuhl
 Reviewed by:

Eppley SO 64821

Date of Certificate October 19, 2016

Remarks: Customer Instrument #866-00030843

End of Report

PSP


THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840
 Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

Calibration Certificate

Instrument: Precision Spectral Pyranometer, Model PSP, Serial Number 33090F3

Procedure: This pyranometer was compared in Eppley's Integrating Hemisphere according to procedures described in *ISO 9847 Section 5.3.7* and Technical Procedure, TP01 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

Transfer Standard: Eppley Standard Precision Pyranometer, Model SPP, Serial Number 37501F3

Results:
Sensitivity: $S = 7.90 \mu V / W m^{-2}$
Uncertainty: $U_{95} = \pm 0.91\%$ (95% confidence level, k=2)
Resistance: 700 Ω at 23°C
Date of Test: October 18, 2016

Traceability: This calibration is traceable to the World Radiation Reference (WRR) through comparisons with Eppley's AHF standard self-calibrating cavity pyrheliometers which participated in the Twelfth International Pyrothermic Comparisons (IPC XII) at Davos, Switzerland in September-October 2015. Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: Leidos Innovations Corp
 Port Hueneme, CA

Signatures:
 In Charge of Test: D. Gieny
 Reviewed by: Nathaniel B. Palmer

Eppley SO: 64822

Date of Certificate: October 19, 2016

Remarks: Customer Instrument ID #866-00038044

End of Report

PAR (Mast)

Biospherical Instruments Inc.

CALIBRATION CERTIFICATE

Calibration Date: 3/8/2017
 Model Number: QSR2200
 Serial Number: 20531
 Operator: TPC
 Standard Lamp: 91453(7/20/16)
 Probe Excitation Voltage Range: 6 to 18 VDC(+)

Output Polarity: POSITIVE

Probe Conditions at Calibration (in air):

Calibration Voltage: 6 VDC(+)

Probe Current: 4.0 mA

Probe Output Voltage:

Probe Illuminated: 90.1 mV
 Probe Dark: 1.0 mV
 Probe Net Response: 89.1 mV
 RG780: 1.1 mV

Corrected Lamp Output:

Output in Air (same condition as calibration):

8.379E+15 quanta/cm²sec
0.013014 uE/cm²sec

Calibration Scale Factor:

(To calculate irradiance, divide the net voltage reading in Volts by this value.)

Dry: 1.0633E-17 V/(quanta/cm²sec)
9.4030E+00 V/(uE/cm²sec)

Notes:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).
3. The collector should be cleaned frequently with alcohol.
4. Calibration was performed with customer cable, when available.

QSR240R 05/24/95

[Insert Page Break]

GUV (Mast)

**Biospherical Instruments Inc.**

System Calibration Certificate

THE INSTRUMENTS REFERENCED BELOW WERE FACTORY TESTED AND CALIBRATED BY

BIOSPHERICAL INSTRUMENTS INC.
5340 Riley Street
San Diego, California 92110 USA

Instruments: GUV-2511 No 25110203113

Optical Calibrations:
NIST Traceability. For wavelengths longer than 313 nm, the specific instruments cited here were calibrated using a 1000W FEL #V-039 (7/20/16) following procedures and standards traceable to NIST Standard of Spectral Irradiance F-616. Traceability paths and all procedures for all calibrated lamps and associated apparatus (shunts, power supplies, DMMs, etc) are maintained following calibration methodologies per National Bureau of Standards (US) (NBS) Special Publication 250-20 Spectral Irradiance Calibrations (1987) and NBS Publication 594-13 Optical Radiation Measurements: The 1973 Scale of Spectral Irradiance (1977).

Solar Calibrations. Lamp calibrations are problematic for solar UV measurements (wavelengths below 320 nm) because the solar spectrum is radically different from the lamp spectrum and changes greatly as a function of wavelength. Solar calibrations are achieved through direct comparison with measurements of a high resolution scanning spectroradiometer in San Diego (SUV-100), which is part of the National Science Foundation's UV Monitoring Network. The SUV-100 instrument has a bandwidth of 1 nm. Calibrated filter radiometer data therefore report spectral irradiance at the channel's nominal wavelengths with a bandwidth of 1 nm. Solar calibrations are typically accurate to within $\pm 10\%$ for solar zenith angles smaller than 75° . At larger solar zenith angles, UV channels have a greater uncertainty due to the rapid change of the solar UV spectrum.

Note that this certificate contains a subset of the information delivered in the calibration database 25110203113v9.mdb. This database is required for operation of this system using Biospherical Instruments Inc.'s Logger® software.



Biospherical Instruments Inc.

GUV-2511 Calibration Certificate

| | | | | | | | | | | | |
|---------------------------|---------|------------|---|--|--|--|---------------------------------|--------------|-------------|--|--|
| System Serial Number | | | 25110203113 | | | | Date of Calibration | | | 5/31/2017 | |
| Calibration database | | | 25110203113v9.mdb | | | | Date of Certificate | | | 5/31/2017 | |
| DASSN | | | 0068 | | | | Standard of Spectral Irradiance | | | V-039(7/20/16) | |
| Microprocessor Tag Number | | | 2 | | | | Operator | | | TC | |
| | | | | | | | | | | | |
| Monochromatic | | Wavelength | Responsivity | ScaleSmall | ScaleMedium | ScaleLarge | OffsetSmall | OffsetMedium | OffsetLarge | Measurement | |
| Channels | Address | [nm] | [Amps per $\mu\text{W}/(\text{cm}^2\text{-nm})$] | [Volts per $\mu\text{W}/(\text{cm}^2\text{-nm})$] | [Volts per $\mu\text{W}/(\text{cm}^2\text{-nm})$] | [Volts per $\mu\text{W}/(\text{cm}^2\text{-nm})$] | [volts] | [volts] | [volts] | Units | |
| Ed0320 | 2 | 320 | 2.5669E-10 | 2.6182E-05 | 7.6493E-03 | 2.7081E+00 | 8.7011E-05 | 8.5714E-05 | 4.3851E-04 | $\mu\text{W}/(\text{cm}^2\text{-nm})$ | |
| Ed0340 | 6 | 340 | 2.0781E-10 | 2.1196E-05 | 6.1927E-03 | 2.1835E+00 | 3.7847E-04 | 3.8104E-04 | 1.3278E-03 | $\mu\text{W}/(\text{cm}^2\text{-nm})$ | |
| Ed0313 | 8 | 313 | 2.2231E-10 | 2.2675E-05 | 6.6247E-03 | 2.3650E+00 | 3.5302E-05 | 3.7066E-05 | 6.5847E-04 | $\mu\text{W}/(\text{cm}^2\text{-nm})$ | |
| Ed0305 | 10 | 305 | 1.6361E-11 | 1.6688E-06 | 4.8755E-04 | 1.5082E-01 | 4.6550E-04 | 4.5669E-04 | -2.1536E-03 | $\mu\text{W}/(\text{cm}^2\text{-nm})$ | |
| Ed0380 | 12 | 380 | 6.2595E-11 | 6.3847E-06 | 1.8653E-03 | 6.2206E-01 | 1.0328E-03 | 1.0241E-03 | -1.5881E-03 | $\mu\text{W}/(\text{cm}^2\text{-nm})$ | |
| Ed0395 | 18 | 395 | 3.3548E-10 | 3.4219E-05 | 9.9974E-03 | 3.2656E+00 | 9.7613E-05 | 9.8620E-05 | 2.5961E-04 | $\mu\text{W}/(\text{cm}^2\text{-nm})$ | |
| | | | | | | | | | | | |
| Broadband | | Wavelength | Responsivity | ScaleSmall | ScaleMedium | ScaleLarge | OffsetSmall | OffsetMedium | OffsetLarge | Measurement | |
| Channels | Address | [nm] | [Amps per $\mu\text{E}/(\text{cm}^2\text{-s})$] | [Volts per $\mu\text{E}/(\text{cm}^2\text{-s})$] | [Volts per $\mu\text{E}/(\text{cm}^2\text{-s})$] | [Volts per $\mu\text{E}/(\text{cm}^2\text{-s})$] | [volts] | [volts] | [volts] | Units | |
| Ed0PAR | 13 | 400-700 | 1.8717E-05 | 1.9091E+00 | 5.5776E+02 | 1.9859E+05 | 9.0359E-04 | 8.9635E-04 | -1.2834E-03 | $\mu\text{E}/(\text{cm}^2\text{-sec})$ | |
| | | | | | | | | | | | |
| Auxiliary | | | | | | | | | | Measurement | |
| Channels | Address | Wavelength | Responsivity | ScaleS | ScaleM | ScaleL | OffsetS | OffsetM | OffsetL | Units | |
| Ed0Gnd | 0 | 0 | 1 | 1.00 | 1.00 | 1.00 | 0 | 0 | 0 | V | |
| Ed0Temp | 22 | 0 | 1 | 0.01 | 0.01 | 0.01 | 0 | 0 | 0 | C | |
| Ed0Vin | 27 | 0 | 1 | -0.25 | -0.25 | -0.25 | 0 | 0 | 0 | V | |

© Biospherical Instruments Inc., 5340 Riley Street, San Diego, California 92110 USA. Contact support@biospherical.com for more information.

Calibration Data – Do Not Destroy

page 2 of 2

Underway Seawater Sampling System

Micro-TSG 1

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0228
 CALIBRATION DATE: 19-Feb-17

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.013733e+000
 h = 1.569130e-001
 i = -5.222989e-004
 j = 6.723089e-005

CPcor = -9.5700e-008
 CTcor = 3.2500e-006
 WBOTC = 9.8072e-007


| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000 | 0.0000 | 0.00000 | 2548.99 | 0.00000 | 0.00000 |
| 1.0000 | 34.7782 | 2.97302 | 5055.47 | 2.97303 | 0.00001 |
| 4.4999 | 34.7582 | 3.27977 | 5245.81 | 3.27976 | -0.00001 |
| 15.0001 | 34.7152 | 4.26053 | 5811.87 | 4.26049 | -0.00004 |
| 18.5000 | 34.7063 | 4.60535 | 5997.94 | 4.60538 | 0.00004 |
| 24.0000 | 34.6972 | 5.16286 | 6286.70 | 5.16287 | 0.00001 |
| 29.0000 | 34.6931 | 5.68441 | 6544.93 | 5.68440 | -0.00001 |
| 32.5000 | 34.6918 | 6.05673 | 6722.93 | 6.05661 | -0.00012 |

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0228
 CALIBRATION DATE: 19-Feb-17

SBE 45 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.625436e-004
 a1 = 3.092515e-004
 a2 = -4.990410e-006
 a3 = 2.190124e-007

| BATH TEMP (° C) | INSTRUMENT OUTPUT (counts) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000 | 744277.3 | 0.9999 | -0.0001 |
| 4.4999 | 636452.3 | 4.5000 | 0.0001 |
| 15.0001 | 405760.1 | 14.9998 | -0.0003 |
| 18.5000 | 351373.2 | 18.5003 | 0.0003 |
| 24.0000 | 281929.6 | 23.9999 | -0.0001 |
| 29.0000 | 232180.3 | 29.0000 | -0.0000 |
| 32.5000 | 203348.2 | 32.5000 | 0.0000 |

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/(a0 + a1[\ln(n)] + a2[\ln^2(n)] + a3[\ln^3(n)]) - 273.15$

Residual (°C) = instrument temperature - bath temperature



Micro-TSG2

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0389
CALIBRATION DATE: 05-Oct-16SBE 45 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter


COEFFICIENTS:

g = -1.011215e+000
h = 1.476175e-001
i = -8.280491e-004
j = 7.730207e-005CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 1.2700e-007

| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000 | 0.0000 | 0.00000 | 2631.99 | 0.00000 | 0.00000 |
| 1.0000 | 34.7178 | 2.96835 | 5231.58 | 2.96833 | -0.00002 |
| 4.5000 | 34.6978 | 3.27464 | 5429.24 | 3.27466 | 0.00002 |
| 14.9999 | 34.6545 | 4.25385 | 6016.97 | 4.25386 | 0.00001 |
| 18.5000 | 34.6448 | 4.59807 | 6210.08 | 4.59806 | -0.00001 |
| 24.0000 | 34.6345 | 5.15456 | 6509.85 | 5.15455 | -0.00002 |
| 29.0000 | 34.6289 | 5.67507 | 6777.88 | 5.67508 | 0.00001 |
| 32.5000 | 34.6261 | 6.04656 | 6962.65 | 6.04663 | 0.00007 |

 $f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$ t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;Conductivity (S/m) = $(g + h * f + i * f^2 + j * f^3) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0389
 CALIBRATION DATE: 05-Oct-16

SBE 45 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = 3.090837e-005
 a1 = 2.658070e-004
 a2 = -1.844963e-006
 a3 = 1.332162e-007

| BATH TEMP (° C) | INSTRUMENT OUTPUT (counts) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000 | 828418.5 | 1.0000 | -0.0000 |
| 4.5000 | 706628.0 | 4.5000 | 0.0000 |
| 14.9999 | 447305.9 | 14.9999 | 0.0000 |
| 18.5000 | 386484.7 | 18.4999 | -0.0001 |
| 24.0000 | 309027.4 | 24.0000 | -0.0000 |
| 29.0000 | 253718.2 | 29.0001 | 0.0001 |
| 32.5000 | 221745.8 | 32.4999 | -0.0001 |

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



Digital Remote Temp

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0323
 CALIBRATION DATE: 23-Jan-16

SBE 38 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = 1.374421e-005
 a1 = 2.751852e-004
 a2 = -2.335169e-006
 a3 = 1.567304e-007

| BATH TEMP (° C) | INSTRUMENT OUTPUT (counts) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| -1.5000 | 711106.6 | -1.5000 | 0.0000 |
| 1.0000 | 635152.2 | 1.0000 | -0.0000 |
| 4.5000 | 543801.6 | 4.4999 | -0.0001 |
| 8.0000 | 467106.4 | 8.0000 | -0.0000 |
| 11.5000 | 402497.5 | 11.5000 | 0.0000 |
| 15.0000 | 347892.6 | 15.0000 | 0.0000 |
| 18.5000 | 301594.1 | 18.5001 | 0.0001 |
| 22.0000 | 262216.7 | 22.0000 | 0.0000 |
| 25.4999 | 228625.3 | 25.4998 | -0.0001 |
| 29.0002 | 199879.9 | 29.0001 | -0.0001 |
| 32.5000 | 175217.4 | 32.5001 | 0.0001 |

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1 [ln(n)] + a_2 [ln^2(n)] + a_3 [ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature



[Insert Page Break]

Transmissometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

| | | | | | |
|--|---------|---------------|----------------|------------|------|
| Date | 3.30.16 | S/N# | CST-1581DR | Pathlength | 25cm |
| | | Analog output | Digital output | | |
| V _d | | 0.004 V | 0 counts | | |
| V _{air} | | 4.806 V | 15795 counts | | |
| V _{ref} | | 4.699 V | 15445 counts | | |
| Temperature of calibration water | | | | 20.5 °C | |
| Ambient temperature during calibration | | | | 22.5 °C | |

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x, in meters): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water; temperature of clean water used to obtain V_{ref}.

Ambient temperature; meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.

Revision L

6/9/09

[Insert Page Break]

Fluorometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 1/27/2016

S/N: FLRTD-855

Chlorophyll concentration expressed in µg/l can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

| | Analog Range 1 | Analog Range 2 | Analog Range 4 (default) | Digital |
|-------------------|-------------------|-------------------|--------------------------------|-------------------|
| Dark Counts | 0.117 | 0.072 | 0.050 V | 72 counts |
| Scale Factor (SF) | 8 | 15 | 30 µg/l/V | 0.0092 µg/l/count |
| Maximum Output | 4.98 | 4.98 | 4.98 V | 16326 counts |
| Resolution | 0.6 | 0.6 | 0.6 mV | 1.0 counts |

Ambient temperature during characterization

22.3 °C

Analog Range: 1 (most sensitive, 0–4,000 counts), 2 (midrange, 0–8,000 counts), 4 (entire range, 0–16,000 counts).

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $SF = x \div (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

FLRTD-855

Revision J

3/17/08

[Insert Page Break]

CTD

Conductivity (primary)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1314
 CALIBRATION DATE: 01-Dec-16

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -3.96047725e+000
 h = 4.57902311e-001
 i = -8.86036213e-005
 j = 2.95672674e-005

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)


| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (kHz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|--------------------|-------------------|--------------------|----------------------------|--------------------------|-------------------|
| 0.0000 | 0.0000 | 0.00000 | 2.94096 | 0.00000 | 0.00000 |
| -1.0000 | 34.6325 | 2.79115 | 8.33098 | 2.79116 | 0.00002 |
| 1.0000 | 34.6329 | 2.96178 | 8.55024 | 2.96177 | -0.00001 |
| 15.0000 | 34.6325 | 4.25145 | 10.05167 | 4.25140 | -0.00005 |
| 18.5000 | 34.6312 | 4.59646 | 10.41610 | 4.59648 | 0.00003 |
| 29.0001 | 34.6295 | 5.67517 | 11.47918 | 5.67521 | 0.00004 |
| 32.5001 | 34.6234 | 6.04616 | 11.82206 | 6.04612 | -0.00003 |

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Conductivity (secondary)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1798
CALIBRATION DATE: 22-Mar-16SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:


g = -9.79125732e+000
h = 1.26410805e+000
i = 1.34277903e-004
j = 4.48411138e-005CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

| BATH TEMP (°C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (kHz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|-------------------|-------------------|--------------------|----------------------------|--------------------------|-------------------|
| 0.0000 | 0.0000 | 0.00000 | 2.78230 | 0.00000 | 0.00000 |
| -1.0000 | 34.5316 | 2.78377 | 5.45147 | 2.78376 | -0.00001 |
| 1.0000 | 34.5320 | 2.95397 | 5.57326 | 2.95399 | 0.00002 |
| 15.0000 | 34.5334 | 4.24056 | 6.41909 | 4.24055 | -0.00001 |
| 18.5000 | 34.5333 | 4.58486 | 6.62706 | 4.58485 | -0.00001 |
| 29.0001 | 34.5313 | 5.66088 | 7.23837 | 5.66090 | 0.00003 |
| 32.4999 | 34.5234 | 6.03065 | 7.43674 | 6.03064 | -0.00002 |

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity



Fluorometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 3/19/2016

S/N: FLRTD-1482

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

| | Analog Range 1 | Analog Range 2 | Analog Range 4 (default) | Digital |
|-------------------|-------------------|-------------------|--------------------------------|------------------------------|
| Dark Counts | 0.066 | 0.035 | 0.020 V | 50 counts |
| Scale Factor (SF) | 6 | 12 | 25 $\mu\text{g/l/V}$ | 0.0075 $\mu\text{g/l/count}$ |
| Maximum Output | 4.99 | 4.99 | 4.99 V | 16380 counts |
| Resolution | 0.5 | 0.5 | 0.5 mV | 1.0 counts |

Ambient temperature during characterization

22.3 °C

Analog Range: 1 (most sensitive, 0–4,000 counts), 2 (midrange, 0–8,000 counts), 4 (entire range, 0–16,000 counts).

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: $\text{SF} = x \div (\text{output} - \text{dark counts})$, where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations *in-situ* is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

FLRTD-1482

Revision J

3/17/08

Pressure Sensor

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1480
 CALIBRATION DATE: 19-Jan-17

SBE 9plus PRESSURE CALIBRATION DATA
 10000 psia S/N 53952

DIGIQUARTZ COEFFICIENTS:


C1 = -5.561704e+004
 C2 = 4.302402e-001
 C3 = 1.582810e-002
 D1 = 4.708200e-002
 D2 = 0.000000e+000
 T1 = 3.029296e+001
 T2 = -2.122954e-004
 T3 = 4.352450e-006
 T4 = 2.626550e-009
 T5 = 0.000000e+000

AD590M, AD590B, SLOPE AND OFFSET:

AD590M = 1.16300e-002
 AD590B = -8.63457e+000
 Slope = 1.00003
 Offset = -3.0883 (dbars)

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT TEMPERATURE (°C) | INSTRUMENT PRESSURE (PSIA) | CORRECTED PRESSURE (PSIA) | RESIDUAL (PSIA) |
|--------------------|---------------------------|--------------------------------|-------------------------------|------------------------------|--------------------|
| 14.368 | 33019.50 | 21.9 | 19.428 | 14.949 | 0.581 |
| 2001.245 | 33602.80 | 22.0 | 2005.236 | 2000.825 | -0.420 |
| 3988.124 | 34174.60 | 22.1 | 3992.274 | 3987.932 | -0.192 |
| 5975.189 | 34735.10 | 22.1 | 5979.167 | 5974.893 | -0.296 |
| 7962.358 | 35285.20 | 22.1 | 7967.167 | 7962.961 | 0.603 |
| 9949.890 | 35824.70 | 22.2 | 9953.696 | 9949.559 | -0.331 |
| 7962.375 | 35285.20 | 22.2 | 7967.156 | 7962.951 | 0.576 |
| 5975.203 | 34735.10 | 22.2 | 5979.148 | 5974.874 | -0.329 |
| 3988.221 | 34174.70 | 22.2 | 3992.602 | 3988.260 | 0.039 |
| 2001.256 | 33602.80 | 22.3 | 2005.213 | 2000.802 | -0.454 |
| 14.365 | 33019.40 | 22.3 | 19.069 | 14.590 | 0.225 |

Residual (PSIA) = corrected instrument pressure - reference pressure



Dissolved Oxygen (primary)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0150
 CALIBRATION DATE: 18-Nov-16

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS: A = -5.1283e-003
 Soc = 0.4594 B = 2.7364e-004
 Voffset = -0.4992 C = -3.5992e-006
 Tau20 = 1.65 E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS
 D1 = 1.92634e-4 H1 = -3.300000e-2
 D2 = -4.64803e-2 H2 = 5.00000e+3
 H3 = 1.45000e+3


| BATH OXYGEN (ml/l) | BATH TEMPERATURE (°C) | BATH SALINITY (PSU) | INSTRUMENT OUTPUT (volts) | INSTRUMENT OXYGEN (ml/l) | RESIDUAL (ml/l) |
|--------------------|-----------------------|---------------------|---------------------------|--------------------------|-----------------|
| 1.10 | 12.00 | 0.00 | 0.827 | 1.10 | -0.00 |
| 1.11 | 2.00 | 0.00 | 0.750 | 1.10 | -0.00 |
| 1.11 | 6.00 | 0.00 | 0.781 | 1.10 | -0.00 |
| 1.12 | 20.00 | 0.00 | 0.891 | 1.12 | 0.00 |
| 1.14 | 26.00 | 0.00 | 0.942 | 1.14 | 0.00 |
| 1.14 | 30.00 | 0.00 | 0.974 | 1.15 | 0.00 |
| 3.85 | 2.00 | 0.00 | 1.374 | 3.85 | 0.00 |
| 3.88 | 12.00 | 0.00 | 1.651 | 3.88 | 0.00 |
| 3.88 | 6.00 | 0.00 | 1.490 | 3.88 | 0.00 |
| 3.90 | 20.00 | 0.00 | 1.862 | 3.90 | 0.00 |
| 3.92 | 26.00 | 0.00 | 2.021 | 3.92 | 0.00 |
| 3.92 | 30.00 | 0.00 | 2.122 | 3.92 | 0.00 |
| 6.60 | 6.00 | 0.00 | 2.185 | 6.60 | -0.00 |
| 6.61 | 2.00 | 0.00 | 1.999 | 6.61 | -0.00 |
| 6.62 | 12.00 | 0.00 | 2.465 | 6.62 | 0.00 |
| 6.66 | 30.00 | 0.00 | 3.251 | 6.66 | -0.00 |
| 6.66 | 20.00 | 0.00 | 2.828 | 6.66 | -0.00 |
| 6.71 | 26.00 | 0.00 | 3.103 | 6.71 | -0.00 |

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen



Dissolved Oxygen (secondary)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0150
 CALIBRATION DATE: 18-Nov-16

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS: A = -5.1283e-003
 Soc = 0.4594 B = 2.7364e-004
 Voffset = -0.4992 C = -3.5992e-006
 Tau20 = 1.65 E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS
 D1 = 1.92634e-4 H1 = -3.300000e-2
 D2 = -4.64803e-2 H2 = 5.00000e+3
 H3 = 1.45000e+3


| BATH OXYGEN (ml/l) | BATH TEMPERATURE (°C) | BATH SALINITY (PSU) | INSTRUMENT OUTPUT (volts) | INSTRUMENT OXYGEN (ml/l) | RESIDUAL (ml/l) |
|--------------------|-----------------------|---------------------|---------------------------|--------------------------|-----------------|
| 1.10 | 12.00 | 0.00 | 0.827 | 1.10 | -0.00 |
| 1.11 | 2.00 | 0.00 | 0.750 | 1.10 | -0.00 |
| 1.11 | 6.00 | 0.00 | 0.781 | 1.10 | -0.00 |
| 1.12 | 20.00 | 0.00 | 0.891 | 1.12 | 0.00 |
| 1.14 | 26.00 | 0.00 | 0.942 | 1.14 | 0.00 |
| 1.14 | 30.00 | 0.00 | 0.974 | 1.15 | 0.00 |
| 3.85 | 2.00 | 0.00 | 1.374 | 3.85 | 0.00 |
| 3.88 | 12.00 | 0.00 | 1.651 | 3.88 | 0.00 |
| 3.88 | 6.00 | 0.00 | 1.490 | 3.88 | 0.00 |
| 3.90 | 20.00 | 0.00 | 1.862 | 3.90 | 0.00 |
| 3.92 | 26.00 | 0.00 | 2.021 | 3.92 | 0.00 |
| 3.92 | 30.00 | 0.00 | 2.122 | 3.92 | 0.00 |
| 6.60 | 6.00 | 0.00 | 2.185 | 6.60 | -0.00 |
| 6.61 | 2.00 | 0.00 | 1.999 | 6.61 | -0.00 |
| 6.62 | 12.00 | 0.00 | 2.465 | 6.62 | 0.00 |
| 6.66 | 30.00 | 0.00 | 3.251 | 6.66 | -0.00 |
| 6.66 | 20.00 | 0.00 | 2.828 | 6.66 | -0.00 |
| 6.71 | 26.00 | 0.00 | 3.103 | 6.71 | -0.00 |

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen



Pump (primary)

SBE SEA-BIRD ELECTRONICS, INC.
13431 NE 20th St. Bellevue, Washington 98005 USA
Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

| | | |
|----------------|-------------------|-------|
| Service | RMA Number | 77298 |
| Report | | |

Customer Information:

| | | | |
|------------------|-----------------------------------|-------------|-----------|
| Company | Lockheed Martin Antarctic Support | Date | 3/31/2014 |
| Contact | Dave Morehouse | | |
| PO Number | 4900044522 | | |

| | |
|----------------------|---------|
| Serial Number | 055641 |
| Model Number | SBE 05T |

Services Requested:

| |
|--|
| 1. Evaluate/Repair Instrumentation. 2. Replace standard impulse connector(s) with NEW wet-pluggable connector(s). |
|--|

Problems Found:

| |
|--|
| |
|--|

Services Performed:

| |
|--|
| 1. Performed initial diagnostic evaluation. 2. Installed NEW two pin wet-pluggable bulkhead connector(s). 3. Performed internal inspection and O-ring and thrust washer replacements. 4. Performed hydrostatic pressure test. |
|--|

Special Notes:

| |
|--|
| |
|--|

Pump (secondary)

SBE SEA-BIRD ELECTRONICS, INC.
13431 NE 20th St. Bellevue, Washington 98005 USA
Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

| | | |
|----------------|-------------------|-------|
| Service | RMA Number | 77298 |
| Report | | |

Customer Information:

| | | | |
|----------------------|-----------------------------------|-------------|-----------|
| Company | Lockheed Martin Antarctic Support | Date | 3/31/2014 |
| Contact | Dave Morehouse | | |
| PO Number | 4900044522 | | |
| Serial Number | 055643 | | |
| Model Number | SBE 05T | | |

Services Requested:

| |
|--|
| 1. Evaluate/Repair Instrumentation. 2. Replace standard impulse connector(s) with NEW wet-pluggable connector(s). |
|--|

Problems Found:

| |
|--|
| |
|--|

Services Performed:

| |
|--|
| 1. Performed initial diagnostic evaluation. 2. Installed NEW two pin wet-pluggable bulkhead connector(s). 3. Performed internal inspection and O-ring and thrust washer replacements. 4. Performed hydrostatic pressure test. |
|--|

Special Notes:

| |
|--|
| |
|--|

Temperature (primary)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1649
 CALIBRATION DATE: 14-Mar-16

SBE 3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


g = 4.80855419e-003
 h = 6.67083809e-004
 i = 2.27462895e-005
 j = 1.61035917e-006
 f0 = 1000.0

| BATH TEMP (° C) | INSTRUMENT OUTPUT (Hz) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|---------------------------|--------------------|-------------------|
| -1.5000 | 5958.826 | -1.4996 | 0.00042 |
| 1.0000 | 6301.309 | 0.9996 | -0.00043 |
| 4.5000 | 6804.364 | 4.4995 | -0.00047 |
| 8.0000 | 7335.524 | 8.0001 | 0.00010 |
| 11.5000 | 7895.401 | 11.5004 | 0.00043 |
| 15.0000 | 8484.680 | 15.0004 | 0.00037 |
| 18.5000 | 9104.035 | 18.4999 | -0.00013 |
| 22.0000 | 9754.306 | 21.9999 | -0.00013 |
| 25.5000 | 10436.029 | 25.4997 | -0.00026 |
| 29.0000 | 11149.901 | 28.9998 | -0.00017 |
| 32.5000 | 11896.554 | 32.5003 | 0.00027 |

f = Instrument Output (Hz)

Temperature ITS-90 (°C) = $1 / (g + h[\ln(f0 / f)] + i[\ln^2(f0 / f)] + j[\ln^3(f0 / f)]) - 273.15$

Residual (°C) = instrument temperature - bath temperature



Temperature (Secondary)

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 5730
 CALIBRATION DATE: 26-Feb-16

SBE 3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


g = 4.35194524e-003
 h = 6.29868351e-004
 i = 1.97821726e-005
 j = 1.44030219e-006
 f0 = 1000.0

| BATH TEMP (° C) | INSTRUMENT OUTPUT (Hz) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|---------------------------|--------------------|-------------------|
| -1.5000 | 3003.663 | -1.5001 | -0.00008 |
| 1.0000 | 3179.306 | 1.0001 | 0.00009 |
| 4.5000 | 3437.566 | 4.5001 | 0.00006 |
| 8.0000 | 3710.656 | 8.0000 | -0.00001 |
| 11.5000 | 3999.011 | 11.4999 | -0.00005 |
| 15.0000 | 4303.052 | 14.9999 | -0.00005 |
| 18.5000 | 4623.188 | 18.5000 | -0.00004 |
| 22.0000 | 4959.824 | 22.0000 | 0.00004 |
| 25.5000 | 5313.340 | 25.5001 | 0.00007 |
| 29.0000 | 5684.113 | 29.0000 | 0.00003 |
| 32.5000 | 6072.511 | 32.4999 | -0.00006 |

f = Instrument Output (Hz)

Temperature ITS-90 (°C) = $1 / (g + h[\ln(f_0 / f)] + i[\ln^2(f_0 / f)] + j[\ln^3(f_0 / f)]) - 273.15$

Residual (°C) = instrument temperature - bath temperature



Transmissometer

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

| | | | | | |
|--|---------|------|-----------|------------|-------|
| Date | 7.29.16 | S/N# | CST-439DR | Pathlength | 25 cm |
| <hr/> | | | | | |
| Analog output | | | | | |
| V_d | 0.057 V | | | | |
| V_{air} | 4.752 V | | | | |
| V_{ref} | 4.659 V | | | | |
| Temperature of calibration water | 23.7 °C | | | | |
| Ambient temperature during calibration | 22.8 °C | | | | |

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x , in meters): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.

Revision M

7/26/11